

Data Path : Z:\HPCHEM1\BNA F\DATA\BF121117\
 Data File : BF101309.D
 Acq On : 12 Dec 2017 2:56
 Operator : SJ/JU
 Sample : I6744-03 2X
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PROS-3-E(7-8)

Quant Time: Dec 12 06:12:43 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF113017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 08 18:17:31 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.83	152	42293	20.00	ng	0.00
21) Naphthalene-d8	8.11	136	169516	20.00	ng	0.00
38) Acenaphthene-d10	9.87	164	71189	20.00	ng	0.00
63) Phenanthrene-d10	11.36	188	113721	20.00	ng	0.00
75) Chrysene-d12	14.02	240	93146	20.00	ng	0.01
86) Perylene-d12	15.51	264	67676	20.00	ng	0.04
System Monitoring Compounds						
5) 2-Fluorophenol	5.45	112	123395	48.21	ng	0.00
7) Phenol-d6	6.46	99	149435	48.09	ng	0.00
23) Nitrobenzene-d5	7.39	82	86224	30.34	ng	0.00
41) 2,4,6-Tribromophenol	10.66	330	28115	32.88	ng	0.00
44) 2-Fluorobiphenyl	9.19	172	183243	35.22	ng	0.00
78) Terphenyl-d14	12.94	244	138849	32.60	ng	0.00
Target Compounds						
49) Dimethylphthalate	9.58	163	16534	2.880	ng	Qvalue # 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF121117\
Data File : BF101309.D
Acq On : 12 Dec 2017 2:56
Operator : SJ/JU
Sample : I6744-03 2X
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PROS-3-E(7-8)

Quant Time: Dec 12 06:12:43 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF113017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 08 18:17:31 2017
Response via : Initial Calibration

