

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121120\
 Data File : BF122659.D
 Acq On : 11 Dec 2020 12:29
 Operator : JU/CG
 Sample : PB133535BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB133535BL

Quant Time: Dec 11 13:19:10 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 07 12:08:49 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.839	152	175121	20.00	ng	0.00
21) Naphthalene-d8	8.122	136	674684	20.00	ng	0.00
39) Acenaphthene-d10	9.874	164	355902	20.00	ng	0.00
64) Phenanthrene-d10	11.363	188	679873	20.00	ng	0.00
76) Chrysene-d12	14.004	240	627786	20.00	ng	0.00
86) Perylene-d12	15.468	264	583541	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.469	112	1324419	119.73	ng	0.01
7) Phenol-d6	6.475	99	1639610	111.76	ng	0.00
23) Nitrobenzene-d5	7.404	82	1031057	76.57	ng	0.00
42) 2,4,6-Tribromophenol	10.669	330	544775	116.94	ng	0.00
45) 2-Fluorobiphenyl	9.198	172	1865644	70.90	ng	0.00
79) Terphenyl-d14	12.951	244	2153480	60.04	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121120\
Data File : BF122659.D
Acq On : 11 Dec 2020 12:29
Operator : JU/CG
Sample : PB133535BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB133535BL

Quant Time: Dec 11 13:19:10 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 07 12:08:49 2020
Response via : Initial Calibration

