

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121124\
 Data File : BF140829.D
 Acq On : 11 Dec 2024 15:15
 Operator : RC/JU
 Sample : P5192-02
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 EFF-WASTE WATER

Quant Time: Dec 11 15:40:43 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 21 15:23:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.857	152	77371	20.000	ng	-0.02
21) Naphthalene-d8	8.139	136	256384	20.000	ng	#-0.02
39) Acenaphthene-d10	9.898	164	159565	20.000	ng	-0.02
64) Phenanthrene-d10	11.398	188	212898	20.000	ng	0.00
76) Chrysene-d12	14.062	240	135537	20.000	ng	0.01
86) Perylene-d12	15.551	264	146438	20.000	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.516	112	237279	52.325	ng	0.02
7) Phenol-d6	6.522	99	203171	33.890	ng	0.00
23) Nitrobenzene-d5	7.428	82	555873	110.900	ng	-0.01
42) 2,4,6-Tribromophenol	10.704	330	138584	81.207	ng	0.00
45) 2-Fluorobiphenyl	9.216	172	1049742	98.020	ng	-0.02
79) Terphenyl-d14	13.045	244	834246	95.843	ng	0.05
Target Compounds						
						Qvalue
15) Benzyl Alcohol	7.022	79	33431	7.520	ng	98
22) Acetophenone	7.275	105	32091	5.134	ng	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121124\
Data File : BF140829.D
Acq On : 11 Dec 2024 15:15
Operator : RC/JU
Sample : P5192-02
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
EFF-WASTE WATER

Quant Time: Dec 11 15:40:43 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 21 15:23:48 2024
Response via : Initial Calibration

