


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121218\
 Data File : BF111312.D
 Acq On : 12 Dec 2018 9:18
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF121218.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Dec 12 13:16:52 2018



AutoFind: Scans 1613, 1614, 1615; Background Corrected with Scan 1608

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	49.3	167853	PASS
68	69	0.00	2	1.4	1897	PASS
69	198	0.00	100	40.9	139301	PASS
70	69	0.00	2	0.6	790	PASS
127	198	10	80	52.5	178792	PASS
197	198	0.00	2	0.7	2340	PASS
198	198	100	100	100.0	340394	PASS
199	198	5	9	6.6	22372	PASS
275	198	10	60	22.8	77696	PASS
365	198	1	100	2.2	7613	PASS
441	198	0.01	100	13.6	46279	PASS
442	198	50	100	86.1	292912	PASS
443	442	15	24	20.3	59360	PASS