

Data Path : Z:\HPCHEM1\BNA F\DATA\BF121315\
 Data File : BF083744.D
 Acq On : 13 Dec 2015 22:12
 Operator : UM/IZ
 Sample : G4743-01
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 NORTHWALL-8

Quant Time: Dec 14 03:28:13 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF121315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 14 01:30:12 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.92	152	129510	20.00	ng	0.00
21) Naphthalene-d8	8.20	136	525320	20.00	ng	-0.01
38) Acenaphthene-d10	9.95	164	241319	20.00	ng	-0.01
63) Phenanthrene-d10	11.44	188	442888	20.00	ng	-0.01
75) Chrysene-d12	14.05	240	302223	20.00	ng	-0.02
86) Perylene-d12	15.49	264	250061	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.53	112	1024152	127.57	ng	0.00
7) Phenol-d6	6.54	99	1296802	127.56	ng	0.00
23) Nitrobenzene-d5	7.48	82	928677	98.90	ng	0.00
41) 2,4,6-Tribromophenol	10.74	330	329491	133.99	ng	-0.01
44) 2-Fluorobiphenyl	9.28	172	1319259	78.17	ng	-0.01
78) Terphenyl-d14	13.01	244	1183134	84.21	ng	-0.01
Target Compounds						Qvalue
49) Dimethylphthalate	9.66	163	421763	22.42	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF121315\
Data File : BF083744.D
Acq On : 13 Dec 2015 22:12
Operator : UM/IZ
Sample : G4743-01
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
NORTHWALL-8

Quant Time: Dec 14 03:28:13 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF121315.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 14 01:30:12 2015
Response via : Initial Calibration

