

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121322\
 Data File : BF131620.D
 Acq On : 13 Dec 2022 19:21
 Operator : CG\JU
 Sample : N5964-08
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SU-518-COMP-04

Quant Time: Dec 14 03:35:03 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF120822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 14 03:22:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.881	152	54076	20.000	ng	0.00
21) Naphthalene-d8	8.169	136	192942	20.000	ng	# 0.00
39) Acenaphthene-d10	9.933	164	112030	20.000	ng	0.00
64) Phenanthrene-d10	11.427	188	219944	20.000	ng	0.00
76) Chrysene-d12	14.080	240	160473	20.000	ng	# 0.00
86) Perylene-d12	15.586	264	113570	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.510	112	459970	141.904	ng	0.01
7) Phenol-d6	6.516	99	556520	131.019	ng	0.00
23) Nitrobenzene-d5	7.451	82	447669	87.720	ng	0.00
42) 2,4,6-Tribromophenol	10.733	330	198340	159.349	ng	0.00
45) 2-Fluorobiphenyl	9.251	172	770063	86.669	ng	0.00
79) Terphenyl-d14	13.027	244	956133	81.711	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121322\
Data File : BF131620.D
Acq On : 13 Dec 2022 19:21
Operator : CG\JU
Sample : N5964-08
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SU-518-COMP-04

Quant Time: Dec 14 03:35:03 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF120822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 14 03:22:29 2022
Response via : Initial Calibration

