

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121322\
 Data File : BF131624.D
 Acq On : 13 Dec 2022 21:28
 Operator : CG\JU
 Sample : N6007-01
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MOO-CONCRETE-1212

Quant Time: Dec 14 03:35:56 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF120822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 14 03:22:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.881	152	52586	20.000	ng	0.00
21) Naphthalene-d8	8.169	136	182541	20.000	ng	# 0.00
39) Acenaphthene-d10	9.933	164	91818	20.000	ng	0.00
64) Phenanthrene-d10	11.439	188	142213	20.000	ng	# 0.00
76) Chrysene-d12	14.110	240	119191	20.000	ng	# 0.02
86) Perylene-d12	15.598	264	123137	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.498	112	313720	99.527	ng	0.00
7) Phenol-d6	6.516	99	482157	116.728	ng	0.00
23) Nitrobenzene-d5	7.445	82	330090	68.366	ng	-0.01
42) 2,4,6-Tribromophenol	10.733	330	61031	59.827	ng	0.00
45) 2-Fluorobiphenyl	9.251	172	553085	75.951	ng	0.00
79) Terphenyl-d14	13.057	244	539345	62.057	ng	0.03

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121322\
Data File : BF131624.D
Acq On : 13 Dec 2022 21:28
Operator : CG\JU
Sample : N6007-01
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MOO-CONCRETE-1212

Quant Time: Dec 14 03:35:56 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF120822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 14 03:22:29 2022
Response via : Initial Calibration

