

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA F\Data\BF121417\
 Data File : BF101395.D
 Acq On : 14 Dec 2017 15:24
 Operator : SJ/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF121417

Quant Time: Dec 14 16:00:17 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF121417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 14 15:42:32 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2	1,4-Dioxane	0.690	0.672	2.6	99	0.00
3	Pyridine	1.917	1.947	-1.6	103	0.00
4	n-Nitrosodimethylamine	0.883	0.893	-1.1	101	0.00
5 S	2-Fluorophenol	1.343	1.334	0.7	102	0.00
6	Aniline	2.322	2.234	3.8	100	0.00
7 S	Phenol-d6	1.671	1.582	5.3	101	0.00
8	2-Chlorophenol	1.412	1.347	4.6	101	0.00
9	Benzaldehyde	1.234	1.212	1.8	102	0.00
10 C	Phenol	1.905	1.865	2.1	103	0.00
11	bis(2-Chloroethyl)ether	1.494	1.462	2.1	101	0.00
12	1,3-Dichlorobenzene	1.594	1.541	3.3	101	0.00
13 C	1,4-Dichlorobenzene	1.628	1.559	4.2	101	0.00
14	1,2-Dichlorobenzene	1.465	1.418	3.2	101	0.00
15	Benzyl Alcohol	1.150	1.088	5.4	100	0.00
16	2,2'-oxybis(1-Chloropropane	2.286	2.279	0.3	102	0.00
17	2-Methylphenol	1.299	1.217	6.3	102	0.00
18	Hexachloroethane	0.566	0.547	3.4	101	0.00
19 P	n-Nitroso-di-n-propylamine	1.097	1.000	8.8	101	0.00
20	3+4-Methylphenols	1.377	1.347	2.2	102	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
22	Acetophenone	0.456	0.433	5.0	100	0.00
23 S	Nitrobenzene-d5	0.359	0.355	1.1	102	0.00
24	Nitrobenzene	0.398	0.388	2.5	103	0.00
25	Isophorone	0.721	0.688	4.6	102	0.00
26 C	2-Nitrophenol	0.171	0.177	-3.5	103	0.00
27	2,4-Dimethylphenol	0.290	0.279	3.8	102	0.00
28	bis(2-Chloroethoxy)methane	0.458	0.442	3.5	103	0.00
29 C	2,4-Dichlorophenol	0.287	0.279	2.8	101	0.00
30	1,2,4-Trichlorobenzene	0.303	0.287	5.3	100	0.00
31	Naphthalene	1.007	0.956	5.1	102	0.00
32	Benzoic acid	0.173	0.170	1.7	105	0.00
33	4-Chloroaniline	0.438	0.418	4.6	102	0.00
34 C	Hexachlorobutadiene	0.175	0.167	4.6	100	0.00
35	Caprolactam	0.100	0.096	4.0	100	0.00
36 C	4-Chloro-3-methylphenol	0.315	0.303	3.8	100	0.00
37	2-Methylnaphthalene	0.656	0.615	6.3	101	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
39	1,2,4,5-Tetrachlorobenzene	0.675	0.666	1.3	103	0.00
40 P	Hexachlorocyclopentadiene	0.088	0.097	-10.2	116	0.00
41 S	2,4,6-Tribromophenol	0.193	0.186	3.6	99	0.00
42 C	2,4,6-Trichlorophenol	0.407	0.403	1.0	99	0.00
43	2,4,5-Trichlorophenol	0.445	0.446	-0.2	101	0.00
44 S	2-Fluorobiphenyl	1.289	1.256	2.6	101	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA F\Data\BF121417\
 Data File : BF101395.D
 Acq On : 14 Dec 2017 15:24
 Operator : SJ/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF121417

Quant Time: Dec 14 16:00:17 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF121417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 14 15:42:32 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.774	1.680	5.3	101	0.00
46	2-Chloronaphthalene	1.357	1.290	4.9	101	0.00
47	2-Nitroaniline	0.469	0.476	-1.5	103	0.00
48	Acenaphthylene	2.144	2.024	5.6	101	0.00
49	Dimethylphthalate	1.609	1.511	6.1	100	0.00
50	2,6-Dinitrotoluene	0.327	0.327	0.0	100	0.00
51 C	Acenaphthene	1.288	1.222	5.1	102	0.00
52	3-Nitroaniline	0.408	0.407	0.2	100	0.00
53 P	2,4-Dinitrophenol	0.087	0.089	-2.3	103	0.00
54	Dibenzofuran	1.637	1.601	2.2	101	0.00
55 P	4-Nitrophenol	0.178	0.185	-3.9	101	0.00
56	2,4-Dinitrotoluene	0.368	0.374	-1.6	103	0.00
57	Fluorene	1.385	1.359	1.9	101	0.00
58	2,3,4,6-Tetrachlorophenol	0.320	0.323	-0.9	102	0.00
59	Diethylphthalate	1.593	1.491	6.4	102	0.00
60	4-Chlorophenyl-phenylether	0.664	0.659	0.8	101	0.00
61	4-Nitroaniline	0.410	0.406	1.0	102	0.00
62	Azobenzene	1.650	1.544	6.4	101	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
64	4,6-Dinitro-2-methylphenol	0.102	0.110	-7.8	104	0.00
65 c	n-Nitrosodiphenylamine	0.738	0.703	4.7	100	0.00
66	4-Bromophenyl-phenylether	0.225	0.217	3.6	100	0.00
67	Hexachlorobenzene	0.238	0.230	3.4	101	0.00
68	Atrazine	0.220	0.214	2.7	100	0.00
69 C	Pentachlorophenol	0.079	0.077	2.5	98	0.00
70	Phenanthrene	1.112	1.049	5.7	101	0.00
71	Anthracene	1.136	1.077	5.2	101	0.00
72	Carbazole	1.064	1.013	4.8	101	0.00
73	Di-n-butylphthalate	1.291	1.232	4.6	102	0.00
74 C	Fluoranthene	1.167	1.105	5.3	101	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	103	0.00
76	Benzidine	0.845	0.824	2.5	101	0.00
77	Pyrene	1.501	1.430	4.7	101	0.00
78 S	Terphenyl-d14	0.830	0.761	8.3	101	0.00
79	Butylbenzylphthalate	0.679	0.643	5.3	98	0.00
80	Benzo(a)anthracene	1.321	1.253	5.1	101	0.00
81	3,3'-Dichlorobenzidine	0.431	0.417	3.2	100	0.00
82	Chrysene	1.247	1.173	5.9	100	0.00
83	Bis(2-ethylhexyl)phthalate	0.860	0.818	4.9	100	0.00
84 c	Di-n-octyl phthalate	1.534	1.493	2.7	103	0.00
85	Indeno(1,2,3-cd)pyrene	1.110	1.009	9.1	99	0.00
86 I	Perylene-d12	1.000	1.000	0.0	102	0.00
87	Benzo(b)fluoranthene	1.219	1.171	3.9	98	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA F\Data\BF121417\
Data File : BF101395.D
Acq On : 14 Dec 2017 15:24
Operator : SJ/JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF121417

Quant Time: Dec 14 16:00:17 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF121417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 14 15:42:32 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.185	1.121	5.4	102	0.00
89 C	Benzo(a)pyrene	1.124	1.071	4.7	100	0.00
90	Dibenzo(a,h)anthracene	0.965	0.886	8.2	99	0.00
91	Benzo(a,h,i)perylene	0.955	0.890	6.8	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0