


Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA F\Data\BF121417\
 Data File : BF101386.D
 Acq On : 14 Dec 2017 10:36
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF121417.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Dec 14 15:42:32 2017



AutoFind: Scans 1614, 1615, 1616; Background Corrected with Scan 1609

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	45.9	48712	PASS
68	69	0.00	2	2.0	792	PASS
69	198	0.00	100	37.7	40013	PASS
70	69	0.00	2	0.7	286	PASS
127	198	10	80	47.3	50114	PASS
197	198	0.00	2	0.6	661	PASS
198	198	100	100	100.0	106029	PASS
199	198	5	9	6.6	7008	PASS
275	198	10	60	23.7	25083	PASS
365	198	1	100	2.7	2898	PASS
441	443	0.01	100	80.8	12783	PASS
442	198	50	100	74.0	78480	PASS
443	442	15	24	20.2	15818	PASS