

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121420\
 Data File : BF122715.D
 Acq On : 14 Dec 2020 20:02
 Operator : JU/CG
 Sample : L5035-05DL 5X
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DUPDL

Manual Integrations
 APPROVED

mohammad
 12/17/2020 9:47:06 AM

Quant Time: Dec 14 23:54:47 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 07 12:08:49 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.834	152	156935	20.00	ng	0.00
21) Naphthalene-d8	8.122	136	608834	20.00	ng	0.00
39) Acenaphthene-d10	9.875	164	317264	20.00	ng	0.00
64) Phenanthrene-d10	11.363	188	532853	20.00	ng	0.00
76) Chrysene-d12	14.010	240	364884	20.00	ng	0.00
86) Perylene-d12	15.474	264	397413	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.445	112	198734	20.05	ng	-0.01
7) Phenol-d6	6.463	99	264487	20.12	ng	-0.02
23) Nitrobenzene-d5	7.392	82	139949	11.52	ng	-0.02
42) 2,4,6-Tribromophenol	10.663	330	55631	13.40	ng	-0.01
45) 2-Fluorobiphenyl	9.192	172	300460	12.81	ng	-0.01
79) Terphenyl-d14	12.945	244	238495	11.44	ng	-0.01
Target Compounds						Qvalue
49) Acenaphthylene	9.733	152	375151	11.64	ng	98
58) Fluorene	10.422	166	84203	3.61	ng	99
71) Phenanthrene	11.392	178	1379904	43.58	ng	99
72) Anthracene	11.439	178	303788	9.67	ng	98
73) Carbazole	11.592	167	85174	2.99	ng	100
75) Fluoranthene	12.586	202	1944627	58.56	ng	99
78) Pyrene	12.810	202	1837818	65.14	ng	100
81) Benzo(a)anthracene	13.998	228	874900m	35.88	ng	
83) Chrysene	14.033	228	948377	39.08	ng	98
87) Indeno(1,2,3-cd)pyrene	16.933	276	502296	19.25	ng	95
88) Benzo(b)fluoranthene	15.051	252	1190173m	42.68	ng	
89) Benzo(k)fluoranthene	15.074	252	305385m	11.98	ng	
90) Benzo(a)pyrene	15.415	252	787014	32.26	ng	98
91) Dibenzo(a,h)anthracene	16.939	278	130058	6.09	ng	95
92) Benzo(g,h,i)perylene	17.374	276	453146	21.93	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121420\
Data File : BF122715.D
Acq On : 14 Dec 2020 20:02
Operator : JU/CG
Sample : L5035-05DL 5X
Misc :
ALS Vial : 22 Sample Multiplier: 1


Instrument :
BNA_F
ClientSampleId :
DUPDL

Manual Integrations
APPROVED

mohammad
12/17/2020 9:47:06 AM

Quant Time: Dec 14 23:54:47 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 07 12:08:49 2020
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 6.834 min Scan# 807
Delta R.T. -0.006 min
Lab File: BF122715.D
Acq: 14 Dec 2020 20:02

Tgt Ion:152 Resp: 156935
Ion Ratio Lower Upper
152 100
150 157.4 125.8 188.6
115 61.7 49.3 73.9

Instrument :
BNA_F
ClientSampleId :
DUPDL




Manual Integrations
APPROVED


mohammad
12/17/2020 9:47:06 AM



#5
2-Fluorophenol
Concen: 20.05 ng
RT: 5.445 min Scan# 571
Delta R.T. -0.012 min
Lab File: BF122715.D
Acq: 14 Dec 2020 20:02

Tgt Ion:112 Resp: 198734
Ion Ratio Lower Upper
112 100
64 53.9 46.9 70.3
63 27.9 24.5 36.7









#7
Phenol-d6
Concen: 20.12 ng
RT: 6.463 min Scan# 744
Delta R.T. -0.018 min
Lab File: BF122715.D
Acq: 14 Dec 2020 20:02

Tgt Ion: 99 Resp: 264487
Ion Ratio Lower Upper
99 100
42 14.9 13.8 20.6
71 32.1 26.9 40.3

Instrument :
BNA_F
ClientSampleId :
DUPDL




Manual Integrations
APPROVED


mohammad
12/17/2020 9:47:06 AM



#21
Naphthalene-d8
Concen: 20.00 ng
RT: 8.122 min Scan# 1026
Delta R.T. -0.006 min
Lab File: BF122715.D
Acq: 14 Dec 2020 20:02

Tgt Ion: 136 Resp: 608834
Ion Ratio Lower Upper
136 100
137 10.7 8.8 13.2
54 6.7 5.5 8.3
68 5.5 4.6 6.8









#23
Nitrobenzene-d5
Concen: 11.52 ng
RT: 7.392 min Scan# 902
Delta R.T. -0.018 min
Lab File: BF122715.D
Acq: 14 Dec 2020 20:02

Tgt Ion: 82 Resp: 139949
Ion Ratio Lower Upper
82 100
128 43.6 37.0 55.4
54 45.4 38.4 57.6

Instrument :
BNA_F
ClientSampleId :
DUPDL




Manual Integrations
APPROVED


mohammad
12/17/2020 9:47:06 AM



#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 9.875 min Scan# 1324
Delta R.T. -0.006 min
Lab File: BF122715.D
Acq: 14 Dec 2020 20:02


Tgt Ion:164 Resp: 317264
Ion Ratio Lower Upper
164 100
162 103.2 80.8 121.2
160 45.0 35.4 53.0





#42
2,4,6-Tribromophenol
Concen: 13.40 ng
RT: 10.663 min Scan# 1458
Delta R.T. -0.012 min
Lab File: BF122715.D
Acq: 14 Dec 2020 20:02




Instrument :
BNA_F
ClientSampleId :
DUPDL




Tgt Ion:330 Resp: 55631
Ion Ratio Lower Upper
330 100
332 95.2 77.5 116.3
141 33.8 23.8 35.8

Manual Integrations
APPROVED



mohammad
12/17/2020 9:47:06 AM




#45
2-Fluorobiphenyl
Concen: 12.81 ng
RT: 9.192 min Scan# 1208
Delta R.T. -0.012 min
Lab File: BF122715.D
Acq: 14 Dec 2020 20:02




Tgt Ion:172 Resp: 300460
Ion Ratio Lower Upper
172 100
171 35.2 29.6 44.4
170 22.9 19.5 29.3









#49
Acenaphthylene
Concen: 11.64 ng
RT: 9.733 min Scan# 1300
Delta R.T. -0.012 min
Lab File: BF122715.D
Acq: 14 Dec 2020 20:02





Tgt Ion:152 Resp: 375151
Ion Ratio Lower Upper
152 100
151 19.7 16.6 25.0
153 13.1 10.9 16.3



#58
Fluorene
Concen: 3.61 ng
RT: 10.422 min Scan# 1417
Delta R.T. -0.011 min
Lab File: BF122715.D
Acq: 14 Dec 2020 20:02




Tgt Ion:166 Resp: 84203
Ion Ratio Lower Upper
166 100
165 98.3 79.4 119.0
167 14.6 10.9 16.3




Instrument :
BNA_F
ClientSampleId :
DUPDL

Manual Integrations
APPROVED




mohammad
12/17/2020 9:47:06 AM




#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 11.363 min Scan# 1577
Delta R.T. -0.006 min
Lab File: BF122715.D
Acq: 14 Dec 2020 20:02





Tgt Ion:188 Resp: 532853
Ion Ratio Lower Upper
188 100
94 9.4 7.4 11.2
80 9.8 8.2 12.4



#71
Phenanthrene
Concen: 43.58 ng
RT: 11.392 min Scan# 1582
Delta R.T. -0.000 min
Lab File: BF122715.D
Acq: 14 Dec 2020 20:02




Tgt Ion:178 Resp: 1379904
Ion Ratio Lower Upper
178 100
176 19.5 16.2 24.2
179 15.5 12.7 19.1



Instrument :
BNA_F
ClientSampleId :
DUPDL


Manual Integrations
APPROVED

mohammad
12/17/2020 9:47:06 AM



#72
 Anthracene
 Concen: 9.67 ng
 RT: 11.439 min Scan# 1590
 Delta R.T. -0.006 min
 Lab File: BF122715.D
 Acq: 14 Dec 2020 20:02




Instrument :
 BNA_F
 ClientSampled :
 DUPDL




Tgt Ion:178 Resp: 303788
 Ion Ratio Lower Upper
 178 100
 176 18.2 15.6 23.4
 179 15.4 12.8 19.2

Manual Integrations
 APPROVED



mohammad
 12/17/2020 9:47:06 AM




#73
 Carbazole
 Concen: 2.99 ng
 RT: 11.592 min Scan# 1616
 Delta R.T. -0.006 min
 Lab File: BF122715.D
 Acq: 14 Dec 2020 20:02



Tgt Ion:167 Resp: 85174
 Ion Ratio Lower Upper
 167 100
 166 22.6 18.0 27.0
 139 13.9 10.9 16.3









#75
Fluoranthene
Concen: 58.56 ng
RT: 12.586 min Scan# 1785
Delta R.T. 0.006 min
Lab File: BF122715.D
Acq: 14 Dec 2020 20:02

Tgt Ion:202 Resp: 1944627
Ion Ratio Lower Upper
202 100
101 11.2 0.0 32.2
203 18.3 0.0 38.1

Instrument :
BNA_F
ClientSampleId :
DUPDL




Manual Integrations
APPROVED


mohammad
12/17/2020 9:47:06 AM




#76
Chrysene-d12
Concen: 20.00 ng
RT: 14.010 min Scan# 2027
Delta R.T. -0.000 min
Lab File: BF122715.D
Acq: 14 Dec 2020 20:02

Tgt Ion:240 Resp: 364884
Ion Ratio Lower Upper
240 100
120 9.2 8.6 12.8
236 25.2 20.3 30.5








#78
Pyrene
Concen: 65.14 ng
RT: 12.810 min Scan# 1823
Delta R.T. -0.000 min
Lab File: BF122715.D
Acq: 14 Dec 2020 20:02




Tgt Ion:202 Resp: 1837818

Ion	Ratio	Lower	Upper
202	100		
200	21.8	17.6	26.4
203	18.3	14.6	21.8





#79
Terphenyl-d14
Concen: 11.44 ng
RT: 12.945 min Scan# 1846
Delta R.T. -0.012 min
Lab File: BF122715.D
Acq: 14 Dec 2020 20:02



Tgt Ion:244 Resp: 238495


Ion	Ratio	Lower	Upper
244	100		
212	6.5	5.7	8.5
122	10.0	8.3	12.5



Instrument :
BNA_F
ClientSampleId :
DUPDL

Manual Integrations
APPROVED

mohammad
12/17/2020 9:47:06 AM







#81
Benzo(a)anthracene
Concen: 35.88 ng m
RT: 13.998 min Scan# 2025
Delta R.T. -0.000 min
Lab File: BF122715.D
Acq: 14 Dec 2020 20:02

Tgt Ion:228 Resp: 874900
Ion Ratio Lower Upper
228 100
226 41.4 23.0 34.4#
229 21.0 16.6 24.8

Instrument :
BNA_F
ClientSampleId :
DUPDL




Manual Integrations
APPROVED


mohammad
12/17/2020 9:47:06 AM



#83
Chrysene
Concen: 39.08 ng
RT: 14.033 min Scan# 2031
Delta R.T. -0.006 min
Lab File: BF122715.D
Acq: 14 Dec 2020 20:02


Tgt Ion:228 Resp: 948377
Ion Ratio Lower Upper
228 100
226 30.9 24.6 36.8
229 22.2 15.8 23.6





#86
Perylene-d12
Concen: 20.00 ng
RT: 15.474 min Scan# 2276
Delta R.T. 0.006 min
Lab File: BF122715.D
Acq: 14 Dec 2020 20:02




Instrument :
BNA_F
ClientSampleId :
DUPDL




Tgt Ion:264 Resp: 397413
Ion Ratio Lower Upper
264 100
260 24.6 18.6 28.0
265 23.6 17.0 25.6

Manual Integrations
APPROVED



mohammad
12/17/2020 9:47:06 AM




#87
Indeno(1,2,3-cd)pyrene
Concen: 19.25 ng
RT: 16.933 min Scan# 2524
Delta R.T. -0.000 min
Lab File: BF122715.D
Acq: 14 Dec 2020 20:02



Tgt Ion:276 Resp: 502296
Ion Ratio Lower Upper
276 100
138 19.5 18.7 28.1
277 24.5 20.2 30.4









#88
Benzo(b)fluoranthene
Concen: 42.68 ng m
RT: 15.051 min Scan# 2204
Delta R.T. 0.006 min
Lab File: BF122715.D
Acq: 14 Dec 2020 20:02

Instrument :
BNA_F
ClientSampled :
DUPDL

Tgt Ion:252 Resp: 1190173
Ion Ratio Lower Upper
252 100
253 22.7 18.3 27.5
125 11.3 8.1 12.1




Manual Integrations
APPROVED


mohammad
12/17/2020 9:47:06 AM



#89
Benzo(k)fluoranthene
Concen: 11.98 ng m
RT: 15.074 min Scan# 2208
Delta R.T. -0.000 min
Lab File: BF122715.D
Acq: 14 Dec 2020 20:02

Tgt Ion:252 Resp: 305385
Ion Ratio Lower Upper
252 100
253 24.0 17.8 26.8
125 9.8 8.4 12.6









#90
Benzo(a)pyrene
Concen: 32.26 ng
RT: 15.415 min Scan# 2266
Delta R.T. 0.006 min
Lab File: BF122715.D
Acq: 14 Dec 2020 20:02

Tgt Ion:252 Resp: 787014
Ion Ratio Lower Upper
252 100
253 23.0 17.5 26.3
125 10.6 8.6 12.8

Instrument :
BNA_F
ClientSampleId :
DUPDL




Manual Integrations
APPROVED


mohammad
12/17/2020 9:47:06 AM



#91
Dibenzo(a,h)anthracene
Concen: 6.09 ng
RT: 16.939 min Scan# 2525
Delta R.T. -0.006 min
Lab File: BF122715.D
Acq: 14 Dec 2020 20:02


Tgt Ion:278 Resp: 130058
Ion Ratio Lower Upper
278 100
139 17.2 12.0 18.0
279 27.0 19.4 29.2





#92
Benzo(g,h,i)perylene
Concen: 21.93 ng
RT: 17.374 min Scan# 2599
Delta R.T. 0.006 min
Lab File: BF122715.D
Acq: 14 Dec 2020 20:02

Instrument :
BNA_F
ClientSampleId :
DUPDL



Tgt Ion: 276 Resp: 453146
Ion Ratio Lower Upper
276 100
277 23.7 19.3 28.9
138 20.9 16.6 24.8

Manual Integrations
APPROVED

mohammad
12/17/2020 9:47:06 AM

