


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121420\
 Data File : BF122694.D
 Acq On : 14 Dec 2020 8:38
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Dec 15 18:14:23 2020



AutoFind: Scans 1619, 1620, 1621; Background Corrected with Scan 1614

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	31.2	26741	PASS
68	69	0.00	2	1.4	399	PASS
69	198	0.00	100	32.5	27848	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	46.1	39496	PASS
197	198	0.00	2	0.3	283	PASS
198	198	100	100	100.0	85736	PASS
199	198	5	9	6.8	5843	PASS
275	198	10	60	29.3	25123	PASS
365	198	1	100	3.5	2989	PASS
441	198	0.01	100	19.9	17104	PASS
442	442	100	100	100.0	109371	PASS
443	442	15	24	18.8	20521	PASS