

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF121515\
 Data File : BF083838.D
 Acq On : 16 Dec 2015 2:46
 Operator : UM/IZ
 Sample : G4739-01DL 10X
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 K205-5-10DL

Quant Time: Dec 16 04:34:28 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF121315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 14 01:30:12 2015
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.90	152	92635	20.00	ng	-0.02
21) Naphthalene-d8	8.19	136	363920	20.00	ng	-0.02
38) Acenaphthene-d10	9.94	164	162397	20.00	ng	-0.02
63) Phenanthrene-d10	11.42	188	266275	20.00	ng	-0.02
75) Chrysene-d12	14.05	240	204580	20.00	ng	-0.02
86) Perylene-d12	15.49	264	158072	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.49	112	64181	11.18	ng	-0.03
7) Phenol-d6	6.52	99	87429	12.02	ng	-0.02
23) Nitrobenzene-d5	7.46	82	49174	7.56	ng	-0.02
41) 2,4,6-Tribromophenol	10.73	330	15891	9.60	ng	-0.02
44) 2-Fluorobiphenyl	9.26	172	104424	8.16	ng	-0.02
78) Terphenyl-d14	13.00	244	80774	8.49	ng	-0.02
Target Compounds						Qvalue
3) Pyridine	3.29	79	18466	2.64	ng	# 24
18) Hexachloroethane	7.41	117	20771	8.08	ng	# 1
25) Isophorone	7.46	82	38575	3.02	ng	# 69
31) Naphthalene	8.20	128	53290	2.81	ng	# 96
37) 2-Methylnaphthalene	9.00	142	67167	5.64	ng	93
48) Acenaphthylene	9.80	152	136358	7.63	ng	# 97
53) 2,4-Dinitrophenol	9.88	184	292	2.67	ng	# 1
57) Fluorene	10.49	166	119477	10.60	ng	99
70) Phenanthrene	11.45	178	548099	38.75	ng	99
71) Anthracene	11.49	178	233800	15.91	ng	100
74) Fluoranthene	12.64	202	281773	19.28	ng	95
77) Pyrene	12.87	202	446173	27.59	ng	99
80) Benzo(a)anthracene	14.04	228	236699	20.35	ng	95
82) Chrysene	14.04	228	236699	20.86	ng	98
85) Indeno(1,2,3-cd)pyrene	16.92	276	23270	2.09	ng	# 100
87) Benzo(b)fluoranthene	15.08	252	105846	10.50	ng	# 95
88) Benzo(k)fluoranthene	15.08	252	105846	11.58	ng	99
89) Benzo(a)pyrene	15.44	252	73310	8.15	ng	99
91) Benzo(g,h,i)perylene	17.35	276	27209	2.96	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF121515\
Data File : BF083838.D
Acq On : 16 Dec 2015 2:46
Operator : UM/IZ
Sample : G4739-01DL 10X
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
K205-5-10DL

Quant Time: Dec 16 04:34:28 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF121315.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 14 01:30:12 2015
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 20.00 ng

RT: 6.90 min Scan# 421

Delta R.T. -0.02 min

Lab File: BF083838.D

Acq: 16 Dec 2015 2:46

Instrument :

BNA_F

ClientSampleId :

K205-5-10DL


Tgt Ion:152 Resp: 92635

Ion Ratio Lower Upper

152 100

150 155.0 126.5 189.7




115 67.5 52.3 78.5



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#3

Pyridine

Concen: 2.64 ng

RT: 3.29 min Scan# 105

Delta R.T. -0.01 min

Lab File: BF083838.D

Acq: 16 Dec 2015 2:46


Tgt Ion: 79 Resp: 18466

Ion Ratio Lower Upper

79 100

52 0.0 63.7 95.5#



51 9.8 30.6 45.8#




Abundance Ion 79.00 (78.70 to 79.70): BF0

Ion 52.00 (51.70 to 52.70): BF0

Ion 51.00 (50.70 to 51.70): BF0





#5

2-Fluorophenol

Concen: 11.18 ng

RT: 5.49 min Scan# 298

Delta R.T. -0.03 min

Lab File: BF083838.D

Acq: 16 Dec 2015 2:46

Instrument :

BNA_F

ClientSampleId :

K205-5-10DL


Tgt Ion: 112 Resp: 64181

Ion Ratio Lower Upper

112 100

64 65.3 50.5 75.7




63 32.8 25.8 38.6



Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF0

Ion 63.00 (62.70 to 63.70): BF0



#7

Phenol-d6

Concen: 12.02 ng

RT: 6.52 min Scan# 388

Delta R.T. -0.02 min

Lab File: BF083838.D

Acq: 16 Dec 2015 2:46


Tgt Ion: 99 Resp: 87429

Ion Ratio Lower Upper

99 100

42 15.0 17.3 25.9#



71 33.6 26.6 40.0




Abundance Ion 99.00 (98.70 to 99.70): BF0

Ion 42.00 (41.70 to 42.70): BF0

Ion 71.00 (70.70 to 71.70): BF0





#18

Hexachloroethane

Concen: 8.08 ng

RT: 7.41 min Scan# 466

Delta R.T. -0.03 min

Lab File: BF083838.D

Acq: 16 Dec 2015 2:46

Instrument :

BNA_F

ClientSampleId :

K205-5-10DL


Tgt Ion:117 Resp: 20771

Ion Ratio Lower Upper

117 100

119 4.5 77.2 115.8#




201 0.0 86.8 130.2#



Abundance Ion 117.00 (116.70 to 117.70): E

Ion 119.00 (118.70 to 119.70): E

Ion 201.00 (200.70 to 201.70): E



#21

Naphthalene-d8

Concen: 20.00 ng

RT: 8.19 min Scan# 534

Delta R.T. -0.02 min

Lab File: BF083838.D

Acq: 16 Dec 2015 2:46

Tgt Ion:136 Resp: 363920


Ion Ratio Lower Upper

136 100

137 10.8 8.7 13.1

54 4.9 7.8 11.6#

68 4.4 5.7 8.5#






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF0

Ion 68.00 (67.70 to 68.70): BF0









#23
 Nitrobenzene-d5
 Concen: 7.56 ng
 RT: 7.46 min Scan# 470
 Delta R.T. -0.02 min
 Lab File: BF083838.D
 Acq: 16 Dec 2015 2:46




Instrument :
 BNA_F
 ClientSampleId :
 K205-5-10DL


Tgt Ion	Ratio	Lower	Upper
82	100		
128	51.3	30.7	46.1#
54	40.2	42.5	63.7#



#25
 Isophorone
 Concen: 3.02 ng
 RT: 7.46 min Scan# 470
 Delta R.T. -0.29 min
 Lab File: BF083838.D
 Acq: 16 Dec 2015 2:46

Tgt Ion	Ratio	Lower	Upper
82	100		
95	0.0	5.0	7.4#
138	0.0	12.1	18.1#









#31
Naphthalene
Concen: 2.81 ng
RT: 8.20 min Scan# 535
Delta R.T. -0.03 min
Lab File: BF083838.D
Acq: 16 Dec 2015 2:46




Instrument :
BNA_F
ClientSampleId :
K205-5-10DL


Tgt Ion:128 Resp: 53290
Ion Ratio Lower Upper
128 100
129 13.3 8.6 13.0#
127 13.8 10.6 16.0



#37
2-Methylnaphthalene
Concen: 5.64 ng
RT: 9.00 min Scan# 605
Delta R.T. 0.08 min
Lab File: BF083838.D
Acq: 16 Dec 2015 2:46

Tgt Ion:142 Resp: 67167
Ion Ratio Lower Upper
142 100
141 93.9 70.3 105.5
115 41.9 29.2 43.8









#38
Acenaphthene-d10
Concen: 20.00 ng
RT: 9.94 min Scan# 687
Delta R.T. -0.02 min
Lab File: BF083838.D
Acq: 16 Dec 2015 2:46

Instrument :
BNA_F
ClientSampleId :
K205-5-10DL

Tgt Ion:164 Resp: 162397
Ion Ratio Lower Upper
164 100
162 104.6 78.8 118.2
160 46.4 33.4 50.2




Abundance Ion 164.00 (163.70 to 164.70): E
Ion 162.00 (161.70 to 162.70): E
Ion 160.00 (159.70 to 160.70): E






#41
2,4,6-Tribromophenol
Concen: 9.60 ng
RT: 10.73 min Scan# 756
Delta R.T. -0.02 min
Lab File: BF083838.D
Acq: 16 Dec 2015 2:46

Tgt Ion:330 Resp: 15891
Ion Ratio Lower Upper
330 100
332 97.0 0.0 0.0#
141 44.2 0.0 0.0#



Abundance Ion 329.80 (329.50 to 330.50): E
Ion 331.80 (331.50 to 332.50): E
Ion 141.00 (140.70 to 141.70): E









#44
 2-Fluorobiphenyl
 Concen: 8.16 ng
 RT: 9.26 min Scan# 628
 Delta R.T. -0.02 min
 Lab File: BF083838.D
 Acq: 16 Dec 2015 2:46

Instrument :
 BNA_F
 ClientSampleId :
 K205-5-10DL

Tgt Ion:172 Resp: 104424
 Ion Ratio Lower Upper
 172 100
 171 34.6 28.6 42.8
 170 22.9 19.4 29.0




Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E






#48
 Acenaphthylene
 Concen: 7.63 ng
 RT: 9.80 min Scan# 675
 Delta R.T. -0.02 min
 Lab File: BF083838.D
 Acq: 16 Dec 2015 2:46

Tgt Ion:152 Resp: 136358
 Ion Ratio Lower Upper
 152 100
 151 20.4 16.3 24.5
 153 16.5 10.8 16.2#



Abundance Ion 152.00 (151.70 to 152.70): E
 Ion 151.00 (150.70 to 151.70): E
 Ion 153.00 (152.70 to 153.70): E









#53
 2,4-Dinitrophenol
 Concen: 2.67 ng
 RT: 9.88 min Scan# 682
 Delta R.T. -0.13 min
 Lab File: BF083838.D
 Acq: 16 Dec 2015 2:46




Instrument :
 BNA_F
 ClientSampleId :
 K205-5-10DL


Tgt Ion:184 Resp: 292
 Ion Ratio Lower Upper
 184 100
 63 511.7 62.4 93.6#
 154 0.0 46.6 69.8#



#57
 Fluorene
 Concen: 10.60 ng
 RT: 10.49 min Scan# 735
 Delta R.T. -0.02 min
 Lab File: BF083838.D
 Acq: 16 Dec 2015 2:46

Tgt Ion:166 Resp: 119477
 Ion Ratio Lower Upper
 166 100
 165 100.5 80.0 120.0
 167 14.2 10.6 16.0





#63

Phenanthrene-d10

Concen: 20.00 ng

RT: 11.42 min Scan# 817

Delta R.T. -0.02 min

Lab File: BF083838.D

Acq: 16 Dec 2015 2:46

Instrument :

BNA_F

ClientSampleId :

K205-5-10DL


Tgt Ion:188 Resp: 266275

Ion Ratio Lower Upper

188 100

94 9.2 9.4 14.2#




80 9.8 10.6 16.0#



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BF0

Ion 80.00 (79.70 to 80.70): BF0



#70

Phenanthrene

Concen: 38.75 ng

RT: 11.45 min Scan# 819

Delta R.T. -0.02 min

Lab File: BF083838.D

Acq: 16 Dec 2015 2:46


Tgt Ion:178 Resp: 548099

Ion Ratio Lower Upper

178 100

176 19.4 15.7 23.5



179 15.8 12.2 18.4




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E





#71

Anthracene

Concen: 15.91 ng

RT: 11.49 min Scan# 823

Delta R.T. -0.02 min

Lab File: BF083838.D

Acq: 16 Dec 2015 2:46

Instrument :

BNA_F

ClientSampleId :

K205-5-10DL


Tgt Ion:178 Resp: 233800

Ion Ratio Lower Upper

178 100

176 18.9 15.2 22.8




179 16.0 12.6 18.8



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#74

Fluoranthene

Concen: 19.28 ng

RT: 12.64 min Scan# 923

Delta R.T. -0.01 min

Lab File: BF083838.D

Acq: 16 Dec 2015 2:46


Tgt Ion:202 Resp: 281773

Ion Ratio Lower Upper

202 100

101 14.4 0.0 31.5



203 18.8 0.0 37.6




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E





#75

Chrysene-d12

Concen: 20.00 ng

RT: 14.05 min Scan# 1047

Delta R.T. -0.02 min

Lab File: BF083838.D

Acq: 16 Dec 2015 2:46

Instrument :

BNA_F

ClientSampleId :

K205-5-10DL


Tgt Ion:240 Resp: 204580

Ion Ratio Lower Upper

240 100

120 16.6 8.1 12.1#




236 25.4 20.8 31.2



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#77

Pyrene

Concen: 27.59 ng

RT: 12.87 min Scan# 943

Delta R.T. -0.01 min

Lab File: BF083838.D

Acq: 16 Dec 2015 2:46


Tgt Ion:202 Resp: 446173

Ion Ratio Lower Upper

202 100

200 21.3 16.7 25.1



203 18.4 14.3 21.5




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E









#78
 Terphenyl-d14
 Concen: 8.49 ng
 RT: 13.00 min Scan# 955
 Delta R.T. -0.02 min
 Lab File: BF083838.D
 Acq: 16 Dec 2015 2:46




Instrument :
 BNA_F
 ClientSampleId :
 K205-5-10DL


Tgt Ion: 244 Resp: 80774
 Ion Ratio Lower Upper
 244 100
 212 7.2 6.1 9.1
 122 14.0 11.0 16.6



#80
 Benzo(a)anthracene
 Concen: 20.35 ng
 RT: 14.04 min Scan# 1046
 Delta R.T. -0.02 min
 Lab File: BF083838.D
 Acq: 16 Dec 2015 2:46

Tgt Ion: 228 Resp: 236699
 Ion Ratio Lower Upper
 228 100
 226 30.6 22.3 33.5
 229 22.0 15.7 23.5





#82

Chrysene

Concen: 20.86 ng

RT: 14.04 min Scan# 1046

Delta R.T. -0.06 min

Lab File: BF083838.D

Acq: 16 Dec 2015 2:46

Instrument :

BNA_F

ClientSampleId :

K205-5-10DL


Tgt Ion:228 Resp: 236699

Ion Ratio Lower Upper

228 100

226 30.6 24.7 37.1




229 22.0 16.3 24.5



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#85

Indeno(1,2,3-cd)pyrene

Concen: 2.09 ng

RT: 16.92 min Scan# 1298

Delta R.T. -0.02 min

Lab File: BF083838.D

Acq: 16 Dec 2015 2:46


Tgt Ion:276 Resp: 23270

Ion Ratio Lower Upper

276 100

138 36.2 0.0 0.0#



277 32.6 0.0 0.0#




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E





#86

Perylene-d12

Concen: 20.00 ng

RT: 15.49 min Scan# 1173

Delta R.T. -0.01 min

Lab File: BF083838.D

Acq: 16 Dec 2015 2:46

Instrument :

BNA_F

ClientSampleId :

K205-5-10DL


Tgt Ion:264 Resp: 158072

Ion Ratio Lower Upper

264 100

260 25.0 19.4 29.0




265 23.3 17.8 26.6



Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E



#87

Benzo(b)fluoranthene

Concen: 10.50 ng

RT: 15.08 min Scan# 1137

Delta R.T. -0.01 min

Lab File: BF083838.D

Acq: 16 Dec 2015 2:46


Tgt Ion:252 Resp: 105846

Ion Ratio Lower Upper

252 100

253 23.2 18.2 27.4



125 15.6 8.2 12.4#




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#88

Benzo(k)fluoranthene

Concen: 11.58 ng

RT: 15.08 min Scan# 1137

Delta R.T. -0.05 min

Lab File: BF083838.D

Acq: 16 Dec 2015 2:46

Instrument :

BNA_F

ClientSampleId :

K205-5-10DL


Tgt Ion:252 Resp: 105846

Ion Ratio Lower Upper

252 100

253 23.2 18.6 27.8




125 15.6 11.5 17.3



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#89

Benzo(a)pyrene

Concen: 8.15 ng

RT: 15.44 min Scan# 1168

Delta R.T. -0.01 min

Lab File: BF083838.D

Acq: 16 Dec 2015 2:46


Tgt Ion:252 Resp: 73310

Ion Ratio Lower Upper

252 100

253 21.9 17.5 26.3



125 13.0 11.0 16.6




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#91

Benzo(g,h,i)perylene

Concen: 2.96 ng

RT: 17.35 min Scan# 1335

Delta R.T. -0.03 min

Lab File: BF083838.D

Acq: 16 Dec 2015 2:46

Instrument :

BNA_F

ClientSampleId :

K205-5-10DL


Tgt Ion: 276 Resp: 27209

Ion Ratio Lower Upper

276 100

277 24.9 19.2 28.8

138 26.9 19.5 29.3



Abundance

Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

