

Data Path : V:\HPCHEM1\BNA F\DATA\BF121515\
 Data File : BF083838.D
 Acq On : 16 Dec 2015 2:46
 Operator : UM/IZ
 Sample : G4739-01DL 10X
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 K205-5-10DL

Quant Time: Dec 16 07:25:13 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF121315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 14 01:30:12 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.90	152	92635	20.00	ng	-0.02
21) Naphthalene-d8	8.19	136	363920	20.00	ng	-0.02
38) Acenaphthene-d10	9.94	164	162397	20.00	ng	-0.02
63) Phenanthrene-d10	11.42	188	266275	20.00	ng	-0.02
75) Chrysene-d12	14.05	240	204580	20.00	ng	-0.02
86) Perylene-d12	15.49	264	158072	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.49	112	64181	11.18	ng	-0.03
7) Phenol-d6	6.52	99	87429	12.02	ng	-0.02
23) Nitrobenzene-d5	7.46	82	49174	7.56	ng	-0.02
41) 2,4,6-Tribromophenol	10.73	330	15891	9.60	ng	-0.02
44) 2-Fluorobiphenyl	9.26	172	104424	8.16	ng	-0.02
78) Terphenyl-d14	13.00	244	80774	8.49	ng	-0.02
Target Compounds						Qvalue
31) Naphthalene	8.20	128	53290	2.81	ng	# 96
37) 2-Methylnaphthalene	8.90	142	49265m	4.14	ng	
48) Acenaphthylene	9.80	152	136358	7.63	ng	# 97
57) Fluorene	10.49	166	88757m	7.87	ng	
70) Phenanthrene	11.45	178	548099	38.75	ng	99
71) Anthracene	11.49	178	233800	15.91	ng	100
74) Fluoranthene	12.64	202	281773	19.28	ng	95
77) Pyrene	12.87	202	446173	27.59	ng	99
80) Benzo(a)anthracene	14.04	228	236699	20.35	ng	95
82) Chrysene	14.08	228	190676m	16.80	ng	
85) Indeno(1,2,3-cd)pyrene	16.92	276	23270	2.09	ng	# 100
87) Benzo(b)fluoranthene	15.08	252	80271m	7.96	ng	
88) Benzo(k)fluoranthene	15.09	252	31251m	3.42	ng	
89) Benzo(a)pyrene	15.44	252	73310	8.15	ng	99
91) Benzo(a,h,i)perylene	17.35	276	27209	2.96	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : V:\HPCHEM1\BNA F\DATA\BF121515\
Data File : BF083838.D
Acq On : 16 Dec 2015 2:46
Operator : UM/IZ
Sample : G4739-01DL 10X
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
K205-5-10DL

Quant Time: Dec 16 07:25:13 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF121315.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 14 01:30:12 2015
Response via : Initial Calibration

