

Data Path : V:\HPCHEM1\BNA F\DATA\BF121515\
 Data File : BF083839.D
 Acq On : 16 Dec 2015 3:14
 Operator : UM/IZ
 Sample : G4739-04DL 5X
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 K206-0-2DL

Quant Time: Dec 16 07:27:04 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF121315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 14 01:30:12 2015
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.90	152	99303	20.00	ng	-0.02
21) Naphthalene-d8	8.19	136	410461	20.00	ng	-0.02
38) Acenaphthene-d10	9.94	164	177645	20.00	ng	-0.02
63) Phenanthrene-d10	11.42	188	294235	20.00	ng	-0.02
75) Chrysene-d12	14.05	240	209332	20.00	ng	-0.02
86) Perylene-d12	15.49	264	170576	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.50	112	152094	24.71	ng	-0.02
7) Phenol-d6	6.52	99	177705	22.80	ng	-0.02
23) Nitrobenzene-d5	7.46	82	116965	15.94	ng	-0.02
41) 2,4,6-Tribromophenol	10.73	330	38021	21.00	ng	-0.02
44) 2-Fluorobiphenyl	9.26	172	198743	13.72	ng	-0.02
78) Terphenyl-d14	13.00	244	134023	13.77	ng	-0.02
Target Compounds						Qvalue
48) Acenaphthylene	9.80	152	152229	7.79	ng	99
49) Dimethylphthalate	9.65	163	52240	3.77	ng	99
70) Phenanthrene	11.45	178	306471	19.61	ng	99
71) Anthracene	11.49	178	79261	4.88	ng	97
74) Fluoranthene	12.64	202	493004	30.52	ng	98
77) Pyrene	12.86	202	721872	43.63	ng	98
80) Benzo(a)anthracene	14.04	228	279996m	23.52	ng	
82) Chrysene	14.08	228	234558	20.20	ng	99
85) Indeno(1,2,3-cd)pyrene	16.92	276	77081	6.76	ng	# 100
87) Benzo(b)fluoranthene	15.08	252	196537m	18.06	ng	
88) Benzo(k)fluoranthene	15.09	252	83560m	8.47	ng	
89) Benzo(a)pyrene	15.44	252	111276	11.46	ng	98
90) Dibenzo(a,h)anthracene	16.93	278	21583m	2.24	ng	
91) Benzo(g,h,i)perylene	17.36	276	80410	8.10	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : V:\HPCHEM1\BNA F\DATA\BF121515\
Data File : BF083839.D
Acq On : 16 Dec 2015 3:14
Operator : UM/IZ
Sample : G4739-04DL 5X
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
K206-0-2DL

Quant Time: Dec 16 07:27:04 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF121315.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 14 01:30:12 2015
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.00 ng
 RT: 6.90 min Scan# 421
 Delta R.T. -0.02 min
 Lab File: BF083839.D
 Acq: 16 Dec 2015 3:14

Instrument :
 BNA_F
 ClientSampleId :
 K206-0-2DL

Tot Ion:152 Resp: 99303
 Ion Ratio Lower Upper
 152 100
 150 159.5 126.5 189.7
 115 70.2 52.3 78.5




Abundance Ion 152.00 (151.70 to 152.70): E
 Ion 150.00 (149.70 to 150.70): E
 Ion 115.00 (114.70 to 115.70): E






#5
 2-Fluorophenol
 Concen: 24.71 ng
 RT: 5.50 min Scan# 299
 Delta R.T. -0.02 min
 Lab File: BF083839.D
 Acq: 16 Dec 2015 3:14

Tgt Ion:112 Resp: 152094
 Ion Ratio Lower Upper
 112 100
 64 45.3 50.5 75.7#
 63 22.3 25.8 38.6#



Abundance Ion 112.00 (111.70 to 112.70): E
 Ion 64.00 (63.70 to 64.70): BF0
 Ion 63.00 (62.70 to 63.70): BF0





#7

Phenol-d6

Concen: 22.80 ng

RT: 6.52 min Scan# 388

Delta R.T. -0.02 min

Lab File: BF083839.D

Acq: 16 Dec 2015 3:14

Instrument :


BNA_F

ClientSampleId :

K206-0-2DL

Tot Ion: 99 Resp: 177705




Ion	Ratio	Lower	Upper
99	100		
42	16.7	17.3	25.9#
71	37.0	26.6	40.0



Abundance Ion 99.00 (98.70 to 99.70): BF0

Ion 42.00 (41.70 to 42.70): BF0

Ion 71.00 (70.70 to 71.70): BF0



#21

Naphthalene-d8

Concen: 20.00 ng

RT: 8.19 min Scan# 534


Delta R.T. -0.02 min

Lab File: BF083839.D

Acq: 16 Dec 2015 3:14

Tgt Ion: 136 Resp: 410461

Ion	Ratio	Lower	Upper
136	100		
137	10.5	8.7	13.1
54	5.0	7.8	11.6#
68	4.5	5.7	8.5#






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF0

Ion 68.00 (67.70 to 68.70): BF0









#23
 Nitrobenzene-d5
 Concen: 15.94 ng
 RT: 7.46 min Scan# 470
 Delta R.T. -0.02 min
 Lab File: BF083839.D
 Acq: 16 Dec 2015 3:14

Instrument :
 BNA_F
 ClientSampleId :
 K206-0-2DL

Tot Ion	82	Resp	116965
Ion Ratio	100	Lower	Upper
128	50.0	30.7	46.1#
54	44.4	42.5	63.7




Abundance Ion 82.00 (81.70 to 82.70): BF0
 Ion 128.00 (127.70 to 128.70): E
 Ion 54.00 (53.70 to 54.70): BF0






#38
 Acenaphthene-d10
 Concen: 20.00 ng
 RT: 9.94 min Scan# 687
 Delta R.T. -0.02 min
 Lab File: BF083839.D
 Acq: 16 Dec 2015 3:14

Tgt Ion	164	Resp	177645
Ion Ratio	100	Lower	Upper
162	103.0	78.8	118.2
160	46.8	33.4	50.2



Abundance Ion 164.00 (163.70 to 164.70): E
 Ion 162.00 (161.70 to 162.70): E
 Ion 160.00 (159.70 to 160.70): E









#41
 2,4,6-Tribromophenol
 Concen: 21.00 ng
 RT: 10.73 min Scan# 756
 Delta R.T. -0.02 min
 Lab File: BF083839.D
 Acq: 16 Dec 2015 3:14

Instrument :
 BNA_F
 ClientSampleId :
 K206-0-2DL

Tot Ion:330 Resp: 38021
 Ion Ratio Lower Upper
 330 100
 332 95.9 0.0 0.0#
 141 41.3 0.0 0.0#




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#44
 2-Fluorobiphenyl
 Concen: 13.72 ng
 RT: 9.26 min Scan# 628
 Delta R.T. -0.02 min
 Lab File: BF083839.D
 Acq: 16 Dec 2015 3:14

Tgt Ion:172 Resp: 198743
 Ion Ratio Lower Upper
 172 100
 171 35.2 28.6 42.8
 170 23.2 19.4 29.0



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E





#48

Acenaphthylene

Concen: 7.79 ng

RT: 9.80 min Scan# 675

Delta R.T. -0.02 min

Lab File: BF083839.D

Acq: 16 Dec 2015 3:14

Instrument :


BNA_F

ClientSampleId :

K206-0-2DL

Tot Ion:152 Resp: 152229




Ion	Ratio	Lower	Upper
152	100		
151	20.0	16.3	24.5
153	13.0	10.8	16.2



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E



#49

Dimethylphthalate

Concen: 3.77 ng

RT: 9.65 min Scan# 662


Delta R.T. -0.03 min

Lab File: BF083839.D

Acq: 16 Dec 2015 3:14

Tgt Ion:163 Resp: 52240



Ion	Ratio	Lower	Upper
163	100		
194	3.4	2.8	4.2
164	10.4	8.6	13.0




Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E





#63

Phenanthrene-d10

Concen: 20.00 ng

RT: 11.42 min Scan# 817

Delta R.T. -0.02 min

Lab File: BF083839.D

Acq: 16 Dec 2015 3:14

Instrument :

BNA_F

ClientSampleId :

K206-0-2DL


Tot Ion:188 Resp: 294235

Ion Ratio Lower Upper

188 100

94 9.0 9.4 14.2#




80 9.2 10.6 16.0#



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BF0

Ion 80.00 (79.70 to 80.70): BF0



#70

Phenanthrene

Concen: 19.61 ng

RT: 11.45 min Scan# 819

Delta R.T. -0.02 min

Lab File: BF083839.D

Acq: 16 Dec 2015 3:14


Tgt Ion:178 Resp: 306471

Ion Ratio Lower Upper

178 100

176 18.8 15.7 23.5



179 15.3 12.2 18.4




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E





#71

Anthracene

Concen: 4.88 ng

RT: 11.49 min Scan# 823

Delta R.T. -0.02 min

Lab File: BF083839.D

Acq: 16 Dec 2015 3:14

Instrument :

BNA_F

ClientSampleId :

K206-0-2DL


Tot Ion:178 Resp: 79261

Ion Ratio Lower Upper

178 100

176 17.5 15.2 22.8




179 14.9 12.6 18.8



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#74

Fluoranthene

Concen: 30.52 ng

RT: 12.64 min Scan# 923

Delta R.T. -0.01 min

Lab File: BF083839.D

Acq: 16 Dec 2015 3:14


Tgt Ion:202 Resp: 493004

Ion Ratio Lower Upper

202 100

101 10.6 0.0 31.5



203 18.1 0.0 37.6




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E





#75

Chrysene-d12

Concen: 20.00 ng

RT: 14.05 min Scan# 1047

Delta R.T. -0.02 min

Lab File: BF083839.D

Acq: 16 Dec 2015 3:14

Instrument :

BNA_F

ClientSampleId :

K206-0-2DL


Tot Ion:240 Resp: 209332

Ion Ratio Lower Upper

240 100

120 15.9 8.1 12.1#




236 25.3 20.8 31.2



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#77

Pyrene

Concen: 43.63 ng

RT: 12.86 min Scan# 943

Delta R.T. -0.01 min

Lab File: BF083839.D

Acq: 16 Dec 2015 3:14


Tgt Ion:202 Resp: 721872

Ion Ratio Lower Upper

202 100

200 22.2 16.7 25.1



203 18.2 14.3 21.5




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E









#78
 Terphenyl-d14
 Concen: 13.77 ng
 RT: 13.00 min Scan# 955
 Delta R.T. -0.02 min
 Lab File: BF083839.D
 Acq: 16 Dec 2015 3:14




Instrument :
 BNA_F
 ClientSampleId :
 K206-0-2DL


Tot Ion:244 Resp: 134023
 Ion Ratio Lower Upper
 244 100
 212 7.4 6.1 9.1
 122 12.4 11.0 16.6



#80
 Benzo(a)anthracene
 Concen: 23.52 ng m
 RT: 14.04 min Scan# 1046
 Delta R.T. -0.02 min
 Lab File: BF083839.D
 Acq: 16 Dec 2015 3:14

Tgt Ion:228 Resp: 279996
 Ion Ratio Lower Upper
 228 100
 226 34.9 22.3 33.5#
 229 20.7 15.7 23.5





#82

Chrysene

Concen: 20.20 ng

RT: 14.08 min Scan# 1049

Delta R.T. -0.02 min

Lab File: BF083839.D

Acq: 16 Dec 2015 3:14

Instrument :

BNA_F

ClientSampleId :

K206-0-2DL


Tot Ion:228 Resp: 234558

Ion Ratio Lower Upper

228 100

226 30.5 24.7 37.1




229 21.2 16.3 24.5



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#85

Indeno(1,2,3-cd)pyrene

Concen: 6.76 ng

RT: 16.92 min Scan# 1298

Delta R.T. -0.02 min

Lab File: BF083839.D

Acq: 16 Dec 2015 3:14


Tgt Ion:276 Resp: 77081

Ion Ratio Lower Upper

276 100

138 24.2 0.0 0.0#



277 25.3 0.0 0.0#




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E









#86
Pervlene-d12
Concen: 20.00 ng
RT: 15.49 min Scan# 1173
Delta R.T. -0.01 min
Lab File: BF083839.D
Acq: 16 Dec 2015 3:14




Instrument :
BNA_F
ClientSampleId :
K206-0-2DL


Tot Ion:264 Resp: 170576
Ion Ratio Lower Upper
264 100
260 23.1 19.4 29.0
265 23.1 17.8 26.6



#87
Benzo(b)fluoranthene
Concen: 18.06 ng m
RT: 15.08 min Scan# 1137
Delta R.T. -0.01 min
Lab File: BF083839.D
Acq: 16 Dec 2015 3:14

Tgt Ion:252 Resp: 196537
Ion Ratio Lower Upper
252 100
253 22.3 18.2 27.4
125 14.0 8.2 12.4#









#88
Benzo(k)fluoranthene
Concen: 8.47 ng m
RT: 15.09 min Scan# 1138
Delta R.T. -0.03 min
Lab File: BF083839.D
Acq: 16 Dec 2015 3:14




Instrument :
BNA_F
ClientSampleId :
K206-0-2DL


Tot Ion:252 Resp: 83560
Ion Ratio Lower Upper
252 100
253 21.9 18.6 27.8
125 13.3 11.5 17.3



#89
Benzo(a)pyrene
Concen: 11.46 ng
RT: 15.44 min Scan# 1168
Delta R.T. -0.01 min
Lab File: BF083839.D
Acq: 16 Dec 2015 3:14

Tgt Ion:252 Resp: 111276
Ion Ratio Lower Upper
252 100
253 23.0 17.5 26.3
125 13.0 11.0 16.6





#90

Dibenzo(a,h)anthracene

Concen: 2.24 ng

RT: 16.93 min Scan# 1299

Delta R.T. -0.02 min

Lab File: BF083839.D

Acq: 16 Dec 2015 3:14

Instrument :

BNA_F

ClientSampleId :

K206-0-2DL


Tot Ion:278 Resp: 21583

Ion Ratio Lower Upper

278 100

139 19.4 12.8 19.2#




279 26.0 18.6 28.0



Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E



#91

Benzo(g,h,i)perylene

Concen: 8.10 ng

RT: 17.36 min Scan# 1336

Delta R.T. -0.02 min

Lab File: BF083839.D

Acq: 16 Dec 2015 3:14


Tgt Ion:276 Resp: 80410

Ion Ratio Lower Upper

276 100

277 24.3 19.2 28.8

138 24.7 19.5 29.3



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

