


Data Path : Z:\HPCHEM1\BNA\_F\DATA\BF121516\  
 Data File : BF091627.D  
 Acq On : 15 Dec 2016 14:13  
 Operator : UM/SJ  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF120916.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Fri Dec 16 00:35:41 2016



AutoFind: Scans 829, 830, 831; Background Corrected with Scan 826

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 36.6      | 27764   | PASS             |
| 68          | 69           | 0.00         | 2            | 0.4       | 108     | PASS             |
| 69          | 198          | 0.00         | 100          | 36.7      | 27896   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 127         | 198          | 10           | 80           | 49.5      | 37602   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.4       | 307     | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 75950   | PASS             |
| 199         | 198          | 5            | 9            | 6.8       | 5166    | PASS             |
| 275         | 198          | 10           | 60           | 22.4      | 17036   | PASS             |
| 365         | 198          | 1            | 100          | 2.2       | 1706    | PASS             |
| 441         | 442          | 0.01         | 24           | 15.4      | 8333    | PASS             |
| 442         | 198          | 50           | 100          | 71.2      | 54110   | PASS             |
| 443         | 442          | 15           | 24           | 18.6      | 10045   | PASS             |