

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121520\
 Data File : BF122731.D
 Acq On : 16 Dec 2020 00:21
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 16 01:57:15 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 15 18:14:23 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	70	0.00
2	1,4-Dioxane	0.594	0.642	-8.1	76	0.00
3	Pyridine	1.569	1.630	-3.9	71	-0.02
4	n-Nitrosodimethylamine	0.629	0.613	2.5	67	-0.02
5 S	2-Fluorophenol	1.263	1.319	-4.4	73	0.00
6	Aniline	1.972	1.942	1.5	68	0.00
7 S	Phenol-d6	1.675	1.653	1.3	69	0.00
8	2-Chlorophenol	1.392	1.403	-0.8	70	0.00
9	Benzaldehyde	1.014	0.871	14.1	68	0.00
10 C	Phenol	1.736	1.725	0.6	70	0.00
11	bis(2-Chloroethyl)ether	1.326	1.297	2.2	68	0.00
12	1,3-Dichlorobenzene	1.648	1.647	0.1	70	0.00
13 C	1,4-Dichlorobenzene	1.661	1.669	-0.5	71	0.00
14	1,2-Dichlorobenzene	1.604	1.548	3.5	71	0.00
15	Benzyl Alcohol	1.154	1.094	5.2	64	0.00
16	2,2'-oxybis(1-Chloropropane	1.891	1.759	7.0	65	0.00
17	2-Methylphenol	1.107	1.078	2.6	66	0.00
18	Hexachloroethane	0.592	0.585	1.2	69	0.00
19 P	n-Nitroso-di-n-propylamine	0.960	0.842	12.3	61	0.00
20	3+4-Methylphenols	1.398	1.308	6.4	67	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	65	0.00
22	Acetophenone	0.497	0.489	1.6	64	0.00
23 S	Nitrobenzene-d5	0.399	0.403	-1.0	64	0.00
24	Nitrobenzene	0.397	0.398	-0.3	65	0.00
25	Isophorone	0.688	0.647	6.0	60	0.00
26 C	2-Nitrophenol	0.194	0.203	-4.6	65	0.00
27	2,4-Dimethylphenol	0.284	0.285	-0.4	64	0.00
28	bis(2-Chloroethoxy)methane	0.471	0.462	1.9	63	0.00
29 C	2,4-Dichlorophenol	0.332	0.331	0.3	64	0.00
30	1,2,4-Trichlorobenzene	0.376	0.383	-1.9	66	0.00
31	Naphthalene	1.104	1.117	-1.2	66	0.00
32	Benzoic acid	0.226	0.182	19.5	48#	-0.02
33	4-Chloroaniline	0.461	0.450	2.4	62	0.00
34 C	Hexachlorobutadiene	0.231	0.233	-0.9	65	0.00
35	Caprolactam	0.100	0.088	12.0	56	-0.01
36 C	4-Chloro-3-methylphenol	0.327	0.305	6.7	59	0.00
37	2-Methylnaphthalene	0.730	0.701	4.0	62	0.00
38	1-Methylnaphthalene	0.691	0.660	4.5	62	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	58	0.00
40	1,2,4,5-Tetrachlorobenzene	0.662	0.722	-9.1	62	0.00
41 P	Hexachlorocyclopentadiene	0.293	0.246	16.0	45#	0.00
42 S	2,4,6-Tribromophenol	0.262	0.255	2.7	55	0.00
43 C	2,4,6-Trichlorophenol	0.458	0.467	-2.0	57	0.00
44	2,4,5-Trichlorophenol	0.473	0.480	-1.5	58	0.00
45 S	2-Fluorobiphenyl	1.479	1.474	0.3	61	0.00
46	1,1'-Biphenyl	1.660	1.750	-5.4	61	0.00
47	2-Chloronaphthalene	1.375	1.459	-6.1	61	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121520\
 Data File : BF122731.D
 Acq On : 16 Dec 2020 00:21
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 16 01:57:15 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 15 18:14:23 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.401	0.390	2.7	54	0.00
49	Acenaphthylene	2.031	2.083	-2.6	59	0.00
50	Dimethylphthalate	1.597	1.535	3.9	55	0.00
51	2,6-Dinitrotoluene	0.337	0.339	-0.6	56	0.00
52 C	Acenaphthene	1.314	1.312	0.2	57	0.00
53	3-Nitroaniline	0.390	0.386	1.0	55	0.00
54 P	2,4-Dinitrophenol	0.165	0.138	16.4	45#	0.00
55	Dibenzofuran	1.849	1.878	-1.6	59	0.00
56 P	4-Nitrophenol	0.278	0.241	13.3	47#	0.00
57	2,4-Dinitrotoluene	0.449	0.433	3.6	53	0.00
58	Fluorene	1.470	1.398	4.9	58	0.00
59	2,3,4,6-Tetrachlorophenol	0.415	0.393	5.3	53	0.00
60	Diethylphthalate	1.554	1.431	7.9	53	0.00
61	4-Chlorophenyl-phenylether	0.724	0.709	2.1	58	0.00
62	4-Nitroaniline	0.396	0.378	4.5	53	0.00
63	Azobenzene	1.428	1.351	5.4	54	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	53	0.00
65	4,6-Dinitro-2-methylphenol	0.122	0.116	4.9	47#	0.00
66 c	n-Nitrosodiphenylamine	0.707	0.739	-4.5	55	0.00
67	4-Bromophenyl-phenylether	0.271	0.287	-5.9	55	0.00
68	Hexachlorobenzene	0.300	0.310	-3.3	54	0.00
69	Atrazine	0.229	0.216	5.7	50#	0.00
70 C	Pentachlorophenol	0.159	0.147	7.5	45#	0.00
71	Phenanthrene	1.189	1.220	-2.6	55	0.00
72	Anthracene	1.179	1.212	-2.8	55	0.00
73	Carbazole	1.069	1.078	-0.8	53	0.00
74	Di-n-butylphthalate	1.343	1.291	3.9	51	0.00
75 C	Fluoranthene	1.246	1.216	2.4	52	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	53	0.00
77	Benzidine	0.433	0.497	-14.8	54	0.00
78	Pyrene	1.546	1.526	1.3	52	0.00
79 S	Terphenyl-d14	1.143	1.112	2.7	54	0.00
80	Butylbenzylphthalate	0.697	0.652	6.5	49#	0.00
81	Benzo(a)anthracene	1.337	1.376	-2.9	55	0.00
82	3,3'-Dichlorobenzidine	0.479	0.490	-2.3	54	0.00
83	Chrysene	1.330	1.343	-1.0	54	0.00
84	Bis(2-ethylhexyl)phthalate	0.954	0.886	7.1	50	0.00
85 c	Di-n-octyl phthalate	1.582	1.521	3.9	52	0.00
86 I	Perylene-d12	1.000	1.000	0.0	65	0.00
87	Indeno(1,2,3-cd)pyrene	1.313	1.612	-22.8	81	-0.01
88	Benzo(b)fluoranthene	1.403	1.309	6.7	58	0.00
89	Benzo(k)fluoranthene	1.283	1.310	-2.1	67	0.00
90 C	Benzo(a)pyrene	1.228	1.247	-1.5	65	0.00
91	Dibenzo(a,h)anthracene	1.074	1.309	-21.9	81	0.00
92	Benzo(g,h,i)perylene	1.040	1.346	-29.4#	87	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121520\
Data File : BF122731.D
Acq On : 16 Dec 2020 00:21
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Dec 16 01:57:15 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 15 18:14:23 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0