

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121520\
 Data File : BF122736.D
 Acq On : 16 Dec 2020 2:55
 Operator : JU/CG
 Sample : L5102-04
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 LLEXSITUTV-015-002-SO

Quant Time: Dec 16 03:43:16 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 15 18:14:23 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.816	152	142334	20.00	ng	0.00
21) Naphthalene-d8	8.098	136	529343	20.00	ng	-0.01
39) Acenaphthene-d10	9.851	164	263643	20.00	ng	-0.01
64) Phenanthrene-d10	11.339	188	434479	20.00	ng	0.00
76) Chrysene-d12	13.974	240	326961	20.00	ng	-0.01
86) Perylene-d12	15.433	264	318943	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.434	112	405657	45.12	ng	0.00
7) Phenol-d6	6.445	99	644333	54.04	ng	-0.01
23) Nitrobenzene-d5	7.375	82	505651	47.86	ng	-0.02
42) 2,4,6-Tribromophenol	10.639	330	91774	26.59	ng	-0.01
45) 2-Fluorobiphenyl	9.174	172	961986	49.35	ng	-0.01
79) Terphenyl-d14	12.921	244	828932	44.38	ng	-0.01


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121520\
Data File : BF122736.D
Acq On : 16 Dec 2020 2:55
Operator : JU/CG
Sample : L5102-04
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
LLEXSITUTV-015-002-SO

Quant Time: Dec 16 03:43:16 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 15 18:14:23 2020
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 6.816 min Scan# 804
Delta R.T. -0.006 min
Lab File: BF122736.D
Acq: 16 Dec 2020 2:55




Instrument :
BNA_F
ClientSampleId :
LLEXSITUTV-015-002-SO


Tgt Ion:152 Resp: 142334
Ion Ratio Lower Upper
152 100
150 154.9 125.8 188.6
115 59.9 49.3 73.9



#5
2-Fluorophenol
Concen: 45.12 ng
RT: 5.434 min Scan# 569
Delta R.T. -0.000 min
Lab File: BF122736.D
Acq: 16 Dec 2020 2:55

Tgt Ion:112 Resp: 405657
Ion Ratio Lower Upper
112 100
64 51.7 46.9 70.3
63 26.5 24.5 36.7









#7
Phenol-d6
Concen: 54.04 ng
RT: 6.445 min Scan# 741
Delta R.T. -0.012 min
Lab File: BF122736.D
Acq: 16 Dec 2020 2:55




Instrument :
BNA_F
ClientSampleId :
LLEXSITUTV-015-002-SO


Tgt Ion: 99 Resp: 644333
Ion Ratio Lower Upper
99 100
42 14.9 13.8 20.6
71 31.4 26.9 40.3



#21
Naphthalene-d8
Concen: 20.00 ng
RT: 8.098 min Scan# 1022
Delta R.T. -0.012 min
Lab File: BF122736.D
Acq: 16 Dec 2020 2:55

Tgt Ion: 136 Resp: 529343
Ion Ratio Lower Upper
136 100
137 11.0 8.8 13.2
54 7.7 5.5 8.3
68 6.2 4.6 6.8









#23
Nitrobenzene-d5
Concen: 47.86 ng
RT: 7.375 min Scan# 899
Delta R.T. -0.018 min
Lab File: BF122736.D
Acq: 16 Dec 2020 2:55




Instrument :
BNA_F
ClientSampleId :
LLEXSITUTV-015-002-SO


Tgt Ion: 82 Resp: 505651
Ion Ratio Lower Upper
82 100
128 45.6 37.0 55.4
54 46.7 38.4 57.6



#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 9.851 min Scan# 1320
Delta R.T. -0.012 min
Lab File: BF122736.D
Acq: 16 Dec 2020 2:55

Tgt Ion: 164 Resp: 263643
Ion Ratio Lower Upper
164 100
162 102.0 80.8 121.2
160 44.8 35.4 53.0









#42
2,4,6-Tribromophenol
Concen: 26.59 ng
RT: 10.639 min Scan# 1454
Delta R.T. -0.012 min
Lab File: BF122736.D
Acq: 16 Dec 2020 2:55




Instrument :
BNA_F
ClientSampleId :
LLEXSITUTV-015-002-SO


Tgt Ion: 330 Resp: 91774
Ion Ratio Lower Upper
330 100
332 98.5 77.5 116.3
141 29.3 23.8 35.8



#45
2-Fluorobiphenyl
Concen: 49.35 ng
RT: 9.174 min Scan# 1205
Delta R.T. -0.012 min
Lab File: BF122736.D
Acq: 16 Dec 2020 2:55

Tgt Ion: 172 Resp: 961986
Ion Ratio Lower Upper
172 100
171 36.2 29.6 44.4
170 23.8 19.5 29.3









#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 11.339 min Scan# 1573
Delta R.T. -0.006 min
Lab File: BF122736.D
Acq: 16 Dec 2020 2:55

Instrument :
BNA_F
ClientSampleId :
LLEXSITUTV-015-002-SO

Tgt Ion:188 Resp: 434479




Ion	Ratio	Lower	Upper
188	100		
94	9.0	7.4	11.2
80	9.8	8.2	12.4




#76
Chrysene-d12
Concen: 20.00 ng
RT: 13.974 min Scan# 2021
Delta R.T. -0.012 min
Lab File: BF122736.D
Acq: 16 Dec 2020 2:55

Tgt Ion:240 Resp: 326961

Ion	Ratio	Lower	Upper
240	100		
120	11.7	8.6	12.8
236	25.0	20.3	30.5









#79
Terphenyl-d14
Concen: 44.38 ng
RT: 12.921 min Scan# 1842
Delta R.T. -0.012 min
Lab File: BF122736.D
Acq: 16 Dec 2020 2:55

Instrument :
BNA_F
ClientSampleId :
LLEXSITUTV-015-002-SO

Tgt Ion:244 Resp: 828932
Ion Ratio Lower Upper
244 100
212 6.7 5.7 8.5
122 11.1 8.3 12.5



#86
Perylene-d12
Concen: 20.00 ng
RT: 15.433 min Scan# 2269
Delta R.T. -0.006 min
Lab File: BF122736.D
Acq: 16 Dec 2020 2:55

Tgt Ion:264 Resp: 318943
Ion Ratio Lower Upper
264 100
260 23.0 18.6 28.0
265 21.5 17.0 25.6

