

Data Path : Z:\HPCHEM1\BNA F\DATA\BF121616\
 Data File : BF091677.D
 Acq On : 16 Dec 2016 17:20
 Operator : UM/SJ
 Sample : PB95306BL
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB95306BL

Quant Time: Dec 16 23:12:38 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF120916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 16 00:35:41 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.77	152	245188	20.00	ng	-0.01
21) Naphthalene-d8	8.05	136	951044	20.00	ng	-0.02
38) Acenaphthene-d10	9.81	164	528042	20.00	ng	-0.01
63) Phenanthrene-d10	11.29	188	838618	20.00	ng	-0.02
75) Chrysene-d12	13.92	240	813637	20.00	ng	-0.02
86) Perylene-d12	15.32	264	677760	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.38	112	1632158	112.15	ng	0.00
7) Phenol-d6	6.42	99	2144291	118.19	ng	-0.01
23) Nitrobenzene-d5	7.33	82	1270404	74.55	ng	-0.02
41) 2,4,6-Tribromophenol	10.60	330	535397	122.07	ng	-0.01
44) 2-Fluorobiphenyl	9.14	172	2319508	75.90	ng	-0.01
78) Terphenyl-d14	12.88	244	2450309	68.34	ng	-0.02

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF121616\
Data File : BF091677.D
Acq On : 16 Dec 2016 17:20
Operator : UM/SJ
Sample : PB95306BL
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB95306BL

Quant Time: Dec 16 23:12:38 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF120916.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 16 00:35:41 2016
Response via : Initial Calibration

