

Data Path : Z:\HPCHEM1\BNA F\DATA\BF121616\
 Data File : BF091694.D
 Acq On : 17 Dec 2016 1:08
 Operator : UM/SJ
 Sample : H6125-05
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-45(8-9)-12-14-16

Quant Time: Dec 19 13:13:09 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF120916.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 16 00:35:41 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.80	152	137560	20.00	ng	0.01
21) Naphthalene-d8	8.11	136	363162	20.00	ng	0.03
38) Acenaphthene-d10	9.87	164	147097	20.00	ng	0.05
63) Phenanthrene-d10	11.31	188	481448	20.00	ng	0.00
75) Chrysene-d12	13.94	240	479092	20.00	ng	0.00
86) Perylene-d12	15.33	264	208779	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.38	112	1358523	166.38	ng	0.00
7) Phenol-d6	6.43	99	1605169	157.69	ng	0.00
23) Nitrobenzene-d5	7.37	82	942875	144.91	ng	0.01
41) 2,4,6-Tribromophenol	10.62	330	216779	177.43	ng	0.01
44) 2-Fluorobiphenyl	9.20	172	553964	65.07	ng	0.05
78) Terphenyl-d14	12.90	244	821956	38.93	ng	0.00
Target Compounds						
37) 2-Methylnaphthalene	8.85	142	2578466	235.77	ng	# 95
57) Fluorene	10.40	166	288023	36.34	ng	# 82
70) Phenanthrene	11.33	178	342988	14.39	ng	# 96
71) Anthracene	11.38	178	63108	2.45	ng	# 62
74) Fluoranthene	12.53	202	243403	9.75	ng	# 91
77) Pyrene	12.75	202	182750	4.81	ng	# 84
80) Benzo(a)anthracene	13.93	228	68673m	2.47	ng	#
83) Bis(2-ethylhexyl)phthalate	13.92	149	199756	10.51	ng	# 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF121616\
Data File : BF091694.D
Acq On : 17 Dec 2016 1:08
Operator : UM/SJ
Sample : H6125-05
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-45(8-9)-12-14-16

Quant Time: Dec 19 13:13:09 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF120916.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 16 00:35:41 2016
Response via : Initial Calibration

