


Data Path : Z:\HPCHEM1\BNA F\DATA\BF121716\  
 Data File : BF091719.D  
 Acq On : 17 Dec 2016 23:21  
 Operator : UM/SJ  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF121716.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Sat Dec 17 22:53:54 2016



AutoFind: Scans 828, 829, 830; Background Corrected with Scan 824

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 35.2      | 23092   | PASS             |
| 68          | 69           | 0.00         | 2            | 1.7       | 410     | PASS             |
| 69          | 198          | 0.00         | 100          | 36.6      | 24031   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 127         | 198          | 10           | 80           | 48.1      | 31552   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.6       | 373     | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 65645   | PASS             |
| 199         | 198          | 5            | 9            | 6.3       | 4142    | PASS             |
| 275         | 198          | 10           | 60           | 23.5      | 15426   | PASS             |
| 365         | 198          | 1            | 100          | 2.8       | 1819    | PASS             |
| 441         | 442          | 0.01         | 24           | 16.1      | 8713    | PASS             |
| 442         | 198          | 50           | 100          | 82.7      | 54275   | PASS             |
| 443         | 442          | 15           | 24           | 19.2      | 10409   | PASS             |