

Data Path : Z:\HPCHEM1\BNA F\DATA\BF121816\
 Data File : BF091777.D
 Acq On : 19 Dec 2016 12:45
 Operator : UM/SJ
 Sample : H6089-01DL 10X
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 T-2DL

Quant Time: Dec 19 17:50:01 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF121716.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Dec 17 22:53:54 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.76	152	205260	20.00	ng	-0.01
21) Naphthalene-d8	8.04	136	722086	20.00	ng	-0.02
38) Acenaphthene-d10	9.80	164	420594	20.00	ng	-0.01
63) Phenanthrene-d10	11.28	188	655317	20.00	ng	-0.01
75) Chrysene-d12	13.91	240	543885	20.00	ng	-0.02
86) Perylene-d12	15.30	264	430403	20.00	ng	-0.03
System Monitoring Compounds						
5) 2-Fluorophenol	5.34	112	143161	11.69	ng	-0.02
7) Phenol-d6	6.40	99	189141	12.66	ng	-0.02
23) Nitrobenzene-d5	7.32	82	142017	11.34	ng	-0.02
41) 2,4,6-Tribromophenol	10.59	330	31218	8.75	ng	-0.01
44) 2-Fluorobiphenyl	9.13	172	218016	10.41	ng	-0.01
78) Terphenyl-d14	12.85	244	207023	9.61	ng	-0.02
Target Compounds						Qvalue
31) Naphthalene	8.06	128	433715	12.57	ng	99
37) 2-Methylnaphthalene	8.76	142	1296449	59.61	ng	99
45) 1,1'-Biphenyl	9.22	154	172287	5.72	ng	95
51) Acenaphthene	9.84	154	89568	3.74	ng	# 63
57) Fluorene	10.35	166	219415	9.92	ng	# 92
70) Phenanthrene	11.30	178	617081	18.85	ng	98
71) Anthracene	11.30	178	617081	17.79	ng	99
77) Pyrene	12.71	202	111968	2.85	ng	# 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF121816\
Data File : BF091777.D
Acq On : 19 Dec 2016 12:45
Operator : UM/SJ
Sample : H6089-01DL 10X
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
T-2DL

Quant Time: Dec 19 17:50:01 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF121716.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Dec 17 22:53:54 2016
Response via : Initial Calibration

