

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF121817\
 Data File : BF101489.D
 Acq On : 18 Dec 2017 19:09
 Operator : SJ/JU
 Sample : I6908-13
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 7

Manual Integrations
 APPROVED

Sohil
 12/19/2017 4:49:58 PM

Quant Time: Dec 19 06:22:05 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF121417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 18 15:38:59 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.81	152	156783	20.00	ng	0.00
21) Naphthalene-d8	8.09	136	606751	20.00	ng	0.00
38) Acenaphthene-d10	9.85	164	239991	20.00	ng	0.00
63) Phenanthrene-d10	11.34	188	381556	20.00	ng	0.00
75) Chrysene-d12	13.99	240	293690	20.00	ng	0.00
86) Perylene-d12	15.46	264	254947	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.43	112	901843	85.68	ng	0.01
7) Phenol-d6	6.45	99	1102350	84.15	ng	0.00
23) Nitrobenzene-d5	7.38	82	651472	59.77	ng	0.00
41) 2,4,6-Tribromophenol	10.65	330	189102	81.77	ng	0.00
44) 2-Fluorobiphenyl	9.16	172	993803	64.24	ng	0.00
78) Terphenyl-d14	12.92	244	707229	58.05	ng	0.00
Target Compounds						
10) Phenol	6.46	94	59966	4.016	ng	Qvalue 90
49) Dimethylphthalate	9.56	163	56147	2.909	ng	98
74) Fluoranthene	12.56	202	95212	4.275	ng	98
77) Pyrene	12.79	202	88331	4.008	ng	99
80) Benzo(a)anthracene	13.98	228	44086	2.273	ng	96
82) Chrysene	14.02	228	40445	2.209	ng	98
87) Benzo(b)fluoranthene	15.03	252	44167m	2.842	ng	
89) Benzo(a)pyrene	15.40	252	29217	2.040	ng	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF121817\
Data File : BF101489.D
Acq On : 18 Dec 2017 19:09
Operator : SJ/JU
Sample : I6908-13
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampled :
7

Manual Integrations
APPROVED

Sohil
12/19/2017 4:49:58 PM

Quant Time: Dec 19 06:22:05 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF121417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 18 15:38:59 2017
Response via : Initial Calibration

