

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF121817\
 Data File : BF101511.D
 Acq On : 19 Dec 2017 6:31
 Operator : SJ/JU
 Sample : I6910-03
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 121317-TIDO-E-D-M

Quant Time: Dec 19 07:19:12 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF121417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 18 15:38:59 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.81	152	123789	20.00	ng	0.00
21) Naphthalene-d8	8.09	136	440147	20.00	ng	0.00
38) Acenaphthene-d10	9.85	164	157128	20.00	ng	0.00
63) Phenanthrene-d10	11.34	188	256266	20.00	ng	0.00
75) Chrysene-d12	13.99	240	284611	20.00	ng	0.00
86) Perylene-d12	15.46	264	260252	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.43	112	317412	38.19	ng	0.01
7) Phenol-d6	6.46	99	224328	21.69	ng	0.00
23) Nitrobenzene-d5	7.38	82	699133	88.42	ng	0.00
41) 2,4,6-Tribromophenol	10.65	330	130133	85.95	ng	0.00
44) 2-Fluorobiphenyl	9.17	172	1097112	108.31	ng	0.00
78) Terphenyl-d14	12.92	244	896399	75.93	ng	0.00
Target Compounds						Qvalue
31) Naphthalene	8.12	128	88600	3.998	ng	99
37) 2-Methylnaphthalene	8.81	142	51170	3.544	ng	94
49) Dimethylphthalate	9.56	163	35443	2.804	ng	99
51) Acenaphthene	9.88	154	50227	4.964	ng	98
54) Dibenzofuran	10.06	168	48352	3.760	ng	98
57) Fluorene	10.40	166	41760	3.839	ng	98
59) Diethylphthalate	10.26	149	28838	2.304	ng	99
70) Phenanthrene	11.36	178	123287	8.651	ng	97
74) Fluoranthene	12.56	202	56182	3.756	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF121817\
Data File : BF101511.D
Acq On : 19 Dec 2017 6:31
Operator : SJ/JU
Sample : I6910-03
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
121317-TIDO-E-D-M

Quant Time: Dec 19 07:19:12 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF121417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 18 15:38:59 2017
Response via : Initial Calibration

