

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121823\
 Data File : BF136621.D
 Acq On : 18 Dec 2023 18:39
 Operator : CG\JU
 Sample : 05805-03RE
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 HR-04-12122023RE

Quant Time: Dec 19 01:05:36 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF120523.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 19 00:56:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.787	152	90400	20.000	ng	0.00
21) Naphthalene-d8	8.063	136	307086	20.000	ng	0.00
39) Acenaphthene-d10	9.822	164	125520	20.000	ng	0.00
64) Phenanthrene-d10	11.316	188	228771	20.000	ng	0.00
76) Chrysene-d12	13.968	240	240468	20.000	ng	# 0.00
86) Perylene-d12	15.451	264	198426	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.428	112	358022	58.489	ng	0.00
7) Phenol-d6	6.428	99	414398	54.261	ng	-0.02
23) Nitrobenzene-d5	7.345	82	235017	41.373	ng	-0.01
42) 2,4,6-Tribromophenol	10.616	330	79091	51.103	ng	-0.01
45) 2-Fluorobiphenyl	9.139	172	402288	43.732	ng	-0.01
79) Terphenyl-d14	12.916	244	496912	27.237	ng	0.00
Target Compounds						
37) 2-Methylnaphthalene	8.780	142	25950	2.409	ng	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121823\
Data File : BF136621.D
Acq On : 18 Dec 2023 18:39
Operator : CG\JU
Sample : 05805-03RE
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
HR-04-12122023RE

Quant Time: Dec 19 01:05:36 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF120523.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 19 00:56:37 2023
Response via : Initial Calibration

