

Data Path : Z:\HPCHEM1\BNA F\DATA\BF121916\
 Data File : BF091802.D
 Acq On : 20 Dec 2016 4:39
 Operator : UM/SJ
 Sample : H6125-06RE
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-45(10-12)-12-14-16RE

Manual Integrations
 APPROVED

sohil
 12/20/2016 6:32:44 PM

Quant Time: Dec 20 06:21:08 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF121716.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Dec 17 22:53:54 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.77	152	189638	20.00	ng	0.00
21) Naphthalene-d8	8.09	136	422780	20.00	ng	0.02
38) Acenaphthene-d10	9.85	164	216035	20.00	ng	0.03
63) Phenanthrene-d10	11.30	188	419365	20.00	ng	0.01
75) Chrysene-d12	13.95	240	429536	20.00	ng	0.02
86) Perylene-d12	15.35	264	401011	20.00	ng	0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.37	112	1763317	155.78	ng	0.00
7) Phenol-d6	6.42	99	2209406	160.05	ng	0.00
23) Nitrobenzene-d5	7.33	82	977601	133.35	ng	-0.01
41) 2,4,6-Tribromophenol	10.61	330	290876	158.78	ng	0.01
44) 2-Fluorobiphenyl	9.17	172	709085	65.94	ng	0.04
78) Terphenyl-d14	12.93	244	537555	31.60	ng	0.06
Target Compounds						
10) Phenol	6.43	94	102274	6.57	ng	# 79
37) 2-Methylnaphthalene	8.82	142	190870	14.99	ng	# 79
57) Fluorene	10.37	166	183894	16.18	ng	# 73
70) Phenanthrene	11.32	178	280486	13.39	ng	# 80
74) Fluoranthene	12.53	202	131954m	6.03	ng	
83) Bis(2-ethylhexyl)phthalate	13.93	149	213812	11.61	ng	# 64

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF121916\
Data File : BF091802.D
Acq On : 20 Dec 2016 4:39
Operator : UM/SJ
Sample : H6125-06RE
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-45(10-12)-12-14-16RE

Quant Time: Dec 20 06:21:08 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF121716.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Dec 17 22:53:54 2016
Response via : Initial Calibration

Manual Integrations
APPROVED

sohil
12/20/2016 6:32:44 PM

