

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF121917\
 Data File : BF101520.D
 Acq On : 19 Dec 2017 14:17
 Operator : SJ/JU
 Sample : I6877-03RE 2X
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-7(0-2)RE

Manual Integrations
 APPROVED

Sohil
 12/20/2017 3:47:45 PM

Quant Time: Dec 19 15:10:49 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF121417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Dec 18 15:38:59 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.81	152	151076	20.00	ng	0.00
21) Naphthalene-d8	8.09	136	591361	20.00	ng	0.00
38) Acenaphthene-d10	9.85	164	250330	20.00	ng	0.00
63) Phenanthrene-d10	11.34	188	397710	20.00	ng	0.00
75) Chrysene-d12	13.99	240	270253	20.00	ng	0.00
86) Perylene-d12	15.47	264	347607	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.43	112	166595	16.43	ng	0.01
7) Phenol-d6	6.45	99	575280	45.58	ng	0.00
23) Nitrobenzene-d5	7.37	82	421138	39.64	ng	0.00
41) 2,4,6-Tribromophenol	10.65	330	1532	0.64	ng	0.00
44) 2-Fluorobiphenyl	9.16	172	772530	47.87	ng	0.00
78) Terphenyl-d14	12.92	244	502361	44.81	ng	0.00
Target Compounds						Qvalue
10) Phenol	6.46	94	33094	2.300	ng	76
31) Naphthalene	8.11	128	109382	3.674	ng	98
37) 2-Methylnaphthalene	8.80	142	59575	3.071	ng	100
49) Dimethylphthalate	9.56	163	77502	3.849	ng	99
51) Acenaphthene	9.88	154	83542	5.183	ng	97
54) Dibenzofuran	10.05	168	76915	3.755	ng	# 94
57) Fluorene	10.40	166	80103	4.622	ng	99
70) Phenanthrene	11.36	178	848353	38.357	ng	99
71) Anthracene	11.42	178	272893	12.079	ng	100
72) Carbazole	11.57	167	79577	3.762	ng	97
74) Fluoranthene	12.56	202	1164354	50.155	ng	97
77) Pylene	12.79	202	1082953	53.394	ng	99
80) Benzo(a)anthracene	13.98	228	638111	35.758	ng	99
82) Chrysene	14.02	228	543382	32.250	ng	98
83) Bis(2-ethylhexyl)phthalate	13.95	149	28700	2.469	ng	# 96
85) Indeno(1,2,3-cd)pyrene	16.96	276	400803	26.713	ng	# 91
87) Benzo(b)fluoranthene	15.04	252	810650m	38.261	ng	
88) Benzo(k)fluoranthene	15.06	252	213589m	10.368	ng	
89) Benzo(a)pyrene	15.41	252	579727	29.681	ng	96
90) Dibenzo(a,h)anthracene	16.94	278	109621m	6.537	ng	
91) Benzo(g,h,i)perylene	17.41	276	378394	22.787	ng	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA F\DATA\BF121917\
Data File : BF101520.D
Acq On : 19 Dec 2017 14:17
Operator : SJ/JU
Sample : I6877-03RE 2X
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SB-7(0-2)RE

Manual Integrations
APPROVED

Sohil
12/20/2017 3:47:45 PM

Quant Time: Dec 19 15:10:49 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF121417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Dec 18 15:38:59 2017
Response via : Initial Calibration

