

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121922\
 Data File : BF131686.D
 Acq On : 19 Dec 2022 12:54
 Operator : CG\JU
 Sample : SP6088
 Misc : 8270/625 SURROGATE
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SP6088

Quant Time: Dec 20 01:02:41 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF120822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 20 00:23:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.875	152	60653	20.000	ng	0.00
21) Naphthalene-d8	8.163	136	220622	20.000	ng	# 0.00
39) Acenaphthene-d10	9.927	164	129817	20.000	ng	0.00
64) Phenanthrene-d10	11.421	188	256530	20.000	ng	0.00
76) Chrysene-d12	14.074	240	177631	20.000	ng	# 0.00
86) Perylene-d12	15.574	264	136012	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.498	112	654510	180.026	ng	0.02
7) Phenol-d6	6.504	99	845957	177.563	ng	0.00
23) Nitrobenzene-d5	7.439	82	594519	101.879	ng	0.00
42) 2,4,6-Tribromophenol	10.722	330	250706	173.823	ng	0.00
45) 2-Fluorobiphenyl	9.245	172	1065984	103.536	ng	0.00
79) Terphenyl-d14	13.015	244	1294025	99.905	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121922\
Data File : BF131686.D
Acq On : 19 Dec 2022 12:54
Operator : CG\JU
Sample : SP6088
Misc : 8270/625 SURROGATE
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SP6088

Quant Time: Dec 20 01:02:41 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF120822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 20 00:23:43 2022
Response via : Initial Calibration

