


Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121922\
 Data File : BF131683.D
 Acq On : 19 Dec 2022 11:09
 Operator : CG\JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF120822.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Dec 20 00:23:43 2022






AutoFind: Scans 1628, 1629, 1630; Background Corrected with Scan 1624

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.1	40344	PASS
68	69	0.00	2	1.6	731	PASS
69	198	0.00	100	41.2	44739	PASS
70	69	0.00	2	0.7	302	PASS
127	198	10	80	47.9	52032	PASS
197	198	0.00	2	0.6	666	PASS
198	198	100	100	100.0	108659	PASS
199	198	5	9	6.7	7322	PASS
275	198	10	60	26.5	28828	PASS
365	198	1	100	3.2	3530	PASS
441	198	0.01	100	11.0	11949	PASS
442	442	50	100	100.0	79218	PASS
443	442	15	24	19.2	15228	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121922\
Data File : BF131683.D
Acq On : 19 Dec 2022 11:09
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Dec 20 03:52:15 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF120822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 20 00:23:43 2022
Response via : Initial Calibration



TIC: BF131683.D\data.ms

(70) Pentachlorophenol (C)

11.221min (-0.000) 27120.43 ng




response 92670

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	64.40	62.43
264.00	63.40	62.85
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121922\
Data File : BF131683.D
Acq On : 19 Dec 2022 11:09
Operator : CG\JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
DFTPP

Quant Time: Dec 20 03:52:15 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF120822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 20 00:23:43 2022
Response via : Initial Calibration



TIC: BF131683.D\data.ms

(77) Benzidine

12.768min (+ 0.006) 0.00 ng m

response 503116

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.20	14.14
183.00	11.60	11.41
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
12/19/2022	BNA_F	<u>BF131683.D</u>
Compound Name	Response	Retention Time
DDT	252864	13.592
DDD	3570	13.321
DDE	169	12.951
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
3739	256603	1.46

Instrument :
BNA_F
ClientSampleId :
DFTPP