

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121923\
 Data File : BF136639.D
 Acq On : 19 Dec 2023 19:46
 Operator : CG\JU
 Sample : PB157746BL
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB157746BL

Quant Time: Dec 20 04:31:56 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF121923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 20 04:16:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.792	152	78888	20.000	ng	0.00
21) Naphthalene-d8	8.069	136	305826	20.000	ng	0.00
39) Acenaphthene-d10	9.822	164	164436	20.000	ng	0.00
64) Phenanthrene-d10	11.316	188	355489	20.000	ng	0.00
76) Chrysene-d12	13.968	240	256743	20.000	ng	0.00
86) Perylene-d12	15.445	264	207398	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.445	112	630755	124.138	ng	0.01
7) Phenol-d6	6.439	99	791641	121.731	ng	0.00
23) Nitrobenzene-d5	7.357	82	505609	85.852	ng	0.00
42) 2,4,6-Tribromophenol	10.622	330	251377	131.096	ng	0.00
45) 2-Fluorobiphenyl	9.151	172	1017901	84.575	ng	0.00
79) Terphenyl-d14	12.916	244	1337666	73.367	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121923\
Data File : BF136639.D
Acq On : 19 Dec 2023 19:46
Operator : CG\JU
Sample : PB157746BL
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB157746BL

Quant Time: Dec 20 04:31:56 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF121923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 20 04:16:58 2023
Response via : Initial Calibration

