

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121923\
 Data File : BF136638.D
 Acq On : 19 Dec 2023 19:15
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF121923

Quant Time: Dec 20 04:18:03 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF121923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 20 04:16:58 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	89	0.00
2	1,4-Dioxane	40.000	40.149	-0.4	91	0.00
3	Pyridine	40.000	41.588	-4.0	95	0.00
4	n-Nitrosodimethylamine	40.000	42.172	-5.4	95	0.00
5 S	2-Fluorophenol	80.000	82.325	-2.9	93	0.00
6	Aniline	40.000	41.685	-4.2	94	0.00
7 S	Phenol-d6	80.000	81.636	-2.0	92	0.00
8	2-Chlorophenol	40.000	41.342	-3.4	92	0.00
9	Benzaldehyde	40.000	38.631	3.4	93	0.00
10 C	Phenol	40.000	41.007	-2.5	94	0.00
11	bis(2-Chloroethyl)ether	40.000	40.142	-0.4	87	0.00
12	1,3-Dichlorobenzene	40.000	41.106	-2.8	93	0.00
13 C	1,4-Dichlorobenzene	40.000	41.149	-2.9	93	0.00
14	1,2-Dichlorobenzene	40.000	41.022	-2.6	93	0.00
15	Benzyl Alcohol	40.000	41.618	-4.0	92	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.190	-3.0	91	0.00
17	2-Methylphenol	40.000	40.866	-2.2	91	0.00
18	Hexachloroethane	40.000	41.549	-3.9	92	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.411	-1.0	90	0.00
20	3+4-Methylphenols	40.000	40.745	-1.9	92	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	89	0.00
22	Acetophenone	40.000	40.399	-1.0	91	0.00
23 S	Nitrobenzene-d5	80.000	82.357	-2.9	91	0.00
24	Nitrobenzene	40.000	41.661	-4.2	92	0.00
25	Isophorone	40.000	40.394	-1.0	89	0.00
26 C	2-Nitrophenol	40.000	42.381	-6.0	92	0.00
27	2,4-Dimethylphenol	40.000	39.989	0.0	90	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.564	-1.4	89	0.00
29 C	2,4-Dichlorophenol	40.000	41.913	-4.8	92	0.00
30	1,2,4-Trichlorobenzene	40.000	41.656	-4.1	92	0.00
31	Naphthalene	40.000	40.662	-1.7	90	0.00
32	Benzoic acid	40.000	45.098	-12.7	96	0.00
33	4-Chloroaniline	40.000	40.080	-0.2	88	0.00
34 C	Hexachlorobutadiene	40.000	40.743	-1.9	89	0.00
35	Caprolactam	40.000	40.860	-2.1	94	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.962	-2.4	89	0.00
37	2-Methylnaphthalene	40.000	40.459	-1.1	89	0.00
38	1-Methylnaphthalene	40.000	40.145	-0.4	89	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	90	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.297	-3.2	90	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.895	2.8	84	0.00
42 S	2,4,6-Tribromophenol	80.000	80.939	-1.2	91	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.054	-2.6	89	0.00
44	2,4,5-Trichlorophenol	40.000	40.304	-0.8	87	0.00
45 S	2-Fluorobiphenyl	80.000	80.781	-1.0	91	0.00
46	1,1'-Biphenyl	40.000	40.434	-1.1	89	0.00
47	2-Chloronaphthalene	40.000	40.431	-1.1	89	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121923\
 Data File : BF136638.D
 Acq On : 19 Dec 2023 19:15
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF121923

Quant Time: Dec 20 04:18:03 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF121923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 20 04:16:58 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.466	-3.7	92	0.00
49	Acenaphthylene	40.000	40.831	-2.1	90	0.00
50	Dimethylphthalate	40.000	41.060	-2.7	93	0.00
51	2,6-Dinitrotoluene	40.000	41.385	-3.5	94	0.00
52 C	Acenaphthene	40.000	40.504	-1.3	91	0.00
53	3-Nitroaniline	40.000	40.623	-1.6	92	0.00
54 P	2,4-Dinitrophenol	40.000	43.014	-7.5	92	0.00
55	Dibenzofuran	40.000	40.466	-1.2	90	0.00
56 P	4-Nitrophenol	40.000	42.245	-5.6	92	0.00
57	2,4-Dinitrotoluene	40.000	41.442	-3.6	92	0.00
58	Fluorene	40.000	39.826	0.4	90	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.618	-1.5	91	0.00
60	Diethylphthalate	40.000	40.484	-1.2	92	0.00
61	4-Chlorophenyl-phenylether	40.000	40.457	-1.1	91	0.00
62	4-Nitroaniline	40.000	40.987	-2.5	93	0.00
63	Azobenzene	40.000	40.831	-2.1	90	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	93	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.369	-5.9	90	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.744	-1.9	93	0.00
67	4-Bromophenyl-phenylether	40.000	40.521	-1.3	90	0.00
68	Hexachlorobenzene	40.000	41.119	-2.8	92	0.00
69	Atrazine	40.000	33.579	16.1	90	0.00
70 C	Pentachlorophenol	40.000	41.852	-4.6	89	0.00
71	Phenanthrene	40.000	40.317	-0.8	94	0.00
72	Anthracene	40.000	39.932	0.2	92	0.00
73	Carbazole	40.000	40.513	-1.3	95	0.00
74	Di-n-butylphthalate	40.000	40.986	-2.5	95	0.00
75 C	Fluoranthene	40.000	40.143	-0.4	96	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	98	0.00
77	Benzydine	40.000	34.984	12.5	88	0.00
78	Pyrene	40.000	41.615	-4.0	94	0.00
79 S	Terphenyl-d14	80.000	83.445	-4.3	97	0.00
80	Butylbenzylphthalate	40.000	42.960	-7.4	99	0.00
81	Benzo(a)anthracene	40.000	40.842	-2.1	98	0.00
82	3,3'-Dichlorobenzidine	40.000	40.139	-0.3	95	0.00
83	Chrysene	40.000	41.413	-3.5	99	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.625	-6.6	97	0.00
85 c	Di-n-octyl phthalate	40.000	43.436	-8.6	98	0.00
86 I	Perylene-d12	20.000	20.000	0.0	99	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	43.558	-8.9	98	0.00
88	Benzo(b)fluoranthene	40.000	39.109	2.2	93	0.00
89	Benzo(k)fluoranthene	40.000	40.706	-1.8	103	0.00
90 C	Benzo(a)pyrene	40.000	40.459	-1.1	99	0.00
91	Dibenzo(a,h)anthracene	40.000	43.954	-9.9	99	0.00
92	Benzo(g,h,i)perylene	40.000	44.083	-10.2	98	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF121923\
Data File : BF136638.D
Acq On : 19 Dec 2023 19:15
Operator : CG\JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF121923

Quant Time: Dec 20 04:18:03 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF121923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 20 04:16:58 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0