

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF122017\
 Data File : BF101560.D
 Acq On : 20 Dec 2017 11:46
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 20 14:26:25 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF121417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 20 14:25:04 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
2	1,4-Dioxane	0.690	0.672	2.6	85	0.00
3	Pyridine	1.917	1.874	2.2	84	0.00
4	n-Nitrosodimethylamine	0.883	0.847	4.1	82	0.00
5 S	2-Fluorophenol	1.343	1.354	-0.8	88	0.00
6	Aniline	2.322	2.190	5.7	84	0.00
7 S	Phenol-d6	1.671	1.557	6.8	85	0.00
8	2-Chlorophenol	1.412	1.338	5.2	85	0.00
9	Benzaldehyde	1.234	1.149	6.9	82	0.00
10 C	Phenol	1.905	1.827	4.1	86	0.00
11	bis(2-Chloroethyl)ether	1.494	1.440	3.6	85	0.00
12	1,3-Dichlorobenzene	1.594	1.532	3.9	86	0.00
13 C	1,4-Dichlorobenzene	1.628	1.567	3.7	87	0.00
14	1,2-Dichlorobenzene	1.465	1.406	4.0	86	0.00
15	Benzyl Alcohol	1.150	1.060	7.8	83	0.00
16	2,2'-oxybis(1-Chloropropane	2.286	2.159	5.6	83	0.00
17	2-Methylphenol	1.299	1.199	7.7	86	0.00
18	Hexachloroethane	0.566	0.550	2.8	87	0.00
19 P	n-Nitroso-di-n-propylamine	1.097	1.020	7.0	88	0.00
20	3+4-Methylphenols	1.377	1.391	-1.0	90	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	88	0.00
22	Acetophenone	0.456	0.438	3.9	86	0.00
23 S	Nitrobenzene-d5	0.359	0.356	0.8	87	0.00
24	Nitrobenzene	0.398	0.385	3.3	87	0.00
25	Isophorone	0.721	0.677	6.1	85	0.00
26 C	2-Nitrophenol	0.171	0.181	-5.8	90	0.00
27	2,4-Dimethylphenol	0.290	0.270	6.9	85	0.00
28	bis(2-Chloroethoxy)methane	0.458	0.431	5.9	86	0.00
29 C	2,4-Dichlorophenol	0.287	0.281	2.1	87	0.00
30	1,2,4-Trichlorobenzene	0.303	0.290	4.3	86	0.00
31	Naphthalene	1.007	0.945	6.2	86	0.00
32	Benzoic acid	0.173	0.185	-6.9	97	0.00
33	4-Chloroaniline	0.438	0.409	6.6	85	0.00
34 C	Hexachlorobutadiene	0.175	0.170	2.9	87	0.00
35	Caprolactam	0.100	0.092	8.0	82	0.00
36 C	4-Chloro-3-methylphenol	0.315	0.301	4.4	85	0.00
37	2-Methylnaphthalene	0.656	0.610	7.0	85	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	88	0.00
39	1,2,4,5-Tetrachlorobenzene	0.675	0.646	4.3	87	0.00
40 P	Hexachlorocyclopentadiene	0.088	0.093	-5.7	98	0.00
41 S	2,4,6-Tribromophenol	0.193	0.191	1.0	88	0.00
42 C	2,4,6-Trichlorophenol	0.407	0.392	3.7	84	0.00
43	2,4,5-Trichlorophenol	0.445	0.431	3.1	85	0.00
44 S	2-Fluorobiphenyl	1.289	1.236	4.1	86	0.00

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF122017\
 Data File : BF101560.D
 Acq On : 20 Dec 2017 11:46
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 20 14:26:25 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF121417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 20 14:25:04 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.774	1.663	6.3	87	0.00
46	2-Chloronaphthalene	1.357	1.270	6.4	87	0.00
47	2-Nitroaniline	0.469	0.473	-0.9	89	0.00
48	Acenaphthylene	2.144	1.923	10.3	84	0.00
49	Dimethylphthalate	1.609	1.432	11.0	83	0.00
50	2,6-Dinitrotoluene	0.327	0.312	4.6	83	0.00
51 C	Acenaphthene	1.288	1.177	8.6	85	0.00
52	3-Nitroaniline	0.408	0.404	1.0	87	0.00
53 P	2,4-Dinitrophenol	0.087	0.098	-12.6	98	0.00
54	Dibenzofuran	1.637	1.580	3.5	87	0.00
55 P	4-Nitrophenol	0.178	0.147	17.4	70	0.00
56	2,4-Dinitrotoluene	0.368	0.377	-2.4	90	0.00
57	Fluorene	1.385	1.337	3.5	87	0.00
58	2,3,4,6-Tetrachlorophenol	0.320	0.308	3.8	85	0.00
59	Diethylphthalate	1.593	1.481	7.0	88	0.00
60	4-Chlorophenyl-phenylether	0.664	0.649	2.3	86	0.00
61	4-Nitroaniline	0.410	0.380	7.3	83	0.00
62	Azobenzene	1.650	1.537	6.8	88	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	87	0.00
64	4,6-Dinitro-2-methylphenol	0.102	0.121	-18.6	98	0.00
65 c	n-Nitrosodiphenylamine	0.738	0.696	5.7	86	0.00
66	4-Bromophenyl-phenylether	0.225	0.219	2.7	87	0.00
67	Hexachlorobenzene	0.238	0.233	2.1	88	0.00
68	Atrazine	0.220	0.209	5.0	84	0.00
69 C	Pentachlorophenol	0.079	0.078	1.3	86	0.00
70	Phenanthrene	1.112	1.040	6.5	87	0.00
71	Anthracene	1.136	1.061	6.6	86	0.00
72	Carbazole	1.064	0.998	6.2	86	0.00
73	Di-n-butylphthalate	1.291	1.247	3.4	89	0.00
74 C	Fluoranthene	1.167	1.092	6.4	86	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	92	0.00
76	Benzidine	0.845	0.777	8.0	85	0.00
77	Pyrene	1.501	1.364	9.1	87	0.00
78 S	Terphenyl-d14	0.830	0.758	8.7	90	0.00
79	Butylbenzylphthalate	0.679	0.650	4.3	89	0.00
80	Benzo(a)anthracene	1.321	1.215	8.0	88	0.00
81	3,3'-Dichlorobenzidine	0.431	0.414	3.9	90	0.00
82	Chrysene	1.247	1.145	8.2	88	0.00
83	Bis(2-ethylhexyl)phthalate	0.860	0.827	3.8	91	0.00
84 c	Di-n-octyl phthalate	1.534	1.506	1.8	93	0.00
85	Indeno(1,2,3-cd)pyrene	1.110	0.957	13.8	85	0.00
86 I	Perylene-d12	1.000	1.000	0.0	89	0.00
87	Benzo(b)fluoranthene	1.219	1.206	1.1	88	0.00

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF122017\
Data File : BF101560.D
Acq On : 20 Dec 2017 11:46
Operator : SJ/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Dec 20 14:26:25 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF121417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 20 14:25:04 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.185	1.039	12.3	83	0.00
89 C	Benzo(a)pyrene	1.124	1.041	7.4	85	0.00
90	Dibenzo(a,h)anthracene	0.965	0.858	11.1	84	0.00
91	Benzo(g,h,i)perylene	0.955	0.857	10.3	84	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0