

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF122017\
 Data File : BF101581.D
 Acq On : 20 Dec 2017 21:20
 Operator : SJ/JU
 Sample : PB105164TB
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB105164TB

Quant Time: Dec 21 04:47:42 2017
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF121417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 20 14:25:04 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.79	152	151542	20.00	ng	0.00
21) Naphthalene-d8	8.07	136	610692	20.00	ng	0.00
38) Acenaphthene-d10	9.83	164	263926	20.00	ng	0.00
63) Phenanthrene-d10	11.32	188	444137	20.00	ng	0.00
75) Chrysene-d12	13.97	240	276160	20.00	ng	0.00
86) Perylene-d12	15.43	264	308355	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.43	112	1245333	122.41	ng	0.02
7) Phenol-d6	6.44	99	1509531	119.22	ng	0.00
23) Nitrobenzene-d5	7.36	82	997903	90.96	ng	0.00
41) 2,4,6-Tribromophenol	10.63	330	277830	109.25	ng	0.00
44) 2-Fluorobiphenyl	9.15	172	1575007	92.57	ng	0.00
78) Terphenyl-d14	12.90	244	1164550	101.66	ng	0.00


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF122017\
Data File : BF101581.D
Acq On : 20 Dec 2017 21:20
Operator : SJ/JU
Sample : PB105164TB
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB105164TB

Quant Time: Dec 21 04:47:42 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF121417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 20 14:25:04 2017
Response via : Initial Calibration









#1

1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.79 min Scan# 800
Delta R.T. -0.00 min
Lab File: BF101581.D
Acq: 20 Dec 2017 21:20

Instrument :
BNA_F
ClientSampled :
PB105164TB




Tgt Ion:152 Resp: 151542
Ion Ratio Lower Upper
152 100
150 162.2 124.6 186.8
115 65.5 50.1 75.1




#5

2-Fluorophenol
Concen: 122.410 ng
RT: 5.43 min Scan# 568
Delta R.T. 0.02 min
Lab File: BF101581.D
Acq: 20 Dec 2017 21:20

Tgt Ion:112 Resp: 1245333
Ion Ratio Lower Upper
112 100
64 61.1 47.9 71.9
63 31.4 23.7 35.5





#7

Phenol-d6

Concen: 119.224 ng

RT: 6.44 min Scan# 740

Delta R.T. -0.00 min

Lab File: BF101581.D

Acq: 20 Dec 2017 21:20

Instrument :

BNA_F

ClientSampleId :

PB105164TB


Tgt Ion: 99 Resp: 1509531

Ion Ratio Lower Upper

99 100

42 20.5 14.5 21.7




71 33.8 25.4 38.0



Abundance Ion 99.00 (98.70 to 99.70): BF1

Ion 42.00 (41.70 to 42.70): BF1

Ion 71.00 (70.70 to 71.70): BF1



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.07 min Scan# 1018

Delta R.T. -0.01 min

Lab File: BF101581.D

Acq: 20 Dec 2017 21:20

Tgt Ion: 136 Resp: 610692


Ion Ratio Lower Upper

136 100

137 11.3 8.9 13.3

54 9.5 6.6 9.8

68 6.9 5.0 7.4






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF1

Ion 68.00 (67.70 to 68.70): BF1





#23

Nitrobenzene-d5

Concen: 90.961 ng

RT: 7.36 min Scan# 897

Delta R.T. -0.01 min

Lab File: BF101581.D

Acq: 20 Dec 2017 21:20

Instrument :

BNA_F

ClientSampleId :

PB105164TB


Tgt Ion: 82 Resp: 997903

Ion Ratio Lower Upper

82 100

128 43.4 36.1 54.1




54 55.9 41.4 62.2



Abundance Ion 82.00 (81.70 to 82.70): BF1

Ion 128.00 (127.70 to 128.70): BF1

Ion 54.00 (53.70 to 54.70): BF1



#38

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.83 min Scan# 1316

Delta R.T. -0.01 min

Lab File: BF101581.D

Acq: 20 Dec 2017 21:20


Tgt Ion: 164 Resp: 263926

Ion Ratio Lower Upper

164 100

162 102.2 86.1 129.1



160 47.4 38.3 57.5




Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E









#41
 2,4,6-Tribromophenol
 Concen: 109.247 ng
 RT: 10.63 min Scan# 1452
 Delta R.T. -0.01 min
 Lab File: BF101581.D
 Acq: 20 Dec 2017 21:20

Instrument :
 BNA_F
 ClientSampleId :
 PB105164TB

Tgt Ion	Ratio	Lower	Upper
330	100		
332	95.3	77.0	115.6
141	33.4	29.9	44.9




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#44
 2-Fluorobiphenyl
 Concen: 92.572 ng
 RT: 9.15 min Scan# 1201
 Delta R.T. -0.00 min
 Lab File: BF101581.D
 Acq: 20 Dec 2017 21:20

Tgt Ion	Ratio	Lower	Upper
172	100		
171	36.1	29.7	44.5
170	24.3	19.6	29.4



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E









#63
Phenanthrene-d10
Concen: 20.000 ng
RT: 11.32 min Scan# 1570
Delta R.T. -0.01 min
Lab File: BF101581.D
Acq: 20 Dec 2017 21:20




Instrument :
BNA_F
ClientSampleId :
PB105164TB

Tgt Ion:188 Resp: 444137
Ion Ratio Lower Upper
188 100
94 10.4 8.5 12.7
80 11.3 9.2 13.8



#75
Chrysene-d12
Concen: 20.000 ng
RT: 13.97 min Scan# 2020
Delta R.T. -0.01 min
Lab File: BF101581.D
Acq: 20 Dec 2017 21:20

Tgt Ion:240 Resp: 276160
Ion Ratio Lower Upper
240 100
120 12.6 9.5 14.3
236 25.8 20.7 31.1





#78

Terphenyl-d14

Concen: 101.661 ng

RT: 12.90 min Scan# 1839

Delta R.T. -0.01 min

Lab File: BF101581.D

Acq: 20 Dec 2017 21:20

Instrument :

BNA_F

ClientSampleId :

PB105164TB


Tgt Ion:244 Resp: 1164550

Ion Ratio Lower Upper

244 100

212 7.1 5.4 8.0




122 11.5 11.0 16.4



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#86

Perylene-d12

Concen: 20.000 ng

RT: 15.43 min Scan# 2269

Delta R.T. -0.00 min

Lab File: BF101581.D

Acq: 20 Dec 2017 21:20


Tgt Ion:264 Resp: 308355

Ion Ratio Lower Upper

264 100

260 24.6 19.2 28.8

265 21.7 17.5 26.3



Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E

