

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF122023\
 Data File : BF136649.D
 Acq On : 20 Dec 2023 15:50
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF122023

Quant Time: Dec 21 01:58:11 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 21 01:54:25 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	1,4-Dioxane	0.534	0.522	2.2	102	0.00
3	Pyridine	1.303	1.328	-1.9	102	0.00
4	n-Nitrosodimethylamine	0.696	0.713	-2.4	101	0.00
5 S	2-Fluorophenol	1.282	1.280	0.2	100	0.00
6	Aniline	1.661	1.639	1.3	99	0.00
7 S	Phenol-d6	1.661	1.639	1.3	99	0.00
8	2-Chlorophenol	1.403	1.392	0.8	99	0.00
9	Benzaldehyde	0.945	0.925	2.1	100	0.00
10 C	Phenol	1.773	1.746	1.5	100	0.00
11	bis(2-Chloroethyl)ether	1.348	1.328	1.5	98	0.00
12	1,3-Dichlorobenzene	1.530	1.507	1.5	99	0.00
13 C	1,4-Dichlorobenzene	1.564	1.552	0.8	100	0.00
14	1,2-Dichlorobenzene	1.455	1.447	0.5	100	0.00
15	Benzyl Alcohol	1.121	1.135	-1.2	100	0.00
16	2,2'-oxybis(1-Chloropropane	2.205	2.180	1.1	99	0.00
17	2-Methylphenol	1.162	1.166	-0.3	102	0.00
18	Hexachloroethane	0.568	0.564	0.7	101	0.00
19 P	n-Nitroso-di-n-propylamine	1.007	1.011	-0.4	102	0.00
20	3+4-Methylphenols	1.452	1.411	2.8	99	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
22	Acetophenone	0.501	0.497	0.8	99	0.00
23 S	Nitrobenzene-d5	0.391	0.391	0.0	98	0.00
24	Nitrobenzene	0.392	0.400	-2.0	100	0.00
25	Isophorone	0.711	0.726	-2.1	101	0.00
26 C	2-Nitrophenol	0.187	0.193	-3.2	99	0.00
27	2,4-Dimethylphenol	0.228	0.234	-2.6	101	0.00
28	bis(2-Chloroethoxy)methane	0.432	0.439	-1.6	99	0.00
29 C	2,4-Dichlorophenol	0.294	0.297	-1.0	99	0.00
30	1,2,4-Trichlorobenzene	0.327	0.331	-1.2	100	0.00
31	Naphthalene	1.099	1.101	-0.2	99	0.00
32	Benzoic acid	0.221	0.243	-10.0	100	0.00
33	4-Chloroaniline	0.406	0.406	0.0	99	0.00
34 C	Hexachlorobutadiene	0.199	0.200	-0.5	100	0.00
35	Caprolactam	0.097	0.101	-4.1	101	0.00
36 C	4-Chloro-3-methylphenol	0.331	0.339	-2.4	101	0.00
37	2-Methylnaphthalene	0.707	0.717	-1.4	101	0.00
38	1-Methylnaphthalene	0.694	0.694	0.0	101	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
40	1,2,4,5-Tetrachlorobenzene	0.619	0.626	-1.1	100	0.00
41 P	Hexachlorocyclopentadiene	0.172	0.179	-4.1	99	0.00
42 S	2,4,6-Tribromophenol	0.236	0.242	-2.5	102	0.00
43 C	2,4,6-Trichlorophenol	0.397	0.405	-2.0	100	0.00
44	2,4,5-Trichlorophenol	0.443	0.463	-4.5	103	0.00
45 S	2-Fluorobiphenyl	1.487	1.477	0.7	101	0.00
46	1,1'-Biphenyl	1.678	1.701	-1.4	102	0.00
47	2-Chloronaphthalene	1.276	1.265	0.9	99	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF122023\
 Data File : BF136649.D
 Acq On : 20 Dec 2023 15:50
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF122023

Quant Time: Dec 21 01:58:11 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 21 01:54:25 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.400	0.411	-2.7	101	0.00
49	Acenaphthylene	1.950	1.986	-1.8	101	0.00
50	Dimethylphthalate	1.461	1.466	-0.3	101	0.00
51	2,6-Dinitrotoluene	0.332	0.345	-3.9	102	0.00
52 C	Acenaphthene	1.209	1.212	-0.2	100	0.00
53	3-Nitroaniline	0.351	0.366	-4.3	101	0.00
54 P	2,4-Dinitrophenol	0.165	0.181	-9.7	99	0.00
55	Dibenzofuran	1.832	1.854	-1.2	102	0.00
56 P	4-Nitrophenol	0.237	0.254	-7.2	101	0.00
57	2,4-Dinitrotoluene	0.430	0.450	-4.7	102	0.00
58	Fluorene	1.454	1.464	-0.7	102	0.00
59	2,3,4,6-Tetrachlorophenol	0.341	0.363	-6.5	103	0.00
60	Diethylphthalate	1.516	1.543	-1.8	101	0.00
61	4-Chlorophenyl-phenylether	0.707	0.705	0.3	100	0.00
62	4-Nitroaniline	0.339	0.342	-0.9	103	0.00
63	Azobenzene	1.416	1.438	-1.6	102	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	102	0.00
65	4,6-Dinitro-2-methylphenol	0.115	0.127	-10.4	101	0.00
66 c	n-Nitrosodiphenylamine	0.650	0.666	-2.5	101	0.00
67	4-Bromophenyl-phenylether	0.222	0.229	-3.2	103	0.00
68	Hexachlorobenzene	0.248	0.249	-0.4	100	0.00
69	Atrazine	0.166	0.141	15.1	101	0.00
70 C	Pentachlorophenol	0.115	0.127	-10.4	101	0.00
71	Phenanthrene	1.026	1.036	-1.0	100	0.00
72	Anthracene	1.039	1.043	-0.4	101	0.00
73	Carbazole	0.980	0.989	-0.9	101	0.00
74	Di-n-butylphthalate	1.196	1.249	-4.4	104	0.00
75 C	Fluoranthene	1.098	1.108	-0.9	102	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	102	0.00
77	Benzydine	0.590	0.507	14.1	92	0.00
78	Pyrene	1.963	2.046	-4.2	102	0.00
79 S	Terphenyl-d14	1.431	1.476	-3.1	101	0.00
80	Butylbenzylphthalate	0.714	0.778	-9.0	105	0.00
81	Benzo(a)anthracene	1.389	1.437	-3.5	102	0.00
82	3,3'-Dichlorobenzidine	0.394	0.401	-1.8	103	0.00
83	Chrysene	1.358	1.393	-2.6	103	0.00
84	Bis(2-ethylhexyl)phthalate	0.899	0.984	-9.5	106	0.00
85 c	Di-n-octyl phthalate	1.390	1.552	-11.7	109	0.00
86 I	Perylene-d12	1.000	1.000	0.0	106	0.00
87	Indeno(1,2,3-cd)pyrene	1.444	1.497	-3.7	105	0.00
88	Benzo(b)fluoranthene	1.336	1.292	3.3	104	0.00
89	Benzo(k)fluoranthene	1.262	1.276	-1.1	104	0.00
90 C	Benzo(a)pyrene	1.128	1.126	0.2	104	0.00
91	Dibenzo(a,h)anthracene	1.185	1.229	-3.7	103	0.00
92	Benzo(g,h,i)perylene	1.213	1.256	-3.5	104	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF122023\
Data File : BF136649.D
Acq On : 20 Dec 2023 15:50
Operator : CG\JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF122023

Quant Time: Dec 21 01:58:11 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 21 01:54:25 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0