

Data Path : Z:\HPCHEM1\BNA F\DATA\BF122116\
 Data File : BF091849.D
 Acq On : 21 Dec 2016 22:29
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 22 03:33:59 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF122116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 21 16:17:51 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2	1,4-Dioxane	0.590	0.666	-12.9	113	0.00
3	Pyridine	2.058	1.892	8.1	91	0.00
4	n-Nitrosodimethylamine	0.776	0.745	4.0	93	0.00
5 S	2-Fluorophenol	1.198	1.319	-10.1	116	0.00
6	Aniline	1.899	1.781	6.2	95	-0.01
7 S	Phenol-d6	1.462	1.412	3.4	96	-0.01
8	2-Chlorophenol	1.362	1.248	8.4	91	-0.01
9	Benzaldehyde	0.904	0.815	9.8	98	0.00
10 C	Phenol	1.666	1.543	7.4	87	0.00
11	bis(2-Chloroethyl)ether	1.326	1.221	7.9	87	-0.01
12	1,3-Dichlorobenzene	1.455	1.390	4.5	92	0.00
13 C	1,4-Dichlorobenzene	1.480	1.360	8.1	92	-0.01
14	1,2-Dichlorobenzene	1.285	1.271	1.1	103	0.00
15	Benzyl Alcohol	1.136	1.006	11.4	89	0.00
16	2,2'-oxybis(1-Chloropropane	1.975	1.811	8.3	93	0.00
17	2-Methylphenol	1.096	1.095	0.1	102	0.00
18	Hexachloroethane	0.549	0.530	3.5	93	0.00
19 P	n-Nitroso-di-n-propylamine	0.973	0.883	9.2	87	-0.01
20	3+4-Methylphenols	1.307	1.216	7.0	90	-0.01
21 I	Naphthalene-d8	1.000	1.000	0.0	98	0.00
22	Acetophenone	0.493	0.480	2.6	94	0.00
23 S	Nitrobenzene-d5	0.360	0.332	7.8	94	0.00
24	Nitrobenzene	0.383	0.377	1.6	87	0.00
25	Isophorone	0.664	0.643	3.2	95	0.00
26 C	2-Nitrophenol	0.189	0.174	7.9	92	0.00
27	2,4-Dimethylphenol	0.313	0.289	7.7	83	-0.01
28	bis(2-Chloroethoxy)methane	0.402	0.371	7.7	88	-0.01
29 C	2,4-Dichlorophenol	0.265	0.270	-1.9	96	0.00
30	1,2,4-Trichlorobenzene	0.291	0.290	0.3	93	0.00
31	Naphthalene	0.915	0.917	-0.2	89	0.00
32	Benzoic acid	0.232	0.221	4.7	88	-0.01
33	4-Chloroaniline	0.413	0.374	9.4	87	0.00
34 C	Hexachlorobutadiene	0.153	0.156	-2.0	97	0.00
35	Caprolactam	0.087	0.087	0.0	95	0.00
36 C	4-Chloro-3-methylphenol	0.305	0.269	11.8	80	-0.01
37	2-Methylnaphthalene	0.628	0.580	7.6	99	0.00
38 I	Acenaphthene-d10	1.000	1.000	0.0	102	0.00
39	1,2,4,5-Tetrachlorobenzene	0.505	0.490	3.0	89	0.00
40 P	Hexachlorocyclopentadiene	0.199	0.194	2.5	86	-0.01
41 S	2,4,6-Tribromophenol	0.166	0.155	6.6	98	0.00
42 C	2,4,6-Trichlorophenol	0.353	0.341	3.4	89	0.00
43	2,4,5-Trichlorophenol	0.364	0.338	7.1	91	-0.01
44 S	2-Fluorobiphenyl	1.042	0.950	8.8	96	0.00

Data Path : Z:\HPCHEM1\BNA F\DATA\BF122116\
 Data File : BF091849.D
 Acq On : 21 Dec 2016 22:29
 Operator : UM/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 22 03:33:59 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF122116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 21 16:17:51 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	1.369	1.429	-4.4	94	0.00
46	2-Chloronaphthalene	1.084	1.070	1.3	96	0.00
47	2-Nitroaniline	0.386	0.339	12.2	79	-0.01
48	Acenaphthylene	1.668	1.645	1.4	100	0.00
49	Dimethylphthalate	1.279	1.272	0.5	97	0.00
50	2,6-Dinitrotoluene	0.280	0.297	-6.1	107	0.00
51 C	Acenaphthene	1.131	1.054	6.8	96	0.00
52	3-Nitroaniline	0.346	0.347	-0.3	89	0.00
53 P	2,4-Dinitrophenol	0.156	0.159	-1.9	91	0.00
54	Dibenzofuran	1.527	1.403	8.1	101	0.00
55 P	4-Nitrophenol	0.235	0.251	-6.8	99	0.00
56	2,4-Dinitrotoluene	0.351	0.338	3.7	100	0.00
57	Fluorene	1.111	1.111	0.0	98	0.00
58	2,3,4,6-Tetrachlorophenol	0.288	0.265	8.0	87	0.00
59	Diethylphthalate	1.242	1.090	12.2	82	-0.01
60	4-Chlorophenyl-phenylether	0.486	0.433	10.9	91	0.00
61	4-Nitroaniline	0.359	0.309	13.9	78	-0.01
62	Azobenzene	1.368	1.241	9.3	93	0.00
63 I	Phenanthrene-d10	1.000	1.000	0.0	95	0.00
64	4,6-Dinitro-2-methylphenol	0.121	0.149	-23.1	109	0.00
65 c	n-Nitrosodiphenylamine	0.593	0.639	-7.8	103	0.00
66	4-Bromophenyl-phenylether	0.208	0.212	-1.9	98	0.00
67	Hexachlorobenzene	0.222	0.229	-3.2	96	0.00
68	Atrazine	0.191	0.156	18.3	78	-0.01
69 C	Pentachlorophenol	0.109	0.118	-8.3	92	0.00
70	Phenanthrene	1.084	1.044	3.7	88	0.00
71	Anthracene	1.086	1.006	7.4	98	0.00
72	Carbazole	1.005	1.028	-2.3	106	0.00
73	Di-n-butylphthalate	1.174	1.046	10.9	91	0.00
74 C	Fluoranthene	1.082	0.991	8.4	97	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	102	0.00
76	Benzidine	0.580	0.508	12.4	94	0.00
77	Pyrene	1.467	1.352	7.8	101	0.00
78 S	Terphenyl-d14	0.825	0.676	18.1	92	0.00
79	Butylbenzylphthalate	0.667	0.622	6.7	89	0.00
80	Benzo(a)anthracene	1.129	0.973	13.8	89	0.00
81	3,3'-Dichlorobenzidine	0.428	0.395	7.7	100	0.00
82	Chrysene	1.132	0.987	12.8	98	0.00
83	Bis(2-ethylhexyl)phthalate	0.786	0.719	8.5	94	0.00
84 c	Di-n-octyl phthalate	1.456	1.274	12.5	90	0.00
85	Indeno(1,2,3-cd)pyrene	0.981	0.866	11.7	93	0.00
86 I	Perylene-d12	1.000	1.000	0.0	99	0.00
87	Benzo(b)fluoranthene	1.245	1.375	-10.4	103	0.00

Data Path : Z:\HPCHEM1\BNA F\DATA\BF122116\
Data File : BF091849.D
Acq On : 21 Dec 2016 22:29
Operator : UM/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Dec 22 03:33:59 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF122116.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 21 16:17:51 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	1.062	0.873	17.8	84	0.00
89 C	Benzo(a)pyrene	1.105	1.068	3.3	94	0.00
90	Dibenzo(a,h)anthracene	0.908	0.872	4.0	94	-0.01
91	Benzo(a,h,i)perylene	0.945	0.891	5.7	92	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0