


Data Path : Z:\HPCHEM1\BNA F\DATA\BF122116\  
 Data File : BF091848.D  
 Acq On : 21 Dec 2016 22:02  
 Operator : UM/SJ  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF122116.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Wed Dec 21 16:17:51 2016



AutoFind: Scans 825, 826, 827; Background Corrected with Scan 822

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 38.2      | 19687   | PASS             |
| 68          | 69           | 0.00         | 2            | 1.3       | 249     | PASS             |
| 69          | 198          | 0.00         | 100          | 37.8      | 19488   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 127         | 198          | 10           | 80           | 49.0      | 25301   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.3       | 141     | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 51601   | PASS             |
| 199         | 198          | 5            | 9            | 6.1       | 3148    | PASS             |
| 275         | 198          | 10           | 60           | 22.7      | 11722   | PASS             |
| 365         | 198          | 1            | 100          | 2.5       | 1314    | PASS             |
| 441         | 442          | 0.01         | 24           | 16.3      | 6675    | PASS             |
| 442         | 198          | 50           | 100          | 79.5      | 41032   | PASS             |
| 443         | 442          | 15           | 24           | 19.0      | 7799    | PASS             |