

Data Path : Z:\HPCHEM1\BNA F\DATA\BF122117\
 Data File : BF101642.D
 Acq On : 22 Dec 2017 8:28
 Operator : SJ/JU
 Sample : I7008-09
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 121817-TIFR-E-D-B

Quant Time: Dec 22 10:30:36 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF121417.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 20 14:25:04 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.80	152	140540	20.00	ng	0.00
21) Naphthalene-d8	8.08	136	531297	20.00	ng	0.00
38) Acenaphthene-d10	9.83	164	202348	20.00	ng	0.00
63) Phenanthrene-d10	11.32	188	305665	20.00	ng	0.00
75) Chrysene-d12	13.97	240	308061	20.00	ng	0.00
86) Perylene-d12	15.44	264	234814	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.42	112	420016	44.52	ng	0.00
7) Phenol-d6	6.44	99	332940	28.35	ng	0.00
23) Nitrobenzene-d5	7.37	82	835979	87.59	ng	0.00
41) 2,4,6-Tribromophenol	10.63	330	140106	71.86	ng	0.00
44) 2-Fluorobiphenyl	9.15	172	1313285	100.68	ng	0.00
78) Terphenyl-d14	12.90	244	944784	73.94	ng	0.00
Target Compounds						
10) Phenol	6.46	94	42391	3.167	ng	Qvalue # 77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF122117\
Data File : BF101642.D
Acq On : 22 Dec 2017 8:28
Operator : SJ/JU
Sample : I7008-09
Misc :
ALS Vial : 44 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
121817-TIFR-E-D-B

Quant Time: Dec 22 10:30:36 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF121417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 20 14:25:04 2017
Response via : Initial Calibration

