



Data Path : Z:\HPCHEM1\BNA F\DATA\BF122117\
Data File : BF101623.D
Acq On : 21 Dec 2017 23:28
Operator : SJ/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Dec 22 06:39:37 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF121417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 20 14:25:04 2017
Response via : Initial Calibration



TIC: BF101623.D

(76) Benzidine



12.663min (-0.000) 0.00ng

response 348446

Ion	Exp%	Act%
184.00	100	100
185.00	15.70	14.56
183.00	11.60	11.60
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA F\DATA\BF122117\
Data File : BF101623.D
Acq On : 21 Dec 2017 23:28
Operator : SJ/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Dec 22 06:39:37 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF121417.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 20 14:25:04 2017
Response via : Initial Calibration



TIC: BF101623.D

(69) Pentachlorophenol (C)

11.145min (+0.006) 19043.92ng

response 30496

Ion	Exp%	Act%
265.70	100	100
268.00	63.90	69.24
264.00	61.90	64.06
0.00	0.00	0.00