


Data Path : Z:\HPCHEM1\BNA F\Data\BF122117\
 Data File : BF101597.D
 Acq On : 21 Dec 2017 10:19
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF121417.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Dec 22 16:07:58 2017



AutoFind: Scans 1610, 1611, 1612; Background Corrected with Scan 1605

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	40.5	49909	PASS
68	69	0.00	2	1.8	743	PASS
69	198	0.00	100	34.2	42156	PASS
70	69	0.00	2	0.7	283	PASS
127	198	10	80	44.2	54533	PASS
197	198	0.00	2	1.1	1301	PASS
198	198	100	100	100.0	123285	PASS
199	198	5	9	6.5	8073	PASS
275	198	10	60	24.7	30424	PASS
365	198	1	100	3.3	4059	PASS
441	443	0.01	100	84.5	17580	PASS
442	198	50	100	91.0	112152	PASS
443	442	15	24	18.6	20807	PASS