

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF122120\
 Data File : BF122799.D
 Acq On : 21 Dec 2020 11:43
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 21 12:08:08 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 15 18:14:23 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
2	1,4-Dioxane	0.594	0.554	6.7	83	0.02
3	Pyridine	1.569	1.503	4.2	83	0.01
4	n-Nitrosodimethylamine	0.629	0.559	11.1	77	0.02
5 S	2-Fluorophenol	1.263	1.171	7.3	82	0.00
6	Aniline	1.972	1.851	6.1	82	0.00
7 S	Phenol-d6	1.675	1.540	8.1	82	0.00
8	2-Chlorophenol	1.392	1.303	6.4	83	0.00
9	Benzaldehyde	1.014	0.983	3.1	98	0.00
10 C	Phenol	1.736	1.584	8.8	82	0.00
11	bis(2-Chloroethyl)ether	1.326	1.257	5.2	84	0.00
12	1,3-Dichlorobenzene	1.648	1.530	7.2	83	0.00
13 C	1,4-Dichlorobenzene	1.661	1.531	7.8	82	0.00
14	1,2-Dichlorobenzene	1.604	1.390	13.3	81	0.00
15	Benzyl Alcohol	1.154	1.049	9.1	78	0.00
16	2,2'-oxybis(1-Chloropropane	1.891	1.664	12.0	77	0.00
17	2-Methylphenol	1.107	1.073	3.1	84	0.00
18	Hexachloroethane	0.592	0.555	6.2	83	0.00
19 P	n-Nitroso-di-n-propylamine	0.960	0.870	9.4	80	0.00
20	3+4-Methylphenols	1.398	1.244	11.0	81	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	91	0.00
22	Acetophenone	0.497	0.446	10.3	82	0.00
23 S	Nitrobenzene-d5	0.399	0.434	-8.8	97	0.00
24	Nitrobenzene	0.397	0.360	9.3	82	0.00
25	Isophorone	0.688	0.643	6.5	83	0.00
26 C	2-Nitrophenol	0.194	0.195	-0.5	87	0.00
27	2,4-Dimethylphenol	0.284	0.269	5.3	84	0.00
28	bis(2-Chloroethoxy)methane	0.471	0.437	7.2	83	0.00
29 C	2,4-Dichlorophenol	0.332	0.308	7.2	83	0.00
30	1,2,4-Trichlorobenzene	0.376	0.348	7.4	84	0.00
31	Naphthalene	1.104	1.023	7.3	84	0.00
32	Benzoic acid	0.226	0.206	8.8	76	0.01
33	4-Chloroaniline	0.461	0.436	5.4	84	0.00
34 C	Hexachlorobutadiene	0.231	0.212	8.2	82	0.00
35	Caprolactam	0.100	0.095	5.0	84	0.01
36 C	4-Chloro-3-methylphenol	0.327	0.309	5.5	84	0.00
37	2-Methylnaphthalene	0.730	0.661	9.5	82	0.00
38	1-Methylnaphthalene	0.691	0.620	10.3	82	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	89	0.00
40	1,2,4,5-Tetrachlorobenzene	0.662	0.613	7.4	81	0.00
41 P	Hexachlorocyclopentadiene	0.293	0.320	-9.2	90	0.00
42 S	2,4,6-Tribromophenol	0.262	0.247	5.7	82	0.00
43 C	2,4,6-Trichlorophenol	0.458	0.423	7.6	80	0.00
44	2,4,5-Trichlorophenol	0.473	0.444	6.1	82	0.00
45 S	2-Fluorobiphenyl	1.479	1.417	4.2	91	0.00
46	1,1'-Biphenyl	1.660	1.524	8.2	82	0.00
47	2-Chloronaphthalene	1.375	1.285	6.5	83	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF122120\
 Data File : BF122799.D
 Acq On : 21 Dec 2020 11:43
 Operator : JU/CG
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 21 12:08:08 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 15 18:14:23 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.401	0.379	5.5	80	0.00
49	Acenaphthylene	2.031	1.892	6.8	83	0.00
50	Dimethylphthalate	1.597	1.517	5.0	84	0.00
51	2,6-Dinitrotoluene	0.337	0.327	3.0	83	0.00
52 C	Acenaphthene	1.314	1.118	14.9	74	0.00
53	3-Nitroaniline	0.390	0.378	3.1	83	0.00
54 P	2,4-Dinitrophenol	0.165	0.152	7.9	76	0.00
55	Dibenzofuran	1.849	1.674	9.5	81	0.00
56 P	4-Nitrophenol	0.278	0.243	12.6	72	0.01
57	2,4-Dinitrotoluene	0.449	0.430	4.2	81	0.00
58	Fluorene	1.470	1.290	12.2	82	0.00
59	2,3,4,6-Tetrachlorophenol	0.415	0.378	8.9	78	0.00
60	Diethylphthalate	1.554	1.443	7.1	82	0.00
61	4-Chlorophenyl-phenylether	0.724	0.664	8.3	83	0.00
62	4-Nitroaniline	0.396	0.389	1.8	85	0.00
63	Azobenzene	1.428	1.298	9.1	80	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	90	0.00
65	4,6-Dinitro-2-methylphenol	0.122	0.119	2.5	81	0.00
66 c	n-Nitrosodiphenylamine	0.707	0.651	7.9	82	0.00
67	4-Bromophenyl-phenylether	0.271	0.252	7.0	82	0.00
68	Hexachlorobenzene	0.300	0.282	6.0	83	0.00
69	Atrazine	0.229	0.188	17.9	73	0.00
70 C	Pentachlorophenol	0.159	0.143	10.1	73	0.00
71	Phenanthrene	1.189	1.086	8.7	83	0.00
72	Anthracene	1.179	1.083	8.1	83	0.00
73	Carbazole	1.069	0.991	7.3	83	0.00
74	Di-n-butylphthalate	1.343	1.203	10.4	80	0.00
75 C	Fluoranthene	1.246	1.116	10.4	80	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	86	0.00
77	Benzydine	0.433	0.507	-17.1	89	0.00
78	Pyrene	1.546	1.472	4.8	81	0.00
79 S	Terphenyl-d14	1.143	1.187	-3.8	92	0.00
80	Butylbenzylphthalate	0.697	0.669	4.0	80	0.00
81	Benzo(a)anthracene	1.337	1.273	4.8	82	0.00
82	3,3'-Dichlorobenzidine	0.479	0.466	2.7	82	0.00
83	Chrysene	1.330	1.254	5.7	81	0.00
84	Bis(2-ethylhexyl)phthalate	0.954	0.867	9.1	79	0.00
85 c	Di-n-octyl phthalate	1.582	1.389	12.2	76	0.00
86 I	Perylene-d12	1.000	1.000	0.0	89	0.00
87	Indeno(1,2,3-cd)pyrene	1.313	1.237	5.8	85	0.01
88	Benzo(b)fluoranthene	1.403	1.268	9.6	76	0.00
89	Benzo(k)fluoranthene	1.283	1.273	0.8	89	0.00
90 C	Benzo(a)pyrene	1.228	1.169	4.8	83	0.01
91	Dibenzo(a,h)anthracene	1.074	1.026	4.5	86	0.01
92	Benzo(g,h,i)perylene	1.040	0.981	5.7	87	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF122120\
Data File : BF122799.D
Acq On : 21 Dec 2020 11:43
Operator : JU/CG
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Dec 21 12:08:08 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF120520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 15 18:14:23 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0