

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF122122\
 Data File : BF131765.D
 Acq On : 22 Dec 2022 01:38
 Operator : CG\JU
 Sample : N6151-06
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 FS-01-(0-2)

Quant Time: Dec 22 05:53:48 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 21 03:12:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.863	152	77228	20.000	ng	0.00
21) Naphthalene-d8	8.151	136	277801	20.000	ng	0.00
39) Acenaphthene-d10	9.916	164	164980	20.000	ng	0.00
64) Phenanthrene-d10	11.410	188	333374	20.000	ng	0.00
76) Chrysene-d12	14.057	240	243649	20.000	ng	0.00
86) Perylene-d12	15.551	264	198568	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.492	112	645901	140.833	ng	0.02
7) Phenol-d6	6.498	99	739589	119.810	ng	0.00
23) Nitrobenzene-d5	7.434	82	681229	99.551	ng	0.00
42) 2,4,6-Tribromophenol	10.710	330	310086	171.720	ng	0.00
45) 2-Fluorobiphenyl	9.233	172	1187593	100.837	ng	0.00
79) Terphenyl-d14	13.004	244	1592817	101.886	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF122122\
Data File : BF131765.D
Acq On : 22 Dec 2022 01:38
Operator : CG\JU
Sample : N6151-06
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
FS-01-(0-2)

Quant Time: Dec 22 05:53:48 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122022.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 21 03:12:37 2022
Response via : Initial Calibration

