

Data Path : Z:\HPCHEM1\BNA F\DATA\BF122215\
 Data File : BF084020.D
 Acq On : 23 Dec 2015 6:08
 Operator : UM/SJ
 Sample : G4916-12
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 20151218MGP222BRV64

Quant Time: Dec 23 06:45:33 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF122215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 22 18:54:09 2015
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.83	152	72235	20.00	ng	0.00
21) Naphthalene-d8	8.11	136	286873	20.00	ng	0.00
38) Acenaphthene-d10	9.87	164	86476	20.00	ng	0.00
63) Phenanthrene-d10	11.36	188	168263	20.00	ng	0.00
75) Chrysene-d12	13.99	240	159925	20.00	ng	-0.01
86) Perylene-d12	15.43	264	126265	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.45	112	132671	31.32	ng	0.01
7) Phenol-d6	6.48	99	173397	34.81	ng	0.00
23) Nitrobenzene-d5	7.39	82	406646	87.08	ng	0.00
41) 2,4,6-Tribromophenol	10.66	330	126772	152.41	ng	0.00
44) 2-Fluorobiphenyl	9.20	172	587211	91.39	ng	0.00
78) Terphenyl-d14	12.93	244	567289	84.59	ng	0.00
Target Compounds						Qvalue
9) Benzaldehyde	6.60	77	10051	3.02	ng	90
15) Benzyl Alcohol	6.98	79	8677	2.21	ng	# 48
17) 2-Methylphenol	7.09	107	19512	5.02	ng	96
18) Hexachloroethane	7.40	117	27730	14.29	ng	# 5
19) n-Nitroso-di-n-propylamine	7.39	70	54883	18.18	ng	# 77
20) 3+4-Methylphenols	7.25	107	29495	6.15	ng	# 48
25) Isophorone	7.39	82	345963	39.10	ng	# 66
27) 2,4-Dimethylphenol	7.78	122	44364	9.83	ng	94
31) Naphthalene	8.13	128	837127	57.12	ng	99
32) Benzoic acid	7.92	122	26216	16.48	ng	# 19
33) 4-Chloroaniline	8.13	127	115435	19.26	ng	# 33
37) 2-Methylnaphthalene	8.83	142	160997	16.96	ng	99
45) 1,1'-Biphenyl	9.29	154	37623	4.92	ng	98
48) Acenaphthylene	9.73	152	83110	9.54	ng	98
51) Acenaphthene	9.90	154	77730	15.00	ng	98
57) Fluorene	10.42	166	65269	10.62	ng	97
66) 4-Bromophenyl-phenylether	10.66	248	9458	4.90	ng	# 1
70) Phenanthrene	11.38	178	82914	8.72	ng	99
71) Anthracene	11.38	178	82914	8.59	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF122215\
Data File : BF084020.D
Acq On : 23 Dec 2015 6:08
Operator : UM/SJ
Sample : G4916-12
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
20151218MGP222BRV64

Quant Time: Dec 23 06:45:33 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF122215.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 22 18:54:09 2015
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 20.00 ng

RT: 6.83 min Scan# 415

Delta R.T. -0.00 min

Lab File: BF084020.D

Acq: 23 Dec 2015 6:08

Instrument :

BNA_F

ClientSampleId :

20151218MGP222BRV64


Tot Ion:152 Resp: 72235

Ion Ratio Lower Upper

152 100

150 155.3 123.0 184.4




115 58.6 51.4 77.2



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#5

2-Fluorophenol

Concen: 31.32 ng

RT: 5.45 min Scan# 294

Delta R.T. 0.01 min

Lab File: BF084020.D

Acq: 23 Dec 2015 6:08


Tgt Ion:112 Resp: 132671

Ion Ratio Lower Upper

112 100

64 45.4 36.6 55.0



63 24.9 19.4 29.0




Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF0

Ion 63.00 (62.70 to 63.70): BF0





#7

Phenol-d6

Concen: 34.81 ng

RT: 6.48 min Scan# 384

Delta R.T. -0.00 min

Lab File: BF084020.D

Acq: 23 Dec 2015 6:08

Instrument :

BNA_F

ClientSampleId :

20151218MGP222BRV64


Tot Ion: 99 Resp: 173397

Ion Ratio Lower Upper

99 100

42 11.4 12.8 19.2#




71 31.8 30.9 46.3



Abundance Ion 99.00 (98.70 to 99.70): BF0

Ion 42.00 (41.70 to 42.70): BF0

Ion 71.00 (70.70 to 71.70): BF0



#9

Benzaldehyde

Concen: 3.02 ng

RT: 6.60 min Scan# 395

Delta R.T. 0.23 min

Lab File: BF084020.D

Acq: 23 Dec 2015 6:08


Tgt Ion: 77 Resp: 10051

Ion Ratio Lower Upper

77 100

105 88.1 83.0 123.0



106 90.7 74.6 114.6




Abundance Ion 77.00 (76.70 to 77.70): BF0

Ion 105.00 (104.70 to 105.70): E

Ion 106.00 (105.70 to 106.70): E





#15

Benzyl Alcohol

Concen: 2.21 ng

RT: 6.98 min Scan# 428

Delta R.T. -0.00 min

Lab File: BF084020.D

Acq: 23 Dec 2015 6:08

Instrument :

BNA_F

ClientSampleId :

20151218MGP222BRV64





Tot Ion: 79 Resp: 8677

Ion Ratio Lower Upper

79 100

108 30.4 69.0 103.4#

77 93.7 50.0 75.0#



#17

2-Methylphenol

Concen: 5.02 ng

RT: 7.09 min Scan# 438

Delta R.T. -0.00 min

Lab File: BF084020.D

Acq: 23 Dec 2015 6:08

Tgt Ion: 107 Resp: 19512




Ion Ratio Lower Upper


107 100

108 115.5 89.8 134.6

77 48.6 35.7 53.5

79 48.7 35.7 53.5









#18
Hexachloroethane
Concen: 14.29 ng
RT: 7.40 min Scan# 465
Delta R.T. 0.06 min
Lab File: BF084020.D
Acq: 23 Dec 2015 6:08




Instrument :
BNA_F
ClientSampleId :
20151218MGP222BRV64


Tot Ion: 117 Resp: 27730
Ion Ratio Lower Upper
117 100
119 2.8 77.4 116.0#
201 0.0 68.6 103.0#



#19
n-Nitroso-di-n-propylamine
Concen: 18.18 ng
RT: 7.39 min Scan# 464
Delta R.T. 0.15 min
Lab File: BF084020.D
Acq: 23 Dec 2015 6:08

Tgt Ion: 70 Resp: 54883
Ion Ratio Lower Upper
70 100
42 40.4 33.3 49.9
101 0.0 9.3 13.9#
130 2.3 23.4 35.0#





#20

3+4-Methylphenols

Concen: 6.15 ng

RT: 7.25 min Scan# 452

Delta R.T. -0.00 min

Lab File: BF084020.D

Acq: 23 Dec 2015 6:08

Instrument :

BNA_F

ClientSampleId :

20151218MGP222BRV64

Tot Ion:107 Resp: 29495


Ion Ratio Lower Upper

107 100

108 87.9 64.1 104.1

77 35.5 128.9 168.9#

79 29.4 9.6 49.6






Abundance Ion 107.00 (106.70 to 107.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 77.00 (76.70 to 77.70): BF0

Ion 79.00 (78.70 to 79.70): BF0



#21

Naphthalene-d8

Concen: 20.00 ng

RT: 8.11 min Scan# 527

Delta R.T. -0.00 min

Lab File: BF084020.D

Acq: 23 Dec 2015 6:08

Tgt Ion:136 Resp: 286873


Ion Ratio Lower Upper

136 100

137 10.9 8.7 13.1

54 7.1 5.8 8.8

68 5.7 4.2 6.2






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF0

Ion 68.00 (67.70 to 68.70): BF0









#23
Nitrobenzene-d5
Concen: 87.08 ng
RT: 7.39 min Scan# 464
Delta R.T. -0.00 min
Lab File: BF084020.D
Acq: 23 Dec 2015 6:08

Instrument :
BNA_F
ClientSampleId :
20151218MGP222BRV64

Tot Ion	Ratio	Lower	Upper
82	100		
128	45.5	34.6	51.8
54	46.3	38.4	57.6




Abundance Ion 82.00 (81.70 to 82.70): BF0
Ion 128.00 (127.70 to 128.70): BF0
Ion 54.00 (53.70 to 54.70): BF0






#25
Isophorone
Concen: 39.10 ng
RT: 7.39 min Scan# 464
Delta R.T. -0.26 min
Lab File: BF084020.D
Acq: 23 Dec 2015 6:08

Tgt Ion	Ratio	Lower	Upper
82	100		
95	0.0	6.6	10.0#
138	0.0	13.6	20.4#



Abundance Ion 82.00 (81.70 to 82.70): BF0
Ion 95.00 (94.70 to 95.70): BF0
Ion 138.00 (137.70 to 138.70): BF0





#27

2,4-Dimethylphenol

Concen: 9.83 ng

RT: 7.78 min Scan# 498

Delta R.T. -0.00 min

Lab File: BF084020.D

Acq: 23 Dec 2015 6:08

Instrument :

BNA_F

ClientSampleId :

20151218MGP222BRV64


Tot Ion:122 Resp: 44364

Ion Ratio Lower Upper

122 100

107 115.6 99.1 148.7




121 56.6 47.9 71.9



Abundance Ion 122.00 (121.70 to 122.70): E

Ion 107.00 (106.70 to 107.70): E

Ion 121.00 (120.70 to 121.70): E



#31

Naphthalene

Concen: 57.12 ng

RT: 8.13 min Scan# 529

Delta R.T. -0.00 min

Lab File: BF084020.D

Acq: 23 Dec 2015 6:08


Tgt Ion:128 Resp: 837127

Ion Ratio Lower Upper

128 100

129 11.7 9.1 13.7



127 14.1 11.1 16.7




Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E





#32

Benzoic acid

Concen: 16.48 ng

RT: 7.92 min Scan# 510

Delta R.T. 0.01 min

Lab File: BF084020.D

Acq: 23 Dec 2015 6:08


Instrument :

BNA_F

ClientSampleId :

20151218MGP222BRV64




Tot Ion	Ratio	Lower	Upper
122	100		
105	6.2	100.3	140.3#
77	38.3	63.5	103.5#



Abundance Ion 122.00 (121.70 to 122.70): E

Ion 105.00 (104.70 to 105.70): E

Ion 77.00 (76.70 to 77.70): BF0



#33

4-Chloroaniline

Concen: 19.26 ng


RT: 8.13 min Scan# 529

Delta R.T. -0.06 min

Lab File: BF084020.D

Acq: 23 Dec 2015 6:08

Tgt Ion	Ratio	Lower	Upper
127	100		
129	82.9	26.5	39.7#
65	0.0	27.2	40.8#
92	0.0	17.7	26.5#






Abundance Ion 127.00 (126.70 to 127.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 65.00 (64.70 to 65.70): BF0

Ion 92.00 (91.70 to 92.70): BF0









#37
2-Methylnaphthalene
Concen: 16.96 ng
RT: 8.83 min Scan# 590
Delta R.T. -0.00 min
Lab File: BF084020.D
Acq: 23 Dec 2015 6:08




Instrument :
BNA_F
ClientSampleId :
20151218MGP222BRV64


Tot Ion:142 Resp: 160997
Ion Ratio Lower Upper
142 100
141 90.4 72.4 108.6
115 36.1 27.9 41.9



#38
Acenaphthene-d10
Concen: 20.00 ng
RT: 9.87 min Scan# 681
Delta R.T. -0.00 min
Lab File: BF084020.D
Acq: 23 Dec 2015 6:08

Tgt Ion:164 Resp: 86476
Ion Ratio Lower Upper
164 100
162 106.0 85.1 127.7
160 49.9 37.5 56.3









#41
 2,4,6-Tribromophenol
 Concen: 152.41 ng
 RT: 10.66 min Scan# 750
 Delta R.T. -0.00 min
 Lab File: BF084020.D
 Acq: 23 Dec 2015 6:08

Instrument :
 BNA_F
 ClientSampleId :
 20151218MGP222BRV64

Tot Ion	Ratio	Lower	Upper
330	100		
332	95.4	76.6	114.8
141	41.0	27.8	41.6




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#44
 2-Fluorobiphenyl
 Concen: 91.39 ng
 RT: 9.20 min Scan# 622
 Delta R.T. 0.00 min
 Lab File: BF084020.D
 Acq: 23 Dec 2015 6:08

Tgt Ion	Ratio	Lower	Upper
172	100		
171	36.4	28.9	43.3
170	23.8	18.6	27.8



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E





#45

1,1'-Biphenyl

Concen: 4.92 ng

RT: 9.29 min Scan# 630

Delta R.T. -0.01 min

Lab File: BF084020.D

Acq: 23 Dec 2015 6:08

Instrument :

BNA_F

ClientSampleId :

20151218MGP222BRV64


Tot Ion:154 Resp: 37623

Ion Ratio Lower Upper

154 100

153 40.6 21.6 61.6




76 11.5 0.0 32.7



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 76.00 (75.70 to 76.70): BF0



#48

Acenaphthylene

Concen: 9.54 ng

RT: 9.73 min Scan# 669

Delta R.T. -0.00 min

Lab File: BF084020.D

Acq: 23 Dec 2015 6:08


Tgt Ion:152 Resp: 83110

Ion Ratio Lower Upper

152 100

151 20.8 16.3 24.5



153 14.2 10.4 15.6




Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E





#51

Acenaphthene

Concen: 15.00 ng

RT: 9.90 min Scan# 684

Delta R.T. -0.00 min

Lab File: BF084020.D

Acq: 23 Dec 2015 6:08

Instrument :

BNA_F

ClientSampleId :

20151218MGP222BRV64


Tot Ion:154 Resp: 77730

Ion Ratio Lower Upper

154 100

153 116.8 91.5 137.3




152 54.1 43.0 64.6



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#57

Fluorene

Concen: 10.62 ng

RT: 10.42 min Scan# 729

Delta R.T. -0.00 min

Lab File: BF084020.D

Acq: 23 Dec 2015 6:08


Tgt Ion:166 Resp: 65269

Ion Ratio Lower Upper

166 100

165 102.2 79.1 118.7



167 13.3 10.4 15.6




Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E







#63
Phenanthrene-d10
Concen: 20.00 ng
RT: 11.36 min Scan# 811
Delta R.T. -0.00 min
Lab File: BF084020.D
Acq: 23 Dec 2015 6:08

Instrument :
BNA_F
ClientSampleId :
20151218MGP222BRV64



Tot Ion	Ratio	Lower	Upper
188	100		
94	6.0	9.0	13.4#
80	6.3	9.6	14.4#



Abundance Ion 188.00 (187.70 to 188.70): E
Ion 94.00 (93.70 to 94.70): BF0
Ion 80.00 (79.70 to 80.70): BF0




Time-->




#66
4-Bromophenyl-phenylether
Concen: 4.90 ng
RT: 10.66 min Scan# 750
Delta R.T. -0.24 min
Lab File: BF084020.D
Acq: 23 Dec 2015 6:08


Tgt Ion	Ratio	Lower	Upper
248	100		
250	191.8	78.4	117.6#
141	449.1	44.0	66.0#




Abundance Ion 248.00 (247.70 to 248.70): E
Ion 250.00 (249.70 to 250.70): E
Ion 141.00 (140.70 to 141.70): E



Time-->





#70

Phenanthrene

Concen: 8.72 ng

RT: 11.38 min Scan# 813

Delta R.T. -0.00 min

Lab File: BF084020.D

Acq: 23 Dec 2015 6:08

Instrument :

BNA_F

ClientSampleId :

20151218MGP222BRV64


Tot Ion:178 Resp: 82914

Ion Ratio Lower Upper

178 100

176 18.5 15.3 22.9




179 14.3 12.0 18.0



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#71

Anthracene

Concen: 8.59 ng

RT: 11.38 min Scan# 813

Delta R.T. -0.05 min

Lab File: BF084020.D

Acq: 23 Dec 2015 6:08


Tgt Ion:178 Resp: 82914

Ion Ratio Lower Upper

178 100

176 18.5 15.4 23.0



179 14.3 12.5 18.7




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E





#75

Chrysene-d12

Concen: 20.00 ng

RT: 13.99 min Scan# 1041

Delta R.T. -0.01 min

Lab File: BF084020.D

Acq: 23 Dec 2015 6:08

Instrument :

BNA_F

ClientSampleId :

20151218MGP222BRV64


Tot Ion:240 Resp: 159925

Ion Ratio Lower Upper

240 100

120 14.4 9.5 14.3#




236 26.0 20.6 31.0



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#78

Terphenyl-d14

Concen: 84.59 ng

RT: 12.93 min Scan# 949

Delta R.T. -0.00 min

Lab File: BF084020.D

Acq: 23 Dec 2015 6:08


Tgt Ion:244 Resp: 567289

Ion Ratio Lower Upper

244 100

212 7.9 5.7 8.5



122 14.6 7.4 11.0#




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E





#86
Pervlene-d12
Concen: 20.00 ng
RT: 15.43 min Scan# 1167
Delta R.T. -0.00 min
Lab File: BF084020.D
Acq: 23 Dec 2015 6:08

Instrument :
BNA_F
ClientSampleId :
20151218MGP222BRV64

Tot Ion	Ratio	Lower	Upper
264	100		
260	22.8	19.4	29.2
265	21.2	17.8	26.6

