

Data Path : Z:\HPCHEM1\BNA F\DATA\BF122215\
 Data File : BF084020.D
 Acq On : 23 Dec 2015 6:08
 Operator : UM/SJ
 Sample : G4916-12
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 20151218MGP222BRV64

Quant Time: Dec 23 06:52:12 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF122215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 22 18:54:09 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.83	152	72235	20.00	ng	0.00
21) Naphthalene-d8	8.11	136	286873	20.00	ng	0.00
38) Acenaphthene-d10	9.87	164	86476	20.00	ng	0.00
63) Phenanthrene-d10	11.36	188	168263	20.00	ng	0.00
75) Chrysene-d12	13.99	240	159925	20.00	ng	-0.01
86) Perylene-d12	15.43	264	126265	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.45	112	132671	31.32	ng	0.01
7) Phenol-d6	6.48	99	173397	34.81	ng	0.00
23) Nitrobenzene-d5	7.39	82	406646	87.08	ng	0.00
41) 2,4,6-Tribromophenol	10.66	330	126772	152.41	ng	0.00
44) 2-Fluorobiphenyl	9.20	172	587211	91.39	ng	0.00
78) Terphenyl-d14	12.93	244	567289	84.59	ng	0.00
Target Compounds						Qvalue
17) 2-Methylphenol	7.09	107	19512	5.02	ng	96
27) 2,4-Dimethylphenol	7.78	122	44364	9.83	ng	94
31) Naphthalene	8.13	128	837127	57.12	ng	99
37) 2-Methylnaphthalene	8.83	142	160997	16.96	ng	99
45) 1,1'-Biphenyl	9.29	154	37623	4.92	ng	98
48) Acenaphthylene	9.73	152	83110	9.54	ng	98
51) Acenaphthene	9.90	154	77730	15.00	ng	98
57) Fluorene	10.42	166	65269	10.62	ng	97
70) Phenanthrene	11.38	178	82914	8.72	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF122215\
Data File : BF084020.D
Acq On : 23 Dec 2015 6:08
Operator : UM/SJ
Sample : G4916-12
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
20151218MGP222BRV64

Quant Time: Dec 23 06:52:12 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF122215.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 22 18:54:09 2015
Response via : Initial Calibration

