


Data Path : Z:\HPCHEM1\BNA F\DATA\BF122215\
 Data File : BF083999.D
 Acq On : 22 Dec 2015 19:50
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF122215.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Dec 22 18:54:09 2015



AutoFind: Scans 833, 834, 835; Background Corrected with Scan 830

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	31.3	8517	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	34.1	9264	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	52.2	14201	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	27184	PASS
199	198	5	9	6.9	1871	PASS
275	198	10	60	28.9	7845	PASS
365	198	1	100	2.9	797	PASS
441	442	0.01	24	16.7	3830	PASS
442	198	50	100	84.6	22990	PASS
443	442	15	24	19.4	4464	PASS