

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF122216\
 Data File : BF091903.D
 Acq On : 23 Dec 2016 4:12
 Operator : UM/SJ
 Sample : H6182-05RE
 Misc :
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-51(11.5-13.5)-121916

Quant Time: Dec 23 06:40:37 2016
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF122116.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 21 16:17:51 2016
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.75	152	202399	20.00	ng	0.00
21) Naphthalene-d8	8.04	136	759775	20.00	ng	0.00
38) Acenaphthene-d10	9.79	164	316199	20.00	ng	0.00
63) Phenanthrene-d10	11.28	188	489397	20.00	ng	0.01
75) Chrysene-d12	13.91	240	388150	20.00	ng	0.00
86) Perylene-d12	15.30	264	316367	20.00	ng	0.00

System Monitoring Compounds						
5) 2-Fluorophenol	5.40	112	1714718	141.46	ng	0.06
7) Phenol-d6	6.45	99	2371839	160.27	ng	0.05
23) Nitrobenzene-d5	7.32	82	1435480	105.06	ng	0.00
41) 2,4,6-Tribromophenol	10.59	330	346604	132.45	ng	0.01
44) 2-Fluorobiphenyl	9.12	172	1885972	128.87	ng	0.00
78) Terphenyl-d14	12.87	244	1448424	90.50	ng	0.01

Target Compounds						Qvalue
10) Phenol	6.46	94	76008	4.51	ng	# 68
15) Benzyl Alcohol	6.91	79	37121	3.23	ng	# 39
18) Hexachloroethane	7.32	117	162203	29.20	ng	# 6
19) n-Nitroso-di-n-propylamine	7.17	70	45542	4.63	ng	# 67
24) Nitrobenzene	7.30	77	68465	4.70	ng	# 45
25) Isophorone	7.56	82	146141	5.80	ng	# 1
27) 2,4-Dimethylphenol	7.69	122	24707	2.08	ng	# 1
28) bis(2-Chloroethoxy)methane	7.84	93	31987	2.09	ng	# 1
35) Caprolactam	8.48	113	100770	30.42	ng	# 1
36) 4-Chloro-3-methylphenol	8.57	107	37332	3.22	ng	# 24
47) 2-Nitroaniline	9.50	65	25152	4.12	ng	# 95
49) Dimethylphthalate	9.52	163	261842	12.95	ng	# 94
50) 2,6-Dinitrotoluene	9.60	165	26321	5.95	ng	# 53
52) 3-Nitroaniline	9.74	138	35619	6.50	ng	# 20
55) 4-Nitrophenol	9.90	139	9506	2.56	ng	# 1
56) 2,4-Dinitrotoluene	10.03	165	14528	2.61	ng	# 60
61) 4-Nitroaniline	10.25	138	24015	4.23	ng	# 90
65) n-Nitrosodiphenylamine	10.45	169	98634	6.79	ng	# 43
73) Di-n-butylphthalate	11.81	149	23609	Below	Cal	# 36
74) Fluoranthene	12.71	202	32065	Below	Cal	# 80
76) Benzidine	12.64	184	1207	Below	Cal	# 1
83) Bis(2-ethylhexyl)phthalate	13.88	149	498223	32.66	ng	# 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

```
Data Path : Z:\HPCHEM1\BNA_F\DATA\BF122216\  
Data File : BF091903.D  
Acq On    : 23 Dec 2016    4:12  
Operator  : UM/SJ  
Sample    : H6182-05RE  
Misc      :  
ALS Vial  : 54    Sample Multiplier: 1
```



#1

1,4-Dichlorobenzene-d4

Concen: 20.00 ng

RT: 6.75 min Scan# 408

Delta R.T. -0.00 min

Lab File: BF091903.D

Acq: 23 Dec 2016 4:12

Instrument :

BNA_F

ClientSampleId :

SB-51(11.5-13.5)-121916


Tgt Ion:152 Resp: 202399

Ion Ratio Lower Upper

152 100

150 156.1 124.9 187.3




115 63.2 46.0 69.0



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#5

2-Fluorophenol

Concen: 141.46 ng

RT: 5.40 min Scan# 290

Delta R.T. 0.06 min

Lab File: BF091903.D

Acq: 23 Dec 2016 4:12


Tgt Ion:112 Resp: 1714718

Ion Ratio Lower Upper

112 100

64 64.1 46.1 69.1



63 34.3 23.8 35.6






Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF0

Ion 63.00 (62.70 to 63.70): BF0





#7

Phenol-d6

Concen: 160.27 ng

RT: 6.45 min Scan# 382

Delta R.T. 0.05 min

Lab File: BF091903.D

Acq: 23 Dec 2016 4:12

Instrument :

BNA_F

ClientSampleId :

SB-51(11.5-13.5)-121916





Tgt Ion: 99 Resp: 2371839

Ion Ratio Lower Upper

99 100

42 18.4 15.6 23.4

71 35.5 27.5 41.3



#10

Phenol

Concen: 4.51 ng

RT: 6.46 min Scan# 383

Delta R.T. 0.05 min

Lab File: BF091903.D

Acq: 23 Dec 2016 4:12


Tgt Ion: 94 Resp: 76008


Ion Ratio Lower Upper

94 100

65 25.7 21.4 61.4

66 35.5 44.0 84.0#





#15

Benzyl Alcohol

Concen: 3.23 ng

RT: 6.91 min Scan# 422

Delta R.T. 0.01 min

Lab File: BF091903.D

Acq: 23 Dec 2016 4:12

Instrument :

BNA_F

ClientSampleId :

SB-51(11.5-13.5)-121916





Tgt Ion: 79 Resp: 37121

Ion Ratio Lower Upper

79 100

108 26.8 61.4 92.0#

77 113.7 50.9 76.3#



#18

Hexachloroethane

Concen: 29.20 ng

RT: 7.32 min Scan# 458

Delta R.T. 0.06 min

Lab File: BF091903.D

Acq: 23 Dec 2016 4:12




Tgt Ion: 117 Resp: 162203




Ion Ratio Lower Upper

117 100

119 10.8 78.1 117.1#

201 0.0 78.6 117.8#





#19

n-Nitroso-di-n-propylamine

Concen: 4.63 ng

RT: 7.17 min Scan# 445

Delta R.T. -0.01 min

Lab File: BF091903.D

Acq: 23 Dec 2016 4:12

Instrument :

BNA_F

ClientSampleId :

SB-51(11.5-13.5)-121916

Tgt Ion: 70 Resp: 45542





Ion Ratio Lower Upper

70 100

42 37.5 49.4 74.0#

101 0.0 7.4 11.2#

130 0.0 14.2 21.4#



#21

Naphthalene-d8

Concen: 20.00 ng

RT: 8.04 min Scan# 521

Delta R.T. -0.00 min

Lab File: BF091903.D

Acq: 23 Dec 2016 4:12

Tgt Ion: 136 Resp: 759775


Ion Ratio Lower Upper


136 100

137 12.4 8.4 12.6

54 9.4 4.5 6.7#

68 8.9 3.6 5.4#





#23

Nitrobenzene-d5

Concen: 105.06 ng

RT: 7.32 min Scan# 458

Delta R.T. -0.00 min

Lab File: BF091903.D

Acq: 23 Dec 2016 4:12

Instrument :

BNA_F

ClientSampleId :

SB-51(11.5-13.5)-121916


Tgt Ion: 82 Resp: 1435480

Ion Ratio Lower Upper

82 100

128 41.3 34.6 52.0




54 51.3 39.5 59.3



Abundance Ion 82.00 (81.70 to 82.70): BF0

Ion 128.00 (127.70 to 128.70): BF0

Ion 54.00 (53.70 to 54.70): BF0



#24

Nitrobenzene

Concen: 4.70 ng

RT: 7.30 min Scan# 456

Delta R.T. -0.05 min

Lab File: BF091903.D

Acq: 23 Dec 2016 4:12


Tgt Ion: 77 Resp: 68465

Ion Ratio Lower Upper

77 100

123 16.8 33.0 49.4#



65 55.4 11.2 16.8#




Abundance Ion 77.00 (76.70 to 77.70): BF0

Ion 123.00 (122.70 to 123.70): BF0

Ion 65.00 (64.70 to 65.70): BF0





#25

Isophorone

Concen: 5.80 ng

RT: 7.56 min Scan# 479

Delta R.T. -0.02 min

Lab File: BF091903.D

Acq: 23 Dec 2016 4:12

Instrument :

BNA_F

ClientSampleId :

SB-51(11.5-13.5)-121916





Tgt Ion: 82 Resp: 146141

Ion Ratio Lower Upper

82 100

95 169.0 5.6 8.4#

138 8.9 14.6 22.0#



#27

2,4-Dimethylphenol

Concen: 2.08 ng

RT: 7.69 min Scan# 490

Delta R.T. -0.02 min

Lab File: BF091903.D

Acq: 23 Dec 2016 4:12




Tgt Ion: 122 Resp: 24707


Ion Ratio Lower Upper

122 100

107 329.0 94.1 141.1#

121 92.7 46.0 69.0#









#28
bis(2-Chloroethoxy)methane
Concen: 2.09 ng
RT: 7.84 min Scan# 503
Delta R.T. 0.03 min
Lab File: BF091903.D
Acq: 23 Dec 2016 4:12




Instrument :
BNA_F
ClientSampleId :
SB-51(11.5-13.5)-121916


Tgt Ion: 93 Resp: 31987
Ion Ratio Lower Upper
93 100
95 488.0 26.5 39.7#
123 231.6 8.6 13.0#



#35
Caprolactam
Concen: 30.42 ng
RT: 8.48 min Scan# 559
Delta R.T. -0.02 min
Lab File: BF091903.D
Acq: 23 Dec 2016 4:12

Tgt Ion: 113 Resp: 100770
Ion Ratio Lower Upper
113 100
55 332.5 119.2 159.2#
56 239.9 83.8 123.8#





#36

4-Chloro-3-methylphenol

Concen: 3.22 ng

RT: 8.57 min Scan# 567

Delta R.T. -0.05 min

Lab File: BF091903.D

Acq: 23 Dec 2016 4:12

Instrument :

BNA_F

ClientSampleId :

SB-51(11.5-13.5)-121916


Tgt Ion:107 Resp: 37332

Ion Ratio Lower Upper

107 100

144 1.7 19.3 28.9#




142 1.5 59.0 88.6#



Abundance Ion 107.00 (106.70 to 107.70): E

Ion 144.00 (143.70 to 144.70): E

Ion 142.00 (141.70 to 142.70): E



#38

Acenaphthene-d10

Concen: 20.00 ng

RT: 9.79 min Scan# 674

Delta R.T. -0.00 min

Lab File: BF091903.D

Acq: 23 Dec 2016 4:12


Tgt Ion:164 Resp: 316199

Ion Ratio Lower Upper

164 100

162 101.3 80.2 120.2



160 46.0 36.9 55.3




Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E









#41
 2,4,6-Tribromophenol
 Concen: 132.45 ng
 RT: 10.59 min Scan# 744
 Delta R.T. 0.01 min
 Lab File: BF091903.D
 Acq: 23 Dec 2016 4:12

Instrument :
 BNA_F
 ClientSampleId :
 SB-51(11.5-13.5)-121916

Tgt Ion:330 Resp: 346604
 Ion Ratio Lower Upper
 330 100
 332 96.5 77.4 116.2
 141 50.4 34.1 51.1




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E








#44
 2-Fluorobiphenyl
 Concen: 128.87 ng
 RT: 9.12 min Scan# 615
 Delta R.T. 0.00 min
 Lab File: BF091903.D
 Acq: 23 Dec 2016 4:12

Tgt Ion:172 Resp: 1885972
 Ion Ratio Lower Upper
 172 100
 171 35.3 29.4 44.0
 170 24.0 20.2 30.4



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E









#47
2-Nitroaniline
Concen: 4.12 ng
RT: 9.50 min Scan# 649
Delta R.T. 0.16 min
Lab File: BF091903.D
Acq: 23 Dec 2016 4:12


Tgt Ion: 65 Resp: 25152
Ion Ratio Lower Upper
65 100
92 61.6 53.9 80.9
138 93.0 76.8 115.2


Instrument :
BNA_F
ClientSampleId :
SB-51(11.5-13.5)-121916



#49
Dimethylphthalate
Concen: 12.95 ng
RT: 9.52 min Scan# 650
Delta R.T. -0.01 min
Lab File: BF091903.D
Acq: 23 Dec 2016 4:12

Tgt Ion: 163 Resp: 261842
Ion Ratio Lower Upper
163 100
194 8.4 3.0 4.4#
164 11.3 8.0 12.0









#50
 2,6-Dinitrotoluene
 Concen: 5.95 ng
 RT: 9.60 min Scan# 657
 Delta R.T. 0.01 min
 Lab File: BF091903.D
 Acq: 23 Dec 2016 4:12




Instrument :
 BNA_F
 ClientSampleId :
 SB-51(11.5-13.5)-121916


Tgt Ion:165 Resp: 26321
 Ion Ratio Lower Upper
 165 100
 63 20.0 36.2 54.4#
 89 12.9 40.2 60.4#



#52
 3-Nitroaniline
 Concen: 6.50 ng
 RT: 9.74 min Scan# 670
 Delta R.T. -0.01 min
 Lab File: BF091903.D
 Acq: 23 Dec 2016 4:12

Tgt Ion:138 Resp: 35619
 Ion Ratio Lower Upper
 138 100
 108 61.0 8.5 12.7#
 92 37.4 97.8 146.8#





#55

4-Nitrophenol

Concen: 2.56 ng

RT: 9.90 min Scan# 684

Delta R.T. -0.02 min

Lab File: BF091903.D

Acq: 23 Dec 2016 4:12

Instrument :

BNA_F

ClientSampleId :

SB-51(11.5-13.5)-121916


Tgt Ion:139 Resp: 9506

Ion Ratio Lower Upper

139 100

109 637.9 52.2 92.2#




65 133.8 85.8 125.8#



Abundance Ion 139.00 (138.70 to 139.70): E

Ion 109.00 (108.70 to 109.70): E

Ion 65.00 (64.70 to 65.70): BF0



#56

2,4-Dinitrotoluene

Concen: 2.61 ng

RT: 10.03 min Scan# 695

Delta R.T. 0.05 min

Lab File: BF091903.D

Acq: 23 Dec 2016 4:12

Tgt Ion:165 Resp: 14528


Ion Ratio Lower Upper

165 100

63 42.6 40.2 60.4

89 26.4 62.5 93.7#

182 9.4 1.8 2.8#






Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BF0

Ion 89.00 (88.70 to 89.70): BF0

Ion 182.00 (181.70 to 182.70): E





#61

4-Nitroaniline

Concen: 4.23 ng

RT: 10.25 min Scan# 714

Delta R.T. -0.13 min

Lab File: BF091903.D

Acq: 23 Dec 2016 4:12

Instrument :

BNA_F

ClientSampleId :

SB-51(11.5-13.5)-121916


Tgt Ion:138 Resp: 24015

Ion Ratio Lower Upper

138 100

92 42.9 31.7 71.7




108 65.7 52.6 92.6



Abundance Ion 138.00 (137.70 to 138.70): E

Ion 92.00 (91.70 to 92.70): BF0

Ion 108.00 (107.70 to 108.70): E



#63

Phenanthrene-d10

Concen: 20.00 ng

RT: 11.28 min Scan# 804

Delta R.T. 0.01 min

Lab File: BF091903.D

Acq: 23 Dec 2016 4:12


Tgt Ion:188 Resp: 489397

Ion Ratio Lower Upper

188 100

94 10.7 10.0 15.0



80 11.4 11.2 16.8




Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BF0

Ion 80.00 (79.70 to 80.70): BF0





#65

n-Nitrosodiphenylamine

Concen: 6.79 ng

RT: 10.45 min Scan# 732

Delta R.T. -0.00 min

Lab File: BF091903.D

Acq: 23 Dec 2016 4:12

Instrument :

BNA_F

ClientSampleId :

SB-51(11.5-13.5)-121916


Tgt Ion:169 Resp: 98634

Ion Ratio Lower Upper

169 100

168 14.2 54.2 81.2#

167 13.5 30.2 45.4#



Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E

100000

80000

60000

40000

20000



0

10.40

10.45

10.50

Time-->



#73

Di-n-butylphthalate

Concen: Below Cal

RT: 11.81 min Scan# 851

Delta R.T. -0.02 min

Lab File: BF091903.D

Acq: 23 Dec 2016 4:12


Tgt Ion:149 Resp: 23609

Ion Ratio Lower Upper

149 100

150 38.6 8.1 12.1#

104 20.0 4.6 7.0#



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E

25000

20000

15000

10000

5000


0


11.80

11.81

11.90

Time-->





#74

Fluoranthene

Concen: Below Cal

RT: 12.71 min Scan# 929

Delta R.T. 0.23 min

Lab File: BF091903.D

Acq: 23 Dec 2016 4:12





Instrument :

BNA_F

ClientSampleId :

SB-51(11.5-13.5)-121916

Tgt Ion:	202	Resp:	32065
Ion Ratio		Lower	Upper
202	100		
101	19.9	0.0	34.6
203	30.4	0.0	38.7



#75

Chrysene-d12

Concen: 20.00 ng




RT: 13.91 min Scan# 1034


Delta R.T. -0.00 min

Lab File: BF091903.D

Acq: 23 Dec 2016 4:12

Tgt Ion:	240	Resp:	388150
Ion Ratio		Lower	Upper
240	100		
120	8.3	12.9	19.3#
236	25.7	20.9	31.3





#78

Terphenyl-d14

Concen: 90.50 ng

RT: 12.87 min Scan# 943

Delta R.T. 0.01 min

Lab File: BF091903.D

Acq: 23 Dec 2016 4:12

Instrument :

BNA_F

ClientSampleId :

SB-51(11.5-13.5)-121916


Tgt Ion:244 Resp: 1448424

Ion Ratio Lower Upper

244 100

212 6.6 6.2 9.2




122 6.0 11.4 17.2#



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#83

Bis(2-ethylhexyl)phthalate

Concen: 32.66 ng

RT: 13.88 min Scan# 1032

Delta R.T. -0.01 min

Lab File: BF091903.D

Acq: 23 Dec 2016 4:12


Tgt Ion:149 Resp: 498223

Ion Ratio Lower Upper

149 100

167 28.9 20.2 30.4



279 2.2 1.8 2.8




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E






#86
Perylene-d12
Concen: 20.00 ng
RT: 15.30 min Scan# 1156
Delta R.T. -0.00 min
Lab File: BF091903.D
Acq: 23 Dec 2016 4:12

Instrument :
BNA_F
ClientSampleId :
SB-51(11.5-13.5)-121916

Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.0	19.2	28.8
265	21.7	17.4	26.0



Abundance Ion 264.00 (263.70 to 264.70): E
Ion 260.00 (259.70 to 260.70): E
Ion 265.00 (264.70 to 265.70): E

