


Data Path : Z:\HPCHEM1\BNA F\DATA\BF122216\
 Data File : BF091876.D
 Acq On : 22 Dec 2016 10:44
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF122116.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Dec 21 16:17:51 2016



AutoFind: Scans 825, 826, 827; Background Corrected with Scan 822

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.8	23058	PASS
68	69	0.00	2	1.4	310	PASS
69	198	0.00	100	37.6	22335	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	49.6	29449	PASS
197	198	0.00	2	0.3	164	PASS
198	198	100	100	100.0	59406	PASS
199	198	5	9	6.5	3853	PASS
275	198	10	60	22.6	13420	PASS
365	198	1	100	2.4	1414	PASS
441	442	0.01	24	15.8	7482	PASS
442	198	50	100	79.5	47231	PASS
443	442	15	24	18.9	8933	PASS