

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF122222\
 Data File : BF131777.D
 Acq On : 22 Dec 2022 15:10
 Operator : CG\JU
 Sample : N6102-03
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PMW-2

Quant Time: Dec 23 01:53:32 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 23 01:24:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.851	152	68858	20.000	ng	0.00
21) Naphthalene-d8	8.145	136	247020	20.000	ng	0.00
39) Acenaphthene-d10	9.904	164	139677	20.000	ng	0.00
64) Phenanthrene-d10	11.392	188	265768	20.000	ng	0.00
76) Chrysene-d12	14.045	240	179584	20.000	ng	0.00
86) Perylene-d12	15.533	264	161463	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.463	112	231217	56.543	ng	0.00
7) Phenol-d6	6.475	99	205003	37.246	ng	-0.01
23) Nitrobenzene-d5	7.422	82	581526	95.570	ng	0.00
42) 2,4,6-Tribromophenol	10.698	330	205283	134.276	ng	0.00
45) 2-Fluorobiphenyl	9.228	172	979451	98.229	ng	0.00
79) Terphenyl-d14	12.986	244	674927	58.574	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF122222\
Data File : BF131777.D
Acq On : 22 Dec 2022 15:10
Operator : CG\JU
Sample : N6102-03
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PMW-2

Quant Time: Dec 23 01:53:32 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122022.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 23 01:24:08 2022
Response via : Initial Calibration

