

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF122222\
 Data File : BF131780.D
 Acq On : 22 Dec 2022 16:52
 Operator : CG\JU
 Sample : PB149800TB
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB149800TB

Quant Time: Dec 23 03:31:21 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 23 03:03:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.851	152	78042	20.000	ng	0.00
21) Naphthalene-d8	8.139	136	291293	20.000	ng	0.00
39) Acenaphthene-d10	9.904	164	184181	20.000	ng	0.00
64) Phenanthrene-d10	11.398	188	381237	20.000	ng	0.00
76) Chrysene-d12	14.045	240	292092	20.000	ng	0.00
86) Perylene-d12	15.533	264	188101	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.475	112	721514	155.679	ng	0.02
7) Phenol-d6	6.486	99	929950	149.076	ng	0.00
23) Nitrobenzene-d5	7.422	82	649448	90.511	ng	0.00
42) 2,4,6-Tribromophenol	10.698	330	305598	151.591	ng	0.00
45) 2-Fluorobiphenyl	9.227	172	1176756	89.500	ng	0.00
79) Terphenyl-d14	12.992	244	1683433	89.824	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF122222\
Data File : BF131780.D
Acq On : 22 Dec 2022 16:52
Operator : CG\JU
Sample : PB149800TB
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB149800TB

Quant Time: Dec 23 03:31:21 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122022.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 23 03:03:57 2022
Response via : Initial Calibration

