

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF122222\
 Data File : BF131779.D
 Acq On : 22 Dec 2022 16:18
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 23 03:30:25 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 23 03:03:57 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
2	1,4-Dioxane	0.555	0.588	-5.9	94	0.00
3	Pyridine	1.583	1.640	-3.6	92	0.00
4	n-Nitrosodimethylamine	0.716	0.741	-3.5	91	0.00
5 S	2-Fluorophenol	1.188	1.213	-2.1	92	0.00
6	Aniline	1.974	1.981	-0.4	90	0.00
7 S	Phenol-d6	1.599	1.606	-0.4	91	0.00
8	2-Chlorophenol	1.276	1.291	-1.2	92	0.00
9	Benzaldehyde	1.073	0.978	8.9	97	0.00
10 C	Phenol	1.909	1.916	-0.4	91	0.00
11	bis(2-Chloroethyl)ether	1.366	1.392	-1.9	91	0.00
12	1,3-Dichlorobenzene	1.442	1.449	-0.5	91	0.00
13 C	1,4-Dichlorobenzene	1.474	1.489	-1.0	92	0.00
14	1,2-Dichlorobenzene	1.380	1.362	1.3	91	0.00
15	Benzyl Alcohol	1.428	1.422	0.4	90	0.00
16	2,2'-oxybis(1-Chloropropane	1.742	1.818	-4.4	95	0.00
17	2-Methylphenol	1.204	1.193	0.9	90	0.00
18	Hexachloroethane	0.584	0.598	-2.4	93	0.00
19 P	n-Nitroso-di-n-propylamine	1.153	1.168	-1.3	92	0.00
20	3+4-Methylphenols	1.550	1.546	0.3	90	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	90	0.00
22	Acetophenone	0.584	0.583	0.2	91	0.00
23 S	Nitrobenzene-d5	0.493	0.498	-1.0	90	0.00
24	Nitrobenzene	0.501	0.510	-1.8	91	0.00
25	Isophorone	0.837	0.848	-1.3	92	0.00
26 C	2-Nitrophenol	0.196	0.195	0.5	89	0.00
27	2,4-Dimethylphenol	0.279	0.279	0.0	91	0.00
28	bis(2-Chloroethoxy)methane	0.444	0.455	-2.5	92	0.00
29 C	2,4-Dichlorophenol	0.337	0.345	-2.4	93	0.00
30	1,2,4-Trichlorobenzene	0.396	0.395	0.3	90	0.00
31	Naphthalene	1.092	1.089	0.3	91	0.00
32	Benzoic acid	0.223	0.220	1.3	83	-0.01
33	4-Chloroaniline	0.426	0.415	2.6	88	0.00
34 C	Hexachlorobutadiene	0.266	0.272	-2.3	93	0.00
35	Caprolactam	0.101	0.098	3.0	88	0.00
36 C	4-Chloro-3-methylphenol	0.384	0.388	-1.0	91	0.00
37	2-Methylnaphthalene	0.745	0.742	0.4	91	0.00
38	1-Methylnaphthalene	0.722	0.726	-0.6	92	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	94	0.00
40	1,2,4,5-Tetrachlorobenzene	0.684	0.704	-2.9	92	0.00
41 P	Hexachlorocyclopentadiene	0.310	0.318	-2.6	87	0.00
42 S	2,4,6-Tribromophenol	0.219	0.225	-2.7	94	0.00
43 C	2,4,6-Trichlorophenol	0.447	0.461	-3.1	92	0.00
44	2,4,5-Trichlorophenol	0.484	0.503	-3.9	92	0.00
45 S	2-Fluorobiphenyl	1.428	1.459	-2.2	93	0.00
46	1,1'-Biphenyl	1.504	1.545	-2.7	92	0.00
47	2-Chloronaphthalene	1.232	1.262	-2.4	91	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF122222\
 Data File : BF131779.D
 Acq On : 22 Dec 2022 16:18
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 23 03:30:25 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122022.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 23 03:03:57 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.444	0.461	-3.8	93	0.00
49	Acenaphthylene	1.848	1.876	-1.5	92	0.00
50	Dimethylphthalate	1.505	1.558	-3.5	95	0.00
51	2,6-Dinitrotoluene	0.324	0.336	-3.7	93	0.00
52 C	Acenaphthene	1.132	1.143	-1.0	91	0.00
53	3-Nitroaniline	0.331	0.334	-0.9	91	0.00
54 P	2,4-Dinitrophenol	0.194	0.188	3.1	86	0.00
55	Dibenzofuran	1.736	1.771	-2.0	93	0.00
56 P	4-Nitrophenol	0.237	0.235	0.8	88	0.00
57	2,4-Dinitrotoluene	0.435	0.446	-2.5	95	0.00
58	Fluorene	1.403	1.422	-1.4	93	0.00
59	2,3,4,6-Tetrachlorophenol	0.404	0.414	-2.5	92	0.00
60	Diethylphthalate	1.450	1.494	-3.0	96	0.00
61	4-Chlorophenyl-phenylether	0.746	0.765	-2.5	95	0.00
62	4-Nitroaniline	0.333	0.326	2.1	89	0.00
63	Azobenzene	1.592	1.659	-4.2	96	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	94	-0.01
65	4,6-Dinitro-2-methylphenol	0.137	0.142	-3.6	92	-0.01
66 c	n-Nitrosodiphenylamine	0.620	0.624	-0.6	94	0.00
67	4-Bromophenyl-phenylether	0.236	0.242	-2.5	94	0.00
68	Hexachlorobenzene	0.260	0.263	-1.2	94	-0.01
69	Atrazine	0.207	0.208	-0.5	93	0.00
70 C	Pentachlorophenol	0.138	0.140	-1.4	91	0.00
71	Phenanthrene	1.091	1.078	1.2	94	0.00
72	Anthracene	1.068	1.053	1.4	94	0.00
73	Carbazole	0.924	0.896	3.0	93	0.00
74	Di-n-butylphthalate	1.150	1.181	-2.7	101	-0.01
75 C	Fluoranthene	1.215	1.170	3.7	94	-0.02
76 I	Chrysene-d12	1.000	1.000	0.0	95	-0.02
77	Benzydine	0.367	0.400	-9.0	101	-0.01
78	Pyrene	1.818	2.064	-13.5	94	-0.01
79 S	Terphenyl-d14	1.283	1.473	-14.8	97	-0.01
80	Butylbenzylphthalate	0.641	0.730	-13.9	100	-0.01
81	Benzo(a)anthracene	1.440	1.501	-4.2	94	-0.02
82	3,3'-Dichlorobenzidine	0.402	0.426	-6.0	100	-0.02
83	Chrysene	1.360	1.431	-5.2	95	-0.02
84	Bis(2-ethylhexyl)phthalate	0.765	0.873	-14.1	104	-0.02
85 c	Di-n-octyl phthalate	1.068	1.220	-14.2	106	-0.02
86 I	Perylene-d12	1.000	1.000	0.0	94	-0.03
87	Indeno(1,2,3-cd)pyrene	1.343	1.557	-15.9	99	-0.05
88	Benzo(b)fluoranthene	1.415	1.395	1.4	97	-0.03
89	Benzo(k)fluoranthene	1.381	1.295	6.2	91	-0.03
90 C	Benzo(a)pyrene	1.121	1.108	1.2	93	-0.03
91	Dibenzo(a,h)anthracene	1.110	1.287	-15.9	99	-0.05
92	Benzo(g,h,i)perylene	1.097	1.298	-18.3	101	-0.05

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF122222\
Data File : BF131779.D
Acq On : 22 Dec 2022 16:18
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Dec 23 03:30:25 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122022.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 23 03:03:57 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0