

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF122223\
 Data File : BF136709.D
 Acq On : 23 Dec 2023 03:52
 Operator : CG\JU
 Sample : 05926-03 10X
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MR-320-0-2

Quant Time: Dec 23 04:16:57 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122023.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 21 01:54:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.787	152	55618	20.000	ng	0.00
21) Naphthalene-d8	8.063	136	166586	20.000	ng	0.00
39) Acenaphthene-d10	9.816	164	71057	20.000	ng	-0.01
64) Phenanthrene-d10	11.310	188	154208	20.000	ng	0.00
76) Chrysene-d12	13.963	240	125664	20.000	ng	0.00
86) Perylene-d12	15.445	264	85756	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.422	112	37392	10.490	ng	0.00
7) Phenol-d6	6.428	99	42756	9.257	ng	-0.01
23) Nitrobenzene-d5	7.340	82	22645	6.956	ng	-0.02
42) 2,4,6-Tribromophenol	10.610	330	8951	10.696	ng	-0.01
45) 2-Fluorobiphenyl	9.134	172	39619	7.499	ng	-0.02
79) Terphenyl-d14	12.898	244	55704	6.195	ng	-0.02

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF122223\
Data File : BF136709.D
Acq On : 23 Dec 2023 03:52
Operator : CG\JU
Sample : 05926-03 10X
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MR-320-0-2

Quant Time: Dec 23 04:16:57 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF122023.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 21 01:54:25 2023
Response via : Initial Calibration

