

Data Path : Z:\HPCHEM1\BNA F\DATA\BF122315\
 Data File : BF084055.D
 Acq On : 24 Dec 2015 1:26
 Operator : UM/SJ
 Sample : G4907-06
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MW-11

Quant Time: Dec 24 07:57:26 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF122215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 22 18:54:09 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.82	152	45794	20.00	ng	-0.01
21) Naphthalene-d8	8.11	136	182019	20.00	ng	0.00
38) Acenaphthene-d10	9.87	164	81697	20.00	ng	0.00
63) Phenanthrene-d10	11.36	188	121368	20.00	ng	0.00
75) Chrysene-d12	13.99	240	87093	20.00	ng	-0.01
86) Perylene-d12	15.41	264	83459	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.45	112	101121	37.65	ng	0.01
7) Phenol-d6	6.49	99	14143m	4.48	ng	0.01
23) Nitrobenzene-d5	7.39	82	196092	66.18	ng	0.00
41) 2,4,6-Tribromophenol	10.66	330	81471	103.68	ng	0.00
44) 2-Fluorobiphenyl	9.18	172	422643	69.62	ng	-0.01
78) Terphenyl-d14	12.93	244	299416	81.99	ng	0.00
Target Compounds						
6) Aniline	6.49	93	45173	10.34	ng	# 84
17) 2-Methylphenol	7.10	107	18607	7.55	ng	# 91
20) 3+4-Methylphenols	7.26	107	318988	104.92	ng	# 44
31) Naphthalene	8.13	128	49702m	5.35	ng	
37) 2-Methylnaphthalene	8.83	142	360964	59.94	ng	98
49) Dimethylphthalate	9.58	163	24186	3.78	ng	# 74
54) Dibenzofuran	10.08	168	16290	2.22	ng	# 68
83) Bis(2-ethylhexyl)phthalate	13.97	149	12253	3.25	ng	# 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF122315\
Data File : BF084055.D
Acq On : 24 Dec 2015 1:26
Operator : UM/SJ
Sample : G4907-06
Misc :
ALS Vial : 40 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MW-11

Quant Time: Dec 24 07:57:26 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF122215.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 22 18:54:09 2015
Response via : Initial Calibration

